

A NEW FIBER BRAIDED SOFT BENDING ACTUATOR
FOR FINGER EXOSKELETON

ILI NAJAA AIMI BINTI MOHD NORDIN

UNIVERSITI TEKNOLOGI MALAYSIA

A NEW FIBER BRAIDED SOFT BENDING ACTUATOR
FOR FINGER EXOSKELETON

ILI NAJAA AIMI BINTI MOHD NORDIN

A thesis submitted in fulfilment of the
requirements for the award of the degree of
Doctor of Philosophy (Electrical Engineering)

Faculty of Electrical Engineering
Universiti Teknologi Malaysia

JUNE 2016

Dedicated, in thankful appreciation
for support, encouragement and understanding
to my beloved mother, father, brothers and sisters

ACKNOWLEDGEMENT

Praise to the Almighty.

First of all, thanks to our Creator for the continuous blessing and for giving me the strength and chances in completing this thesis.

Special thanks to my project supervisor, Associate Professor Ir. Dr. Ahmad ‘Athif bin Mohd Faudzi, my co-supervisor, Dr. Dyah Ekashanti Octorina Dewi and my co-supervisors during my attachment in Okayama University, Japan, Associate Professor Dr. Shuichi Wakimoto, Professor Dr. Koichi Suzumori, and Professor Dr. Takefumi Kanda for their guidance, and helpful supports.

My appreciation also goes to my family members who have been very encouraging in supporting me throughout the study. Many thanks for the love they have given me. To all my friends who have involved directly or indirectly towards the completion of my Ph.D. thesis, I really appreciate all their help, support and interest.

I would also like to thank Universiti Teknologi Malaysia (UTM), *MyPhd* scholarship under Ministry of Education (MOE) Malaysia and Okayama University for providing the facilities and equipment for this research.

ABSTRACT

This thesis presents a design, development and analysis of a novel bending-type pneumatic soft actuator as a drive source for a finger exoskeleton. Soft actuators are gaining momentum in robotic applications due to their simple structure, high compliance, high power-to-weight ratio and low production cost. Smaller and lighter soft actuator that can provide higher power transmission at lower operating air pressure will benefit finger actuation mechanism compared to motorized cable and pulley-driven finger rehabilitation devices. In this study, a soft actuator with new bending method is proposed. It is based on fibre reinforcement of two fibre braided angles of contraction and extension characteristics combined in a single-chamber cylindrical actuator. Another four design parameters identified that affect the bending motion and the actuating force were the air chamber diameter, position of fibre layer reinforcement, fibre reinforcement coverage angle, and silicone rubber materials. Geometrical and material parameters were varied in Finite Element Method (FEM) simulation for design optimization and some parameters were tested experimentally to validate the FEM models. The effects of fibre angles (contraction and extension) on the bending motion and force were analyzed. The optimized actuator can generate bending motion up to 131° bending angle and the end tip of the actuator can make contact with the other base tip at only 240 kPa given input pressure. Both displacement simulation and experimental testing results matched closely. Maximum bending force of 5.42 N was generated at 350 kPa. A wearable finger soft exoskeleton prototype with five optimized bending actuators was tested to drive finger flexion motion of eight healthy subjects with simulated paralysis conditions. The finger soft exoskeleton demonstrated the ability to provide gripping force of 3.61 ± 0.22 N, gained at 200 kPa given air pressure. The device can successfully provide assistance to weak fingers in gripping at least 240 g object. It shows potential in helping people with weakened finger muscle to be more independent in their finger rehabilitation exercise.

ABSTRAK

Tesis ini membentangkan pembangunan penggerak lenturan lembut jenis pneumatik terbaru sebagai sumber pemacu dalam menggerakkan jari. Penggerak lembut mendapat momentum dalam aplikasi robotik kerana strukturnya yang mudah, sifat pematuhan yang tinggi, nisbah kuasa kepada berat yang tinggi dan memerlukan kos produksi yang rendah. Penggerak lembut yang lebih kecil, ringan dan dapat menjana kuasa penghantaran yang tinggi pada pengendalian tekanan udara yang lebih rendah dapat memanfaatkan mekanisme penggerak jari jika dibandingkan dengan penggerak jari menggunakan motor dan takal. Dalam kajian ini, kaedah lenturan baru untuk penggerak lembut diusulkan. Ia adalah berdasarkan kepada dua sudut corak gentian yang mempunyai sifat penguncupan dan pemanjangan digabungkan dalam penggerak silinder tunggal. Empat lagi parameter reka bentuk telah dikenal pasti dapat memberi kesan terhadap gerakan lentur dan daya penggerakannya, iaitu isipadu ruang udara, kedudukan lapisan gentian, liputan sudut gentian dan bahan getah silikon. Parameter geometri dan bahan telah diubah dalam simulasi Kaedah Unsur Terhingga (FEM) dalam pengoptimuman reka bentuk dan ada di antaranya yang diuji secara eksperimen untuk mengesahkan model FEM yang direka. Kesan daripada sudut gentian penguncupan dan pemanjangan terhadap gerakan lentur dan daya lenturan telah dianalisa. Penggerak yang telah dioptimumkan boleh menjana 131° sudut lenturan dan hujung akhir penggerak dapat menyentuh hujung yang lainnya hanya pada 240 kPa tekanan udara. Kedua-dua keputusan anjakan daripada analisis simulasi dan eksperimen hampir berpadanan. Daya lenturan sebanyak 5.42 N dapat dijana pada 350 kPa. Penggerak lembut yang telah dioptimumkan menunjukkan keupayaan cengkaman kuasa sebanyak 3.61 ± 0.22 N pada 200 kPa tekanan udara. Sarung tangan kerangka luar menggunakan lima penggerak lenturan yang optimum telah diuji untuk membengkokkan jari lapan orang yang sihat yang dilemahkan. Berdasarkan ujian daya cengkaman yang dilakukan, sarung tangan kerangka luar ini dapat membantu menggerakkan jari yang lemah sekurang-kurangnya dalam menggenggam objek seberat 240 g. Ia menunjukkan potensi dalam membantu orang yang lemah otot jarinya untuk lebih berdikari dalam melakukan rehabilitasi senaman jari.

TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	ACKNOWLEDGEMENT	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	x
	LIST OF FIGURES	xi
	LIST OF ABBREVIATIONS	xv
	LIST OF SYMBOLS	xvi
	LIST OF APPENDICES	xviii
1	INTRODUCTION	1
	1.1 Background	1
	1.2 Problem Statements	4
	1.3 Objectives	5
	1.4 Scope	5
	1.5 Contributions	6
	1.6 Organization of the Thesis	6
2	LITERATURE REVIEW	8
	2.1 Introduction	8
	2.2 Rehabilitation Robotics	8
	2.3 Design of Pneumatic Soft Actuator	15
	2.3.1 Contracted and Extended	
	Linear-type PAM Soft Actuator	16
	2.3.2 Bending-type Soft Actuator	18

2.4	FEM Simulation of Soft Actuator	28
2.5	Summary	29
3	RESEARCH METHODOLOGY	31
3.1	Introduction	31
3.2	Proposed Bending Concept	34
3.3	Design Specifications of FBBA	35
	3.3.1 Variations of Fiber Angle Combination	40
	3.3.2 Optimization Parameters	43
3.4	FEM Simulation of FBBA	45
	3.4.1 Basic Protocol	45
	3.4.2 3-D Model Design Development	46
	3.4.3 FEM Analysis Conditions	48
	3.4.4 Mechanical Tensile Test of Rubber Materials	51
3.5	Fabrication of FBBA	53
	3.5.1 Mold Fabrication of FBBA	54
	3.5.1.1 AutoCAD Mold Design	54
	3.5.1.2 CNC Machining Process	57
	3.5.2 Rubber Parts and Fiber Knitting Fabrication Process	58
3.6	Experimental Testing of FBBA	60
	3.6.1 Displacement Data Quantification and Analysis	60
	3.6.2 Force Data Quantification and Analysis	63
3.7	Design and Development of SOFT-EXOS	65
3.8	Experimental Setup and Analysis of SOFT-EXOS in Healthy Subjects	70
	3.8.1 Subjects Population Criteria	70
	3.8.2 Measuring Set-Up	70
	3.8.3 Data Quantification and Analysis Method	72
3.9	Summary	73

4	RESULTS AND DISCUSSION	75
4.1	Introduction	75
4.2	FEM Simulation of Contracted and Extended Linear-type Actuator	76
4.3	Variations of Fiber Angle Combination (CS 1)	78
4.3.1	Displacement FEM Simulation Results	79
4.3.2	Displacement Experimental Results	81
4.3.3	Displacement Validation Results	84
4.3.4	Force Experimental Results	86
4.4	Variations of Air Chamber Diameter (CS 2)	88
4.5	Variations of Position of Fiber Layer Reinforcement (CS 3)	91
4.6	Variations of Fiber Coverage Angle (CS 4)	92
4.7	Variations of Rubber Material by Layer (CS 5)	94
4.8	Variations of Actuator Size (CS 6)	98
4.9	Repeatability and Reproducibility Test of the Optimized Actuator	99
4.10	SOFT-EXOS Grip Force Test in Healthy Subjects	101
4.10.1	Grip Force Measurement in Experiment A	101
4.10.2	Grip Force Measurement in Experiment B	103
4.11	Summary	106
5	CONCLUSIONS AND FUTURE WORKS	108
5.1	Conclusions	108
5.2	Suggestions for Future Works	110
	REFERENCES	111
	Appendices A – D	123 – 131

LIST OF TABLES

TABLE NO.	TITLE	PAGE
2.1	Summary of finger exoskeleton/ finger orthosis research	12
2.2	Summary of bending-type soft actuator	22
3.1	Definition of labeling a until l	36
3.2	Geometrical design parameters being fixed	37
3.3	Actuator comparison grouped by case studies	38
3.4	Details of design parameters for all actuator models	39
3.5	The derived fiber angles	42
3.6	Detail specifications of FEM simulation	51
3.7	Pitch and revolution of fiber angle with braid structure	55
3.8	Finger SOFT- EXOS design requirements	65
4.1	Single fiber angle actuator characteristic	78
4.2	Bending angle comparison in CS 1	86
4.3	Bending angle comparison in CS 3	92
4.4	Variations of fiber coverage angle in CS4	93
4.5	Variations of rubber materials by layer in CS 5	95
4.6	Bending angle comparison in CS 6	99
4.7	Grip force induced by right and left hand without wearing SOFT-EXOS	102
4.8	Grip force generated by SOFT-EXOS (worn on left hand)	104
4.9	Grip force generated by SOFT-EXOS (worn on right hand)	104

LIST OF FIGURES

FIGURE NO.	TITLE	PAGE
2.1	Functional hand rehabilitation tasks	9
2.2	Soft actuation robotic device for (a) shoulder, and (b) knee	14
2.3	Soft actuation robotic device for (a) wrist, (b) elbow	15
2.4	Contracted and Extended linear-type PAM	17
2.5	Contraction and extension PAM actuator, (a) at rest state (b) after 1.5 MPa water pressurization	21
2.6	Two types of PAM actuator bundled together, (a) at rest state (b) after water pressurization	21
2.7	Deformation of two chamber-type actuator	29
2.8	Deformation of bellow-type actuator	29
3.1	Methodology design flow chart	33
3.2	The proposed fiber braided bending mechanism	34
3.3	3-D and cross sectional view of FBBA	35
3.4	Fiber angle derivation model	40
3.5	Six models of different combination of contraction and extension fiber angles	43
3.6	Variations of air chamber diameter in CS 2	44
3.7	Variations of fiber layer reinforcement position from center point in CS 3	44
3.8	Variations of contraction and extension fiber coverage angle in CS 4	44
3.9	Variations of rubber materials by layer in CS 5	45
3.10	Variations of actuator total diameter in CS 6	45
3.11	Cylindrical structure composed from 10° base elements	46

3.12	Layout model of a fully crossed interlaced fiber braid structure of fiber angle 81°, 72°, 46° and 35°	47
3.13	A line element of two connected nodes	48
3.14	Boundary conditions	50
3.15	Increment of steps from 0 to 1 second	50
3.16	Samples as per ASTM D412 for uniaxial tensile testing	52
3.17	Uniaxial tensile test result for RTV silicone rubber	52
3.18	Stages of actuator fabrication	53
3.19	Layout of cylindrical actuator body at fiber reinforcement layer	55
3.20	Top views of all mold designs	56
3.21	3-D view of mold designs of inner rubber layer, outer rubber layer and end cap	57
3.22	Tools and equipment for CNC machining	58
3.23	Rubber fabrication process	59
3.24	Fiber reinforcement steps	60
3.25	Experimental setup of displacement measurement	61
3.26	Definition of bending angle, Φ_B	62
3.27	Experimental setup of force measurement	64
3.28	Top-view of SOFT-EXOS	66
3.29	Side-view of SOFT-EXOS	66
3.30	Prototypes of SOFT-EXOS	67
3.31	SOFT-EXOS system	68
3.32	Control box	69
3.33	Relationship of PWM input signals with the desired output pressures	69
3.34	Gripping and sitting posture for grip force measurement	71
4.1	Behavior of single fiber angle 3-D FEM Models before and after pressurization	76
4.2	Displacement plots of contracted-type and extended-type actuator at Z axis as input pressure increases	77
4.3	YZ cross section (horizontal view) of six models	

	constructed from different combination of θ_C and θ_E	79
4.4	Simulation results of (a) displacement in X axis versus input pressure, (b) Y axis versus input pressure, (c) Z axis versus input pressure, (d) tip trajectory until 300 kPa, and (e) bending angle versus input pressure	80
4.5	All molds produced by the CNC Roland Machine Model MDX-40	82
4.6	Fibers reinforcement surrounding the inner rubber layer	82
4.7	Six models of different combination of θ_C and θ_E	83
4.8	Experimental results of (a) tip trajectory until 300 kPa, and (b) bending angle versus input pressure	83
4.9	Displacement of actuator Model A	84
4.10	Comparison of tip trajectories between simulation and experiment for (a) Model A, (b) Model B, (c) Model C, (d) Model D, (e) Model E, and (f) Model F	85
4.11	Experimental results of the generated force at tip versus input pressure, of (a) all models, (b) Model A, and (c) Model B	87
4.12	XY cross sections of three actuator models constructed from different d_{cham}	88
4.13	Simulation results of (a) tip trajectory until 300 kPa, and (b) bending angle versus input pressure	89
4.14	Experimental results of (a) tip trajectory until 300 kPa, (b) bending angle versus input pressure, and (c) generated force at tip versus input pressure	90
4.15	Comparison of tip trajectory operated until 300 kPa between simulation and experimental results of Model H	90
4.16	3-D view of r_{fiber} variations	91
4.17	Simulation results of tip trajectory until 300 kPa	92
4.18	3-D view of α_C and α_E variations	93
4.19	Simulation results of (a) tip trajectory until 300 kPa,	

	and (b) bending angle versus input pressure	94
4.20	Simulation results of (a) tip trajectory until 130 kPa, and (b) bending angle versus input pressure	95
4.21	Experimental results of (a) tip trajectory until 190 kPa, (b) bending angle versus input pressure, and (c) generated force at tip versus input pressure	96
4.22	Tip trajectories comparison between simulation and experimental results of Model R until 190 kPa	97
4.23	Variations of d_{total}	98
4.24	Simulation results of bending angle versus input pressure	99
4.25	Bending displacement of the optimized actuator model at 240 kPa air pressurization	100
4.26	Experimental results of (a) displacement Y, (b) displacement Z, and (c) tip trajectory operated until 240 kPa	100
4.27	Experimental results of generated force at tip versus input pressure	101
4.28	Continuous grip force visualized in Vernier Logger Lite software	102
4.29	Grip forces measured at 200 kPa of a right hand	103
4.30	Mean grip force induced by SOFT-EXOS	105
4.31	Relationship of input pressure with actuator design variables	106

LIST OF ABBREVIATIONS

FBBA	-	Fiber Braided Bending Actuator
SOFT-EXOS	-	Wearable Finger Soft Exoskeleton
FEA	-	Finite Element Analysis
FEM	-	Finite Element Method
3-D	-	Three-Dimensional
ADL	-	Activities in Daily Living
MCP	-	Metacarpal-Phalangeal
PIP	-	Proximal Interphalangeal
DIP	-	Distal Interphalangeal
DOF	-	Degree-of-freedom
OSHA	-	Occupational Safety and Health Administration
PAM	-	Pneumatic Artificial Muscle
FMA	-	Flexible Microactuator
CS	-	Case Study
SIM	-	Simulation
EXP	-	Experiment
YM	-	Young's Modulus
PR	-	Poisson's Ratio
CAD	-	Computer-Aided Design
CNC	-	Computer Numerical Control
MS	-	Microsoft
RTV	-	Room Temperature Vulcanizing
SD	-	Standard Deviation
ABS	-	Acrylonitrile-Butadiene-Styrene

LIST OF SYMBOLS

F	-	contraction force
D_o	-	actual diameter of the fiber layer at initial state
P	-	applied input pressure
θ_o	-	fiber braid angle before pressurization
ε	-	contraction ratio
θ_C	-	extension fiber angle
θ_E	-	contraction fiber angle
d_{cham}	-	air chamber diameter
r_{fiber}	-	fiber layer radius
t_i	-	inner rubber layer thickness
t_o	-	outer rubber layer thickness
Z	-	fiber spacing between two consecutive fiber
α_C	-	contraction fiber coverage angle
α_E	-	extension fiber coverage angle
M_{in}	-	inner layer rubber material
M_{out}	-	outer layer rubber material
d_{total}	-	actuator total diameter
L_{total}	-	actuator total length
L_{cap}	-	end cap fittings length
L_f	-	length of actuator body with fiber reinforcement
α_{braid}	-	coverage angle of fully crossed interlaced fiber braid

C_{180}	-	circumference length of 81° fiber braid structure
C_{90}	-	circumference length of 72° fiber braid structure
C_{30}	-	circumference length of 46° fiber braid structure
C_{20}	-	circumference length of 35° fiber braid structure
Z_{small}	-	fiber spacing between two consecutive smallest element
C	-	circumference length of full circle
P	-	pitch
R	-	revolution
Z_{new}	-	displacement of the tip at Z direction from the centre end point of actuator's proximal end
Φ_B	-	bending angle
$\#SIM$	-	bending angle from simulation data
$\#EXP$	-	bending angle from experiment data
\bar{X}_{TF}	-	mean of the generated tip force from 3 experiment trials
s_{TF}	-	standard deviation of the generated tip force from 3 experiment trials
X_i	-	generated tip force value of each trial
N	-	total number of trials
\bar{X}_G	-	mean of the generated grip force from 3 experiment trials
s_G	-	standard deviation of the generated grip force from 3 experiment trials
X_G	-	generated grip force value of each trial
μ_G	-	mean of the generated grip force representing total of 8 subjects
SD_G	-	standard deviation of the generated grip force representing total of 8 subjects
N_P	-	total number of subjects participated in the testing

LIST OF APPENDICES

APPENDIX	TITLE	PAGE
A	List of Publications	123
B	List of Achievements	125
C	Valve Control Source Code	126
D	Displacement Data of the Optimized Actuator	129

CHAPTER 1

INTRODUCTION

1.1 Background

It was reported that 49% of older people that face difficulties in performing physical tasks were caused by musculoskeletal diseases such as Arthritis, 13.7% by heart disease, 12% by injury, 11.7% by old age, 6% by lung disease, and remaining 2.9% were caused by stroke [1]. Among those contributing factors, people with stroke was reported to be facing difficulties in using upper extremities and performing basic activities of daily living [1].

Worldwide, stroke is the second commonest cause of death and a leading cause of adult disability [2]. According to statistics announced by National Stroke Association of Malaysia (NASAM), stroke is the third highest cause of death in Malaysia [3]. Nearly 40,000 people in Malaysia suffer from stroke every year, affecting adults more than children. Some studies shows that majority of stroke cases occurred due to cerebral infarction (50% - 87%), followed by cerebral hemorrhage (20% - 30%) [4]–[7]. Other most frequent risk factors include hypertension, diabetes mellitus and previous stroke [5], [8]–[13].

Paralysis, a common disability resulting from stroke, may range from complete inability to move to less than total strength, thus affecting the stroke survivors in their daily activities, such as causing difficulty in walking and grasping objects. Depending on the severity of neurological deficits, 19% of stroke survivors

were very severely disabled, 4% of them severely disabled, 26% moderately disabled, 41% minor disabled and no disability shown for the remaining 10% [14].

After completing stroke rehabilitation, a study shows 11% of the stroke survivors still very severely and severely disabled, 11% moderately disabled and another 78% shows minimum or no disability [14]. Other study in Singapore shows 91.9% of the stroke survivors can independently conduct self-care activities after completed rehabilitation [13]. Other than stroke survivors, elderly and patients with prolonged intensive care (ICU) stay [15] can also exhibit general weakness and problems in coordinating that need physical rehabilitation.

Early stage, repetitive and continuous rehabilitation can help the brain relearns lost skills much faster and with more significant results. The assistance of robotic devices, might also promote a cost-effective therapy for stroke survivors to maintain their ability to move after receiving standard in-hospital rehabilitation [16]–[18]. Rehabilitation devices specifically designed for restoring hand function, usually known as finger exoskeleton must be able to at least assist flexion motion of fingers. Rondi Blackburn, a medical professional developed a theory that repeated exercises of the affected hand and fingers will open up new pathways of communication between the brain and the stroke-affected area (American Heart Association). Strength, mobility and precision exercises are types of exercises usually being performed to stroke survivor rehabilitation.

With the advancements in robotics and mechatronics research in the last decade, rehabilitation robotics has become an active research area. Rehabilitation Robotics is the application of robotic technology to the rehabilitative needs of people with disabilities as well as the growing elderly population [19]. The main restriction in the current finger rehabilitation robot system is the complexity in its structure which requires metallic or plastic alignment in every finger joint in the finger actuation system [20]–[27]. Heavy unit from plastic and metal load can contribute to wearer discomfort.

To counter these problems, Soft Actuator, also known as Rubber Actuator, is a pneumatic driven actuator which can offer simpler structure, high power-to-weight ratio, lightweight, comparative low cost, and easy maintenance [28]–[31], suitable to be utilized in finger exoskeleton. It converts energy from compressed air into various motions depending on its design. In comparison to hydraulic actuator, pneumatic actuator is relatively small in size, requires smaller tank for air storage and is lightweight. It is also easy to control with only simple on-off type control. It does not produce heat except for friction, thus the risk of accidental fire is low. This can promote safer interaction with human.

Soft actuation by using Soft Actuator is still a young approach in robotics engineering and currently developing. There are a few research groups around the world that implement soft actuator in robotics application. To name a few, there was Prof. Koichi Suzumori Research Group established at Okayama University in the early 2000s. The group focuses on developing soft actuators mechanism in various fields [28]–[50], especially in object manipulation and medical assistance. In the same university, there was also Prof. Toshiro Noritsugu Research Group that implements soft actuation in rehabilitation assistance [51]–[54]. Other groups that are currently actively involved in soft actuation research are the Whitesides and Conor J Walsh Research Group from Harvard University. Since 2011, they have been rapidly developing soft actuator technology especially in the field of biomimetic and assistive wearable rehabilitation [55]–[63].

Although some studies in the recent years have focused on applying soft actuator in finger exoskeleton for object grasping manipulation [51]–[54], [64], there has been very little discussion on the gripping force assistance and these studies have been limited to the generated actuator force, not gripping force from the soft exoskeleton unit.

This research proposed a novel soft exoskeleton to assist fingers weak in finger flexion motion by utilizing a new bending-type soft actuator. The research is expected to contribute highly to the development of a user-friendly, comfortable,

safer and more powerful finger exoskeleton prototype, where it can help in reducing therapists' workloads in performing therapy tasks to stroke survivors in the future.

1.2 Problem Statements

Wearable devices designed to suit many parts of body such as shoulder, elbow and finger assist system are gaining popularity at present due to their portability and can support movement in weak muscles especially in stroke survivors. Robots or devices that are comfortable to wear and easy to be used will increase user motivation in performing the exercises, thus hastening recovery. Simple mechanical design with high flexibility and proper softness in motion and touch is required in order to provide efficient therapy [58], [63], [65].

Although some finger exoskeletons show promising results in providing grasping motion [51], [53], [54], [61] the size and force performance of bending-type actuator utilized in the actuation system can still be improved. In addition, little data of the assisted grasping force has been found [51], [53], [54], [61]. Smaller size bending-type soft actuator with increased power transmission at lower operated air pressure is required.

A new soft actuator design suitable for power soft actuation is proposed based on fiber braided reinforcement in McKibben actuator. McKibben pneumatic artificial muscle (PAM) is known to be high achieving contraction force actuator due to fiber braided layer structure incorporated at the outer layer of its cylindrical body structure [66]. The fiber layer restrains radial contraction while promoting contraction forces. Due to its high force capability driven by the fiber braided reinforcement, two different fiber braided angles in a single chamber were proposed to obtain desired bending motion and force for finger flexion actuation. Currently, bending soft actuator prototype shown from literature that uses fiber braided contraction and extension fiber angles reinforcement are not mechanically interlocked and bulky [38].

1.3 Objectives

The followings are the objectives of the research:

1. To simulate, fabricate and optimize the proposed bending soft actuator using two fiber braided angles (contraction and extension) in one chamber.
2. To validate the 3-D FEM models of the proposed bending soft actuator with the fabricated actuator models.
3. To implement finger soft exoskeleton that utilized the optimized bending soft actuator in finger flexion.

The main objective of this research is to develop a new single chamber fiber braided bending-type actuator (FBBA) using combined contraction and extension-type fiber reinforcement. The new bending mechanism of soft actuator developed in this research is utilized in finger soft exoskeleton and is expected to be able to provide grasping assistance in weak fingers at least in holding light objects.

1.4 Scope

The followings are the scope of the research:

1. FBBA design development using technical mathematical drawings of combined contraction and extension fiber angles reinforcement in one chamber.
2. Proof of FBBA bending concept in FEM simulation analysis using MARC[®] Mentat software.
3. FBBA 3-D FEM optimization and analysis based on several design parameters (air chamber diameter, position of fiber reinforcement, rubber materials, fiber reinforcement coverage angle) performed with FEM software, MARC[®].
4. Fabrications of the proposed FBBA using molding, rubber bonding and fiber knitting techniques.
5. Evaluation of the FBBA based on tip trajectory plot, bending angle and generated tip force performance.

6. Implementation of the optimized FBBA's onto the glove (SOFT-EXOS) utilizing rubber band and elastic textile attachment.
7. Evaluation of FBBA in SOFT-EXOS implementation based on the measured mean value of the assisted grip force in 8 healthy subjects.

1.5 Contributions

The main research contributions of the research are as follows:

1. A new actuator bending actuator concept using two fiber braided angles (contraction and extension) in one chamber that can produce bending motion is introduced.
2. A new 3-D FEM model designed with combined fiber pattern reinforcement is developed and validated by experimental testing of the fabricated actuator model.
3. A new prototype of finger exoskeleton utilizing the novel bending actuator and flexible glove attachment is developed.

1.6 Organization of the Thesis

The thesis is organized in five chapters. Background of the research field, introduction to the recognized problems that need to be solved, the proposed solutions to the research problems, the scopes of the study, and some recognized contributions of the research are introduced in Chapter 1.

In Chapter 2, the literature review on finger flexion methods in finger exoskeleton are presented in several actuating mechanisms, for example pulley system-operated, motor-operated, and pneumatic-operated. Soft actuation mechanisms showing different resulting motions, particularly in pneumatic and hydraulic operated system applied in diverse applications are also presented. Various bending concepts proposed from different groups of researcher studied in flexible and high force applications are also studied.

The research flow and methodology used in the development of the proposed actuator design and the implementation of the actuator design in simulation, experimental testing and application- based study are shown in Chapter 3.

Chapter 4 mainly presents the results of FEM simulations and experimental testing of the proposed actuators that were evaluated by displacement and force performance. FEM model validation and optimization results in several geometrical parameter changes are also presented. In addition, the feasibility study conducted on healthy subjects in order to evaluate the performance of the proposed actuator design implemented in the finger soft exoskeleton system is presented.

Finally, the summary of research contributions and future solutions gained from the study are covered in Chapter 5.

REFERENCES

1. Ettinger, J. W. H., Fried, L. P., Harris, T., Shemanski, L., Schulz, R. and Robbins, J. Self-Reported Causes of Physical Disability in Older People: The Cardiovascular Health Study. *Journal of the American Geriatrics Society*, 1994, 42(10): 1035–1044.
2. Murray, C. J. L. and Lopez, A. D. Mortality by Cause for Eight Regions of the World: Global Burden of Disease Study. *The Lancet*, 1997, 349(9061): 1269–7126.
3. Stroke Statistic. Retrieved from the NASAM website: http://www.nasam.org/english/prevention-what_is_a_stroke.php.
4. Hu, H. H., Sheng, W. Y., Chu, F. L. and Chiang, B. N. Incidence of Stroke in Taiwan. *Stroke*, 1992, 23(9): 1237–1241.
5. Nwosu, C. M., Nwabueze, A. C. and Ikeh, V. O. Stroke at the Prime of Life: A Study of Nigerian Africans between the Ages of 16 and 45 Years. *East Afr. Med J*, 1992, 69(7): 384–390.
6. Rozenhul-Sorokin, N., Ronen, R., Tamir, A., Geva, H. and Eldar, R. Stroke in the Young in Israel: Incidence and Outcomes. *Stroke*, 1996, 27(5): 838–841.
7. Ng, W. K., Goh, K. J., George, J., *et al.* A Comparative Study of Stroke Subtypes between Asians and Caucasians in Two Hospital-Based Stroke Registries. *Neurol J Southeast Asia*, 1998, 3: 19-26.
8. Fu-chang, D., Hai-yan, W., Cai-liang, Y., *et al.* A Cohort Study on the Difference of Stroke Incidence Between Urban and Rural Population and Related Influential Factors. *Chung-Hua-Liu-Hsing-Ping-Hsueh-Tsa-Chih*, 1995, 16(5): 278–280.
9. Feigin, V. L., Wiebers, D. O., Nikitin, Y. P., O’Fallon, W. M. and Whisnant, J. P. Stroke Epidemiology in Novosibirsk, Russia: A Population-Based Study. *Mayo Clin Proc*, 1995, 70(9): 847–852.

10. Shimamoto, T., Iso, H., Iida, M., and Komachi, Y. Epidemiology of Cerebrovascular Disease: Stroke Epidemic in Japan. *J Epidemiol*, 1996, 6(3): S43–47.
11. Bharucha, N. and Kuruvilla, T. Epidemiology of Stroke in India. *Neurol J Southeast Asia*, 1998, 3: 5–8.
12. Ryglewicz, D., Baranska-Gieruszczak, M., Czlonkowska, A., Lechowicz, W. and Hier D. B. Stroke Recurrence among 30 Days Survivors of Ischaemic Stroke in a Prospective Community-Based Study. *Neurol Res*, 1997, 19(4): 377–379.
13. Venketasubramanian, N. The Epidemiology of Stroke in ASEAN Countries- A Review. *Neurol J Southeast Asia*, 1998, 3: 9–14.
14. Jorgensen, H. S., Nakayama H., Raaschou, H. O., Vive-Larsen, J., Stoier, M. and Olsen, T. S. Outcome and time course of recovery in stroke. Part I: Outcome: The Copenhagen Stroke Registry. *Arch Phys Med Rehabil*, 1995, 76(5): 399–405.
15. Jones, C. and Griffiths, R. D., Identifying Post Intensive Care Patients who may Need Physical Rehabilitation. *Clin. Intensive Care*, 2000, 11(1): 35–38.
16. Reinkensmeyer, D. J., Kahn, L. E., Averbuch, M., McKenna-Cole, B. D. S. and Rymer, W. Z. Understanding and Treating Arm Movement Impairment after Chronic Brain Injury: Progress with the ARM Guide. *J. Rehabil. Res. Dev.*, 2000, 37(6): 653–662.
17. Jack, D., Boian, R., Merians, A. S., Tremaine, M., Burdea, G. C., Adamovich, S. V., Recce, M. and Poizner, H. Virtual Reality-Enhanced Stroke Rehabilitation. *IEEE Trans. Neural Syst. Rehabil. Eng.*, 2001, 9(3): 308–318.
18. Krebs, H. I., Volpe, B. T., Aisen, M. L. and Hogan, N. Increasing Productivity and Quality of Care: Robot-Aided Neuro-Rehabilitation. *J. Rehabil. Res. Dev.*, 2000, 37(6): 639–652.
19. Hillman, M. Introduction to the Special Issue on Rehabilitation Robotics. *Robotica*, 1998, 16(5): 485.
20. Kawasaki, H., Ito, S., Ishigure, Y., Nishimoto, Y., Aoki, T., Mouri, T., Sakaeda, H. and Abe, M. Development of a Hand Motion Assist Robot for Rehabilitation Therapy by Patient Self-Motion Control. *Proceedings of the 2007 IEEE 10th Int. Conf. Rehabil. Robot.* June 13-15. 2007. 234–240.

21. Godfrey, S. B., Schabowsky, C. N., Holley, R. J. and Lum, P. S. Hand Function Recovery in Chronic Stroke with HEXORR Robotic Training: A Case Series. *Proceedings of the 2010 Annual International Conference of the IEEE Engineering in Medicine and Biology*. August 31- Sept 4. 2010. 4485–4488.
22. Schabowsky, C. N., Godfrey, S. B., Holley, R. J. and Lum, P. S. Development and Pilot Testing of HEXORR: Hand EXOskeleton Rehabilitation Robot. *J. Neuroeng. Rehabil.*, 2010, 7: 36.
23. Wege, A. and Hommel, G. Development and Control of a Hand Exoskeleton for Rehabilitation Construction of the Exoskeleton. *Proceedings of the 2005 IEEE/RSJ International Conference on Intelligent Robots and Systems*. August 2-6. 2006. 149–157.
24. Lucas, L., DiCicco, M. and Matsuoka, Y. An EMG-Controlled Hand Exoskeleton for Natural Pinching. *J. Robot. Mechatronics*, 2004, 16(5): 1–7.
25. Dovat, L., Lamercy, O., Johnson, V., Salman, B., Wong, S., Gassert, R., Burdet, E., Teo, C. L. and Milner, T. A. Cable Driven Robotic System to Train Finger Function After Stroke. *Proceedings of the 2007 IEEE 10th International Conference on Rehabilitation Robotics*. June 13-15. 2007. 222–227.
26. Bouzit, M., Burdea, G., Popescu, G. and Boian, R. The Rutgers Master II- New design Force-Feedback Glove. *IEEE/ASME Transactions on Mechatronics*, 2002, 7(2): 256–263.
27. Chiri, A., Giovacchini, F., Vitiello, N., Cattin, E., Roccella, S., Vecchi, F. and Carrozza, M. C. HANDEXOS: Towards an Exoskeleton Device for the Rehabilitation of the Hand. *Proceedings of the 2009 IEEE/RSJ International Conference on Intelligent Robots and Systems*. October 10-15. 2009. 1106–1111.
28. Suzumori, K., Hama, T. and Kanda, T. New Pneumatic Rubber Actuators to Assist Colonoscope Insertion. *Proceedings of the 2006 IEEE International Conference on Robotics and Automation*. May 15-19. 2006. 1824–1829.
29. Suzumori, K., Endo, S., Kanda, T., Kato, N. and Suzuki, H. A Bending Pneumatic Rubber Actuator Realizing Soft-Bodied Manta Swimming Robot. *Proceedings of the 2007 IEEE International Conference on Robotics and Automation*. April 10-14. 2007. 4975–4980.

30. Wakimoto, S., Ogura, K., Suzumori, K. and Nishioka, Y. Miniature Soft Hand with Curling Rubber Pneumatic Actuators. *Proceedings of the 2009 IEEE International Conference on Robotics and Automation*. May 12-17. 2009. 556–561.
31. Nagase, J. Wakimoto, S. Satoh, T. Saga, N. and Suzumori, K. Design of a Variable-Stiffness Robotic Hand Using Pneumatic Soft Rubber Actuators. *Smart Mater. Struct.*, 2011, 20(10): 1–9.
32. Suzumori K. and Saito F. Micro Rubber Structure Realizing Multi-Legged Passive Walking. *Proceedings of the 2008 IEEE/RSJ International Conference on Intelligent Robots and Systems*. September 22-26. 2008. 445–450.
33. Wakimoto, S., Suzumori, K., Kanda, T. and Kure, K. Soft Pressure Sensor Embedded in McKibben Actuator. *Proceedings of the 2007 IEEE/ASME International Conference on Advanced Intelligent Mechatronics*. September 4-7. 2007. 1-6.
34. Kure, K., Kanda, T., Suzumori, K. and Wakimoto, S. Flexible Displacement Sensor Using Injected Conductive Paste. *Sensors Actuators A Phys.*, 2008, 143(2): 272–278.
35. Wakimoto, S., Suzumori, K. and Kanda, T. Development of Intelligent McKibben Actuator with Built-In Soft Conductive Rubber Sensor. *Proceedings of the 2005 International Conference on Solid State Sensors, Actuators and Microsystems*. June 5-9. 2005. 1:745–748.
36. Onoe, H., Suzumori, K. and Wakimoto, S. Optimum Design of Pneumatic Multi-Chamber Rubber Tube Actuator Generating Traveling Deformation Waves for Colonoscope Insertion. *Proceedings of the 2008 IEEE/ASME International Conference on Advanced Intelligent Mechatronics*. July 2-5. 2008. 31–36.
37. Saito, F. and Suzumori, K. Micro Rubber Structure Realizing Multi-Legged Passive Walking- Integration and Miniaturization by Micro Rubber Molding Process. *Proceedings of the 2009 IEEE/RSJ International Conference on Intelligent Robots and Systems*. October 10-15. 2009. 3025–3030.
38. Iwata, K., Suzumori, K. and Wakimoto, S. Bundling Several Extending/Contracting Muscles to Power Soft Mechanisms. *8th JFPS International Symposium on Fluid Power*. October 25-28. 2011. 338–343.

39. Iwata, K., Suzumori, K. and Wakimoto, S. A Method of Designing and Fabricating McKibben Muscles Driven by 7 MPa Hydraulics. *Int. J. Autom. Technol.*, 2012, 6(4): 482–487.
40. Ozaki, K., Wakimoto, S., Suzumori, K. and Yamamoto, Y. Novel Design of Rubber Tube Actuator Improving Mountability and Drivability for Assisting Colonoscope Insertion. *Proceedings of the 2011 IEEE International Conference on Robotics and Automation*. May 9-13. 2011. 3263–3268.
41. Wakimoto, S., Suzumori, K. and Ogura, K. Miniature Pneumatic Curling Rubber Actuator Generating Bidirectional Motion with One Air-Supply Tube. *Adv. Robot.*, 2011, 25(9–10): 1311–1330.
42. Mori, M., Suzumori, K., Takahashi, M. and Hasoya, T. Very High Force Hydraulic McKibben Artificial Muscle with a Cord Sleeve. *Adv. Robot.*, 2012, 24(1-2): 233-254.
43. Wakimoto, S. and Suzumori, K. Fabrication and Basic Experiments of Pneumatic Multi-Chamber Rubber Tube Actuator for Assisting Colonoscope Insertion. *Proceedings of the 2010 IEEE International Conference on Robotics and Automation*. May 3-7. 2010. 3260–3265.
44. Iwata, K. Suzumori, K. and Wakimoto, S. Combination of Extending/Contracting Artificial Muscles Realizing Power Soft Mechanisms. *Proceedings of the 2011 4th International Conference on Manufacturing, Machine Design and Tribology*. April 22-25. 2011. 1: 1–2.
45. Yang, Q., Zhang, L., Bao, G., Xu, S. and Ruan, J. Research on Novel Flexible Pneumatic Actuator FPA. *Proceedings of the 2004 IEEE International Conference on Robotics and Automation*. December 1-3. 2004. 385–389.
46. Udupa, G. and Sreedharan, P. Robotic Gripper Driven by Flexible Microactuator Based on an Innovative Technique. *Proceedings of the 2010 IEEE Workshop on Advanced Robotics and its Social Impacts*. October 26-28. 2010. 111-116.
47. Deng, M., Wang, A., Wakimoto, S. and Kawashima, T. Characteristic Analysis and Modeling of a Miniature Pneumatic Curling Rubber Actuator. *Proceedings of the 2011 International Conference on Advanced Mechatronic Systems*. August 11-13. 2011. 534–539.
48. Shapiro, Y., Wolf, A. and Gabor, K. Bi-Bellows: Pneumatic Bending Actuator. *Sensors Actuators A Phys.*, 2011, 167(2): 484–494.

49. Higuchi, T. Next Generation Actuators Leading Breakthroughs. *J. Mech. Sci. Technol.*, 2010, 24(1): 13–18.
50. Chen, G., Pham, M. T., Redarce, T. and Villeurbanne, I. D. L. Development and Kinematic Analysis of a Silicone- Rubber Bending Tip for Colonoscopy. *Proceedings of the 2006 IEEE/RSJ International Conference on Intelligent Robots and Systems*. October. 2006. 168–173.
51. Noritsugu, T., Yamamoto, H., Sasaki, D. and Takaiwa M. Wearable Power Assist Device for Hand Grasping Using Pneumatic Artificial Rubber Muscle. *Proceedings of the 2004 SICE Annual Conference*. August 4-6. 2004. 420–425.
52. Noritsugu, T. Pneumatic Soft Actuator for Human Assist Technology. *Proceedings of the 6th JFPS International Symposium on Fluid Power*. November 7-20. 2005. 11–20.
53. Noritsugu, T., Takaiwa, M. and Sasaki, D. Power Assist Wear Driven with Pneumatic Rubber Artificial Muscles. *Proceedings of the 2008 15th International Conference on Mechatronics and Machine Vision in Practice*. December 2-4. 2008. 539–544.
54. Noritsugu, T., Takaiwa, M. and Sasaki, D. Development of Power Assist Wear Using Pneumatic Rubber Artificial Muscles. *Asia International Symposium on Mechatronics*. August 27-31. 2008. 371–375.
55. Polygerinos, P., Wang, Z., Overvelde, J. T. B., Galloway, K. C., Wood, R. J., Bertoldi, K. and Walsh, C. J. Modeling of Soft Fiber-Reinforced Bending Actuators. *IEEE Transactions on Robotics*, 2015, 31(3): 778–789.
56. Maeder-York, P., Clites, T., Boggs, E., Neff, R., Polygerinos, P., Holland, D., Stirling, L., Galloway, K., Wee, C. and Walsh, C. Biologically Inspired Soft Robot for Thumb Rehabilitation. *J. Med. Device.*, 2014, 8(2): 1–3.
57. Polygerinos, P., Wang, Z., Galloway, K. C., Wood, R. J. and Walsh, C. J. Soft Robotic Glove for Combined Assistance and At-Home Rehabilitation. *Rob. Auton. Syst.*, 2014, 73: 135-143.
58. Aubin, P. M., Sallum, H., Walsh, C., Stirling, L. and Correia, A. A Pediatric Robotic Thumb Exoskeleton for At-Home Rehabilitation: The Isolated Orthosis for Thumb Actuation (IOTA). *Proceedings of the 2013 IEEE International Conference on Rehabilitation Robotics*. June 24-26. 2013. 1-6.

59. Connolly, F., Polygerinos, P., Walsh, C. J. and Bertoldi, K. Mechanical Programming of Soft Actuators by Varying Fiber Angle *Soft Robot.*, 2015, 2(1): 26–32.
60. Galloway, K. C., Polygerinos, P., Walsh, C. J. and Wood R. J. Mechanically Programmable Bend Radius for Fiber-Reinforced Soft Actuators. *Proceedings of the 2013 16th International Conference on Advanced Robotics*. November 25-29. 2013. 1–6.
61. Polygerinos, P., Lyne, S., Nicolini, L. F., Mosadegh, B., Whitesides, G. M. and Walsh, C. J. Towards a Soft Pneumatic Glove for Hand Rehabilitation. *Proceedings of the 2013 IEEE/RSJ International Conference on Intelligent Robots and Systems*. November 3-7. 2013. 1512–1517.
62. Mosadegh, B., Polygerinos, P., Keplinger, C., Wennstedt, S., Shepherd, R. F., Gupta, U., Shim, J., Bertoldi, K., Walsh, C. J. and Whitesides, G. M. Pneumatic Networks for Soft Robotics that Actuate Rapidly *Adv. Funct. Mater.*, 2014, 24 (15): 2163–2170.
63. Polygerinos, P., Galloway, K. C., Savage, E., Herman, M., Donnell, K. O. and Walsh, C. J. Soft Robotic Glove for Hand Rehabilitation and Task Specific Training. *Proceedings of the 2015 IEEE International Conference on Robotics and Automation*. May 26-30. 2015. 2913–2919.
64. Sasaki, D., Noritsugu, T., Yamamoto, H. and Takaiwa, M. Wearable Power Assist Device for Hand Grasping Using Pneumatic Artificial Rubber Muscle. *Proceedings of the 2004 SICE Annual Conference*. August 4-6. 2004. 655–660.
65. Heo, P., Gu, G. M., Lee, S. J., Rhee, K. and Kim, J. Current Hand Exoskeleton Technologies for Rehabilitation and Assistive Engineering. *Int. J. Precis. Eng. Manuf.*, 2012, 13(5): 807–824.
66. Andrikopoulos, G., Nikolakopoulos, G. and Manesis, S. A Survey on Applications of Pneumatic Artificial Muscles. *Proceedings of the 2011 19th Mediterranean Conference on Control and Automation*. June 20-23. 2011. 1439–1446.
67. Wolf, S. L., Lecraw, D. E., Barton, L., A. and Jann, B. B. Forced Use of Hemiplegic Upper Extremities to Reverse the Effect of Learned Nonuse Among Chronic Stroke and Head-Injured Patients. *Exp. Neurol.*, 1989, 104(2): 125–132.

68. Fischer, H. C., Stubblefield, K., Kline, T., Luo, X., Kenyon, R. V. and Kamper, D. G. Hand Rehabilitation Following Stroke: A Pilot Study of Assisted Finger Extension Training in a Virtual Environment. *Top. Stroke Rehabil.*, 2007, 14(1): 1–12.
69. Huang, Y. Y. and Low, K. H. Initial Analysis and Design of an Assistive Rehabilitation Hand Device with Free Loading and Fingers Motion Visible to Subjects. *Proceedings of the 2008 IEEE International Conference on Systems, Man and Cybernetics*. October 12-15. 2008. 2584–2590.
70. Matheus, K. and Dollar, A. M. Benchmarking Grasping and Manipulation: Properties of the Objects of Daily Living. *Proceedings of the 2010 IEEE/RSJ International Conference on Intelligent Robots and Systems*. October 18-22. 2010. 5020–5027.
71. Dovat, L., Lamercy, O., Gassert, R., Maeder, T., Milner, T., Leong, T. C. and Burdet, E. HandCARE: A Cable-Actuated Rehabilitation System to Train Hand Function After Stroke. *IEEE Trans. Neural Syst. Rehabil. Eng.*, 2008, 16(6): 582–591.
72. Chiri, A., Vitiello, N., Giovacchini, F., Roccella, S., Vecchi, F. and Carrozza, M. C. Mechatronic Design and Characterization of the Index Finger Module of A Hand Exoskeleton for Post-Stroke Rehabilitation. *IEEE/ASME Trans. Mechatronics*, 2012, 17(5): 884–894.
73. Wu, J., Huang, J., Wang, Y. and Xing, K. A Wearable Rehabilitation Robotic Hand Driven by PM-TS Actuators. *Lect. Notes Comput. Sci. (including Subser. Lect. Notes Artif. Intell. Lect. Notes Bioinformatics)*. 2010. 440–450.
74. Tadano, K., Akai, M., Kadota, K. and Kawashima, K. Development of Grip Amplified Glove Using Bi-Articular Mechanism with Pneumatic Artificial Rubber Muscle. *Proceedings of the 2010 IEEE International Conference on Robotics and Automation*. May 3-7. 2010. 2363–2368.
75. Wilkening, A., Baiden, D. and Ivlev, O. Assistive Control of Motion Therapy Devices Based on Pneumatic Soft-Actuators with Rotary Elastic Chambers. *Proceedings of the 2011 IEEE International Conference on Rehabilitation Robotics*. June 29- July 1. 2011. 1–6.
76. Sasaki, D., Noritsugu, T. and Takaiwa, M. Development of Active Support Splint Driven by Pneumatic Soft Actuator (ASSIST). *Proceedings of the*

- 2005 *IEEE International Conference on Robotics and Automation*. April 18-22. 2005. 520–525.
77. Shepherd, R. F., Ilievski, F., Choi, W., Morin, S. A., Stokes, A. A., Mazzeo, A. D., Chen, X., Wang, M. and Whitesides, G. M. Multigait Soft Robot. *Proc. Natl. Acad. Sci. U.S.A.*, 2011, 108: 20400-20403.
 78. Laschi, C., Cianchetti, M., Mazzolai, B., Margheri, L., Follador, M., and Dario, P. Soft Robot Arm Inspired by the Octopus. *Adv. Robot.*, 2012, 26(7): 709–727.
 79. Kumagai, I., Wakimoto, S. and Suzumori, K. Development of Large Intestine Endoscope Changing Its Stiffness. *Proceedings of the 2009 IEEE International Conference on Robotics and Biomimetics*. December 19-23. 2009. 2320–2325.
 80. Trivedi, D., Rahn, C. D., Kier, W. M. and Walker, I. D. Soft Robotics: Biological Inspiration, State of The Art, and Future Research. *Appl. Bionics Biomech.*, 2008, 5(3): 99–117.
 81. Suzumori, K., Iikura, S. and Tanaka, H. Flexible Microactuator for Miniature. *Proceedings of the 1991 IEEE Micro Electro Mechanical Systems, An Investigation of Micro Structures, Sensors, Actuators, Machines and Robots*. January 30- February 2. 1991. 204–209.
 82. Zhao, F., Dohta, S., Akagi, T. and Matsushita, H. Development of a Bending Actuator using a Rubber Artificial Muscle and its Application to a Robot Hand. *Proceedings of the 2006 SICE-ICASE International Joint Conference*. October 18-21. 2006. 381–384.
 83. Gupta, A. and O'Malley, M. K. Design of a Haptic Arm Exoskeleton for Training and Rehabilitation. *IEEE/ASME Transactions on Mechatronics*, 2006, 11(3): 280–289.
 84. Caldwell, D. G., Tsagarakis, N. G., Kousidou, S., Costa, N. and Sarakoglou I. 'Soft' Exoskeletons for Upper and Lower Body Rehabilitation — Design, Control and Testing. *Int. J. Humanoid Robot.*, 2007, 4(3): 549–573.
 85. Shepherd, R. F., Stokes, A. A., Freake, J., Barber, J., Snyder, P. W., Mazzeo, A. D., Cademartiri, L., Morin, S. A. and Whitesides, G. M. Using Explosions to Power a Soft Robot. *Angew. Chem. Int. Ed. Engl.*, 2013, 52(10): 2892–2896.

86. Suzumori, K., Maeda, T., Watanabe, H. and Hisada, T. Fiberless Flexible Microactuator Designed by Finite-Element Method. *IEEE/ASME Transactions on Mechatronics*, 1997, 2(4): 281–286.
87. Suzumori, K. and Tanaka, H. Applying a Flexible Microactuator to Robotic Mechanisms. *IEEE Control Systems*, 1992, 12(1): 21-27.
88. Ogura, K., Wakimoto, S., Suzumori, K. and Nishioka, Y. Micro Pneumatic Curling Actuator— Nematode Actuator. *Proceedings of the 2008 IEEE International Conference on Robotics and Biomimetics*. February 21-26. 2009. 462–467.
89. Ilievski, F., Mazzeo, A. D., Shepherd, R. F., Chen, X. and Whitesides, G. M. Soft Robotics for Chemists. *Angew. Chemie*, 2011, 123(8): 1930–1935.
90. Martinez, R. V., Fish, C. R., Chen, X. and Whitesides, G. M. Elastomeric Origami: Programmable Paper-Elastomer Composites as Pneumatic Actuators. *Adv. Funct. Mater.*, 2012, 22(7): 1376–1384.
91. Zhao, F., Dohta, S. and Akagi, T. Development and Analysis of Bending Actuator Using McKibben Artificial Muscle. *J. Syst. Des. Dyn.*, 2012, 6(2): 158–169.
92. Suzumori, K., Iikura, S. and Tanaka, H. Development of Flexible Microactuator and Its Applications to Robotic Mechanisms. *Proceedings of the 1991 IEEE International Conference on Robotics and Automation*. April 9-11. 1991. 1622–1627.
93. Martinez, R. V., Branch, J. L., Fish, C. R., Jin, L., Shepherd, R. F., Nunes, R. M. D., Suo, Z. and Whitesides, G. M. Robotic Tentacles with Three-Dimensional Mobility Based on Flexible Elastomers. *Adv. Mater.*, 2013, 25(2): 205–212.
94. Wang, J., Li, J., Zhang, Y. and Wang, S. Design of an Exoskeleton for Index Finger Rehabilitation. *Proceedings of the 2009 Annual International Conference of the IEEE Engineering in Medicine and Biology Society*. September 3-6. 2009. 5957–5960.
95. Tsagarakis, N., Caldwell, D. G. and Medrano-Cerda, G. A. A. 7 DOF Pneumatic Muscle Actuator (pMA) Powered Exoskeleton. *Proceedings of the 1999 8th IEEE International Workshop on Robot and Human Interaction*. 1999. 327–333.

96. Gopura R. A. R. C. and Kiguchi, K. Mechanical Designs of Active Upper-Limb Exoskeleton Robots. *Proceedings of the 2009 IEEE International Conference on Rehabilitation Robotics*. June 23-26. 2009. 178–187.
97. Poli, P., Morone, G., Rosati, G. and Masiero, S. Robotic Technologies and Rehabilitation: New Tools for Stroke Patients' Therapy. *Biomed Res. Int.*, 2013, 2013: 1–8.
98. Choi, H. R., Ryew, S. M., Jung, K. M., Kim, H. M., Jeon, J. W., Nam, J. D., Maeda, R. and Tanie, K.. Micro Robot Actuated by Soft Actuators Based on Dielectric Elastomer. *Proceedings of the 2002 IEEE/RSJ International Conference on Intelligent Robots and Systems*. September 30- October 4. 2002. 1730–1735.
99. Spinks, G. M., Mottaghitalab, V., Bahrami-Samani, M., Whitten, P. G. and Wallace, G. G. Carbon-Nanotube-Reinforced Polyaniline Fibers for High-Strength Artificial Muscles. *Adv. Mater.*, 2006, 18(5): 637–640.
100. Rahman, S. M. M. A Novel Variable Impedance Compact Compliant Series Elastic Actuator: Analysis of Design, Dynamics, Materials and Manufacturing. *Appl. Mech. Mater.*, 2012, 245: 99–106.
101. Mazzolai, B., Margheri, L., Cianchetti, M., Dario, P. and Laschi, C. Soft Robotic Arm Inspired by the Octopus: II. From Artificial Requirements to Innovative Technological Solutions. *Bioinspiration & Biomimetics*, 2012, 7(2): 1-14.
102. Chuc, N. H., Park, J. K., Huu, N., Vuong, L., Kim, D., Koo, J. C., Lee, Y., Nam, J., Choi, H. R and Vuong, N. H. L. Multi-Jointed Robot Finger Driven by Artificial Muscle Actuator. *Proceedings of the 2009 IEEE International Conference on Robotics and Automation*. May 12-17. 2009. 587–592.
103. Bundhoo, V. and Park, E. J. Design of an Artificial Muscle Actuated Finger Towards Biomimetic Prosthetic Hands. *Proceedings of the 2005 12th International Conference on Advanced Robotics*. July 18-20. 2005. 368–375.
104. Peters, M., Mackenzie, K. and Bryden, P. Finger Length and Distal Finger Extent Patterns in Humans. *Am. J. Phys. Anthropol.*, 2002, 117(3): 209–217.
105. Suzumori, K. Optimal Design of Bending Pneumatic Rubber Actuator Based on Non-linear Finite Element Analysis. *12th IFToMM World Congress*. 2007.
106. Schulte, H. F. The Characteristics of the McKibben Artificial Muscle. *Appl. Extern. Power Prosthet. Orthotics*, 1961, 874: 94–115.

107. Konishi, S., Nokata, M., Jeong, O. C., Kusuda, S., Sakakibara, T., Kuwayama, M. and Tsutsumi, H. Pneumatic Micro Hand and Miniaturized Parallel Link Robot for Micro Manipulation Robot System. *Proceedings of the 2006 IEEE International Conference on Robotics and Automation*. May 15-19. 2006. 1036–1041.
108. Caldwell, D. G., Medrano-Cerda, G. A., and Goodwin, M. Control of Pneumatic Muscle Actuators. *IEEE Control Syst. Mag.*, 1995, 15(1): 40–48.
109. Widia, M. and Dawal, S. Z. Investigation on Upperlimb Muscle Activity and Grip Strength During Drilling Task. *Proceedings of the 2010 International MultiConference of Engineers and Computer Scientists*. March 17-19. 2010. 1-5.
110. Sidek, S. N. and Mohideen, A. J. H. Mapping of EMG Signal to Hand Grip Force at Varying Wrist Angles. *Proceedings of the 2012 IEEE EMBS Conference on Biomedical Engineering and Sciences (IECBES)*. December 17-19. 2012. 648–653.
111. Heck, S., Zilleken, C., Pennig, D. and Koslowsky, T. C. Reconstruction of Radial Capitellar Fractures using Fine-Threaded Implants (FFS). *Injury*, 2012, 43(2): 164–168.
112. Krause, K. E., McIntosh, E. I. and Vallis, L. A. Sarcopenia and Predictors of the Fat Free Mass Index in Community-Dwelling and Assisted-Living Older Men and Women. *Gait Posture*, 2012, 35(2): 180–185.
113. Taniguchi, H., Toyooka, H., Minatohara, T., Fukuda, M. and Zhao, F. Pneumatic Rubber Actuators for a Rehabilitation Device to Prevent Contractures of Finger Joints. *Proceedings of the 2012 19th International Conference on Mechatronics and Machine Vision in Practice*. November 28-30. 2012. 420–424.