

PARTICLE SIZE MEASUREMENT USING ELECTROSTATIC SENSOR
THROUGH SPATIAL FILTERING METHOD

TEIMOUR TAJDARI

A thesis submitted in fulfilment of the
requirements for the award of the degree of
Doctor of Philosophy (Electrical Engineering)

Faculty of Electrical Engineering
Universiti Teknologi Malaysia

MARCH 2014

This thesis is dedicated to my respected parents,
Yarmohammad Tajdari and Malekkhatoon Paydar,
and to my beloved wife,
Fatemeh Tajdari

ACKNOWLEDGEMENT

I would like to thank my supervisor, Prof. Dr. Mohd Fua'ad bin Hj. Rahmat, for his inspirational guidance, selfless service, patience and encouragement, without which the study could not have been possible. His blend of character has been unique and worth emulating.

I also wish to thank my entire family members, my father-in-law and my mother-in-law for their understanding and encouragement during my PhD program. Without such support, my mind could not have been at peace to allow me to concentrate on and carry out the PhD research that has led to this thesis.

My sincere appreciation also goes to everyone whom I may not have mentioned above, including all those who have helped directly or indirectly in the completion of my project.

ABSTRACT

Particle size measurement is important in powder and particle industries in which the particle size affects the productivity and efficiency of the machine, for example, in coal-fired power plants. An electrostatic sensor detects the electric charge from dry particles moving in a pipeline. Analysis of the detected signal can provide useful information about the particle velocity, mass flow rate, concentration and size. Using electrostatic sensors, previous researches studied particle sizing using magnitude dependent analysis which is a highly conditional method where the results can be affected by other parameters such as particle mass flow rate, velocity and concentration. This research proposes a magnitude independent analysis for particle sizing in the frequency domain called spatial filtering method. The solution was started by modeling and analysis of the charge induced to the ring electrode using finite-element analysis to find the sensitivity of electrode. A mathematical model was provided to compute particle position on the radial axis of the electrode and then a new technique was proposed to extract a single particle size from the calculated particle radial position. To validate the proposed method experimentally, a sensor was designed and five test particles ranging from 4 mm to 14 mm were selected for measurement. The results show a 0.44 mm estimation error between the estimated and expected results. The results also show that the method is promising for the establishment of a reliable and cost-effective solid particle sizing system.

ABSTRAK

Pengukuran saiz zarah memainkan peranan yang penting dalam industri serbuk dan zarah, di mana ianya boleh mempengaruhi tahap produktiviti dan kecekapan mesin, sebagai contoh di loji kuasa arang batu. Sensor elektrostatik mengesan cas elektrik dari zarah kering yang bergerak dalam saluran paip. Analisis terhadap isyarat yang dikesan memberikan maklumat yang berguna mengenai halaju, kadar aliran jisim, kepekatan dan saiz zarah. Dengan menggunakan sensor elektrostatik, penyelidikan terdahulu telah membuat kajian terhadap saiz zarah dengan menggunakan analisis magnitud bersandar iaitu satu kaedah yang sangat bersyarat di mana keputusannya boleh dipengaruhi oleh parameter lain seperti kadar aliran jisim, halaju dan kepekatan zarah. Kajian ini mencadangkan satu analisis berdasarkan magnitud bebas untuk mengukur saiz zarah dalam frekuensi domain, iaitu kaedah ruang penapisan. Penyelesaiannya bermula dengan pemodelan dan analisis terhadap cas yang teraruh pada elektrod bentuk cincin dengan menggunakan analisis unsur-terhingga untuk mencari sensitiviti elektrod. Model matematik telah diterbitkan untuk mengira kedudukan zarah pada paksi jejarian elektrod dan satu teknik baru telah dicadangkan untuk mendapatkan saiz zarah tunggal dengan berpandukan kedudukan jejari zarah yang dikira. Untuk mengesahkan kaedah yang telah dicadangkan secara uji kaji, sensor telah direka dan lima ujian yang terdiri daripada zarah yang bersaiz dari 4 mm hingga 14 mm telah dipilih untuk pengukuran. Keputusan menunjukkan ralat anggaran di antara keputusan yang dianggarkan dan diharapkan adalah sebanyak 0.44 mm. Keputusan juga menunjukkan bahawa kaedah ini dapat menghasilkan sistem saiz zarah yang boleh dipercayai dan kos efektif.

TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	DEDICATION	iii
	ACKNOWLEDGEMENTS	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	xi
	LIST OF FIGURES	xii
	LIST OF ABBREVIATIONS	xv
	LIST OF SYMBOLS	xvi
	LIST OF APPENDICES	xx
1	INTRODUCTION	1
	1.1 Motivation and Introduction to Electrostatic Sensor	1
	1.2 Electrostatic Sensor	2
	1.3 Research Background	3
	1.3.1 Mass Flow Rate Measurement	4
	1.3.2 Velocity Measurement	5
	1.3.3 Process Tomography System	6
	1.3.4 Miscellaneous Applications	7
	1.4 Problem Statement	9
	1.5 Research Objectives	10
	1.6 Research Scope and Limitations	11

1.7	Research Contributions	12
1.8	Thesis Outline	12
1.9	Summary of the Introduction	14
2	LITERATURE REVIEW	15
2.1	Introduction to Literature Review	15
2.2	Particle Sizing Methods	16
2.2.1	Physical Methods	16
2.2.2	Laser Diffraction	17
2.2.3	Imaging Method	20
2.2.4	Electrical Methods	21
	2.2.4.1 Differential Mobility Analyzer	21
	2.2.4.2 Coulter Count Method	24
	2.2.4.3 Electrostatic Sensor	26
2.2.5	Miscellaneous Methods	27
2.3	Electrostatic measurement	30
2.3.1	Electrostatic Charge Measurement Methods	31
	2.3.1.1 Faraday Pail	31
	2.3.1.2 Electrostatic Sensor	33
2.3.2	Electrostatic Sensor Design	34
	2.3.2.1 Electrode Design	35
	2.3.2.2 Signal Conditioning Circuit	36
2.3.3	Mathematical Modeling of the Electrostatic Sensor	40
2.3.4	Sensitivity of the Electrode	45
2.3.5	Spatial Filtering Effect	48
	2.3.5.1 Spatial Filtering Method	52
	2.3.5.2 Velocity Measurement Using Spatial Filtering Method	52
2.4	Summary of the Literature Review	54
3	RESEARCH METHODOLOGY	56
3.1	Introduction to Methodology	56
3.2	Electrostatic Sensor Design	57

3.2.1	Electrode Design	57
3.2.2	Electrode Spatial Sensitivity	58
3.2.2.1	Induced Charge to the Electrode	59
3.2.2.2	Finite Element Modeling	60
3.2.2.3	Spatial Filtering Effect of the Electrode	63
3.2.3	Frequency Response of the Electrode	64
3.2.4	Frequency Response of the Electrostatic Sensor	67
3.2.4.1	Process to Model the Frequency Spectrum	69
3.2.5	Signal Conditioning Circuit	70
3.2.6	Hardware Setup	75
3.3	Particle Size Measurement	75
3.3.1	Spatial Filtering Method for Particle Size Measurement	76
3.4	Summary of the Methodology	82
4	RESULTS AND ANALYSIS	83
4.1	Introduction to Results and Analysis	83
4.2	Electrode Design	84
4.2.1	Electric Charge Induction to Electrode	85
4.2.2	The Electrode's Spatial Sensitivity	87
4.2.2.1	Axial Axis Sensitivity	87
4.2.2.2	Radial Axis Sensitivity	90
4.2.2.3	The Electrode Spatial Filtering Effect	91
4.2.3	The Electrode Frequency Response	91
4.3	The Electrostatic Sensor Frequency response	93
4.4	Signal Conditioning Circuit	95
4.4.1	Intrinsic Noise Analysis	97
4.4.2	Extrinsic Noise	98
4.5	Experimental Setup	101
4.6	Particle Size Measurement	105
4.7	Summary of Results and Analysis	113

5	DISCUSSION AND FUTURE WORK	115
	5.1 Introduction	115
	5.2 Research Contributions	116
	5.3 Research Limitations	117
	5.4 Future Work	118
	REFERENCES	119
	Appendices A-D	131-154

LIST OF TABLES

TABLE NO.	TITLE	PAGE
2.1	Particle sizing methods and application areas	30
2.2	Important studies in modeling and application of spatial filtering effect	54
4.1	The best fit equation's constants and the SEE for three curves	89
4.2	The constants a, b, c and d values for axial sensitivity curves in 17 points along the electrode radius	107
4.3	Particle size versus CS frequency (from experimental tests)	111

LIST OF FIGURES

FIGURE NO.	TITLE	PAGE
1.1	Application and research areas for the electrostatic sensor	8
2.1	Basic Laser Diffraction hardware setup	18
2.2	Image based particle sizing procedure	20
2.3	Schematic of Differential Mobility Method	22
2.4	Coulter Count or Electric Sensing Zone basic set up	24
2.5	A micro Coulter Counter by Richards et al. (2012)	26
2.6	Schematic of basic ultrasonic measurement setup	28
2.7	Schematic of the resonant structure	29
2.8	Schematic diagram of the basic Faraday Pail	32
2.9	Schematic diagram of the basic Electrostatic Sensor setup	33
2.10	Schematic representations of conventional electrostatic senor electrodes	35
2.11	Electrostatic noise collector using a capacitor	37
2.12	Electrostatic charge collector using resistor	38
2.13	inverting amplifier	39
2.14	Schematic of the ring electrode	40
2.15	Schematic representation of electrostatic sensor head as a combination of three parts	41
2.16	Equivalent circuit diagram of the electrostatic sensor head	42
2.17	A ring electrode sensitivity on z axis	46
2.18	Sensitivity of a circular palate electrodes	47
2.19	The circuit model of electrostatic sensor	51
3.1	Specification of the selected ring electrode	57
3.2	The procedure of performing simulation in Ansoft Maxwell	61

3.3	Ansoft Maxwell interface and the schematic of the ring electrode	62
3.4	Schematic of ring electrode in cylindrical coordinate system	64
3.5	Equivalent circuit diagram of the electrostatic sensor	67
3.6	Current-to-voltage converter used as a pre-amplifier circuit	71
3.7	Pre-amplifier circuit diagram	72
3.8	Signal conditioning circuit diagram includes pre-amplifier and amplifier	73
3.9	Hardware setup for particle size measurement using electrostatic sensor	75
3.10	Schematic representation of particle position inside electrode head	78
3.11	Particles fall tangent with pipe wall get different radial position	80
3.12	schematic representation of particle dropping considerations	81
4.1	Electrode and pipe cross-section	84
4.2	Electric charge density distributions on the ring surface	86
4.3	Axial sensitivity of the electrode at $r = 0$	88
4.4	The axial sensitivity at three different radial positions	89
4.5	Radial sensitivity of the electrode at $z = 0$	90
4.6	Frequency properties of the designed electrode	93
4.7	frequency properties of sensor output signal	95
4.8	Schematic representation of signal conditioning circuit	95
4.9	The Gain versus frequency in the signal conditioning circuit	97
4.10	The curve indicates the output noise spectrum	98
4.11	Output signal in the presence of 50Hz extrinsic noise	99
4.12	A Dual voltage power supply circuit using ICL7660	99
4.13	The electrostatic sensor	100
4.14	The noise-less output signal from the designed sensor	100
4.15	The setup for experimental tests	101
4.16	Comparison of the induced current to the electrode	103
4.17	Normalized frequency spectrum of the sensor output	105
4.18	Relation between the radial position r and the constant d	108
4.19	Mathematical estimation of particle radial	

	position (r) versus f_{cs}	109
4.20	Sensor output in frequency domain (top) and time domain (bellow) for a 10mm diameter test particle	110
4.21	Estimated results and actual results	111
4.22	Estimated results and actual results (proportionality k is implemented)	113

LIST OF ABBREVIATIONS

PC	-	Personal computer
AC	-	Alternating Current
DC	-	Direct Current
2D	-	Two Dimensions
3D	-	Three Dimension
PSD	-	Power Spectrum Density
CCD	-	Charged-Coupled Device
DMA	-	Differential Mobility Analyzer
FEM	-	Finite Element Modeling
FFT	-	Fast Fourier Transform
CS frequency	-	Frequency at the Crest of Spectrum
HV	-	High Voltage
PCB	-	Printed Circuit Board
ESZ	-	Electrical Sensing Zone
OD	-	Outside Diameter
ID	-	Inside Diameter
CAD	-	Computer-Aided Design
DAQ	-	Data Acquisition
SEE	-	Standard Error of Estimation
RMSE	-	Root Mean Square Error
FET	-	Field-Effect Transistor
HHT	-	Hilbert-Huang Transform

LIST OF SYMBOLS

t	-	Time
m	-	meter
mm	-	millimeter
kg	-	kilogram
k	-	kilo (1000), Proportionality Constant
M	-	Mega (10^6)
μ	-	micron (10^{-6})
n	-	nano (10^{-9}), Number of Electric Charges
p	-	pico (10^{-12})
f	-	Frequency
Hz	-	Hertz
d_p	-	Particle Diameter
e	-	Electric Charge Unit
C_c	-	Cunningham Slip Correction
λ	-	Mean Free Pass, Wave Length
r_{inner}	-	Radius of the Inner Electrode
r_{outer}	-	Radius of the Outer Electrode
L	-	Length of the Electrode
x	-	x axis, Wave Distance to Transmitter
α	-	Attenuation Coefficient, Constant
A_x	-	Wave Amplitude in Receiver
A_0	-	Wave Amplitude in Transmitter
q	-	Electric Charge on Particle, Induced Charge to Electrode
Q	-	Electric Charge on Particle, Induced Charge to Electrode
q'	-	Induced Charge to Electrode

C	-	Capacitance, Coulomb (Electric Charge Unit)
s	-	Second, Laplace Notation, Surface Notation
A	-	Ampere, Constant
R	-	Resistance
R_f	-	Feedback Resistance
C_f	-	Feedback Capacitance
F	-	Farad
I_s	-	Electric Current
W	-	Axial Length
θ	-	A selected Angle
ϕ	-	Electric Potential
U	-	Output Voltage
τ	-	Time Constant
σ	-	Charge Density
∇	-	Del Operator
ρ_v	-	Volume Charge Density
ρ	-	Volume Charge Density
Γ_s	-	Boundary Conditions for Metal Screen
Γ_e	-	Boundary Conditions for Electrode
Γ_p	-	Boundary Conditions for Pipe
r_0	-	Radius of the Particle, Initial Radial Distance to Rotation Axis
r	-	Distance from Center of the Particle, Distance from Center of the Ring electrode, Particle Radial Position
ε_0	-	Electric Permittivity for Free Space
f_{cs}	-	Frequency at the Crest of the Frequency Spectrum
D	-	Particle Diameter, Electrode Diameter
η	-	Fluid Viscosity, Air Viscosity
ρ_s	-	Density of the Solid
ρ_f	-	Density of the Liquid
g	-	Gravity Acceleration, Constant

V	-	Voltage
v	-	velocity
ω	-	Angular frequency
ϖ	-	Angular frequency
r_f	-	Final Radial Distance to Rotation Axis
Z_p	-	Electrical Mobility
E	-	Electric Field
Q_s	-	Air Flow for Aerosol-Free Air
Q_a	-	Air Flow for Positively Charged Aerosols
$S(x)$	-	Electrode Sensitivity in x axis
$S(z)$	-	Electrode Sensitivity in z axis
$S(r)$	-	Electrode Sensitivity in r axis
$i(x)$	-	Electrostatic Noise
$h(t)$	-	Impulse Response
$H(f)$	-	Frequency Response
$P(\varpi)$	-	Sensor Frequency Response
$T(\varpi)$	-	Sensor Frequency Response
$H_r(\varpi)$	-	Electrode Frequency Response
b	-	Axial Length
f_c	-	Cutoff frequency
C_d	-	Capacitance between Particle and Electrode
C_e	-	Capacitance between Particle and Electrode
r_d	-	Input Resistor of the Pre-Amplifier
R_a	-	Input Resistor of the Pre-Amplifier
C_n	-	Capacitance between Electrode and Ground
r_n	-	Resistance between Electrode and Ground
r_1	-	Gain adjusting Resistor of the Amplifier
r_2	-	Gain adjusting Resistor of the Amplifier
a	-	Sensitivity Equation's Constant
b	-	Sensitivity Equation's Constant

c	-	Sensitivity Equation's Constant
d	-	Sensitivity Equation's Constant
f_{\max}	-	Frequency at the Crest of the Frequency Spectrum
v_m	-	Velocity
\vec{D}	-	Electric Flux
ϵ_r	-	Relative Permittivity
C_a	-	Input Capacitance of Pre-Amplifier
r_p	-	Pipe Radius
r_r	-	Ring Electrode Radius

LIST OF APPENDICES

APPENDIX	TITLE	PAGE
A	The PCB design of the signal conditioning circuit	126
B	The electrode axial sensitivity in 17 radial position	127-136
C	The experimental test results to measure the f_{cs}	137-147
D	List of Publications	148-149

CHAPTER 1

INTRODUCTION

1.1 Motivation and Introduction to Electrostatic Sensor

One of the interesting areas to develop a reliable and cost-effective instrumentation in particle and powder industries have been designing the measurement systems based on electrostatic sensors, which has led to a great deal of research and development on the sensor application. Particle movement inside a pipe or a conveyor produces a small amount of electric charge on the particle surface due to particle-to-particle and particle-to-pipe wall friction and collision. This charge on the particle surface can be detected using an electrostatic sensor. The magnitude and frequency component of the detected signal depends on the physical characteristics of the particle and its dense flow parameters such as velocity, concentration and mass flow rate. The particles flow information can be extracted using suitable signal processing algorithms from the electrostatic sensor output signal. Using the electrostatic sensor the velocity of the moving particle can be measured using the cross-correlation technique (Xu *et al.*, 2010b) or the spatial filtering method (Xu *et al.*, 2013). Particles' mass flow rate measurement in direct method (Gajewski, 1999a), concentration profile-map utilizing process tomography method (Rahmat *et al.*, 2009c), and particle flow dense mean-size measurement (Zhang and Yan, 2003) are some other studies that employed the electrostatic sensor capabilities.

This research deals with the electrostatic sensor application to measure the mean-size of a single particle. To the sensor, a particle seems as a point charge and particle size does not mean anything to the sensor. A new technique is proposed and studied in detail to acquire the particle size information utilizing electrostatic sensor.

1.2 Electrostatic Sensor

Electrostatic sensor (electrodynamic sensor) consists of two main parts which are the sensor electrode and the signal conditioning circuit. The electrode is a conductive metal that can detect the electrostatic flow noise from a moving charged particle. The charge induced to the electrode needs to be collected and amplified using a suitable signal conditioning circuit to an acceptable level. Then the output signal from the sensor can be sent to a PC using a data-transfer card for signal visualization or further analysis.

Geometrical properties of the electrode dramatically affect the output signal magnitude and its frequency. Depending on their area of application, the electrodes are designed in different shapes including ring electrode, pin electrode and plate electrode. The electrode can be installed using either the intrusive or non-intrusive method. In the intrusive method, the electrode will be installed inside the conveyor in which it is in direct contact with the particle flow, following the same method as that employed by Rahmat and Lee (2004) in their study. In the non-intrusive method, the electrode will be installed in the pipe circumference (Gajewski, 1999b).

A typical signal conditioning circuit for electrostatic sensor consists of a signal collector (pre-amplifier) and a signal amplifier. In some applications such as process tomography, an AC-to-DC converter is added to the circuit to convert the output signal to its equivalent DC level. The signal conditioning circuit deals with a random and a very small range of electric charge fluctuations. Due to the high level of amplification, the sensor is very susceptible to detect the noise from external electromagnetic sources.

Sensitivity of the electrostatic sensor describes the ability of the electrode in detecting the electric field strength at any point within the space. It can usually be described as the ratio of the charge induced to the electrode over the total charge existing in the source (Xu *et al.*, 2007). The sensitivity of the sensor is highly dependent upon the physical properties of the electrode, especially the size and physical geometry of the electrode. In addition, it is a function of the electric charge location in the coordinate system. When the electrode sensitivity is considered

against particle location within the space, it is usually called the electrode spatial sensitivity. The electrode sensitivity can describe the electrode spatial filtering effect.

Spatial filtering effect of the sensor describes the frequency bandwidth properties and the frequency response of the sensor electrode. It has a direct relation with the size and shape of the electrode. In fact, a moving particle remains for a longer time in the electrode detection area with longer axial length. As a result, the induction to a bigger size electrode happens in smaller frequency and vice versa.

In designing a measurement system using an electrostatic sensor, both sensitivity and spatial filtering effect of the electrode are important to be known and analyzed. Information regarding the electrode sensitivity and frequency response is needed in making decisions about the amplifier gain and the design of the noise reduction circuit.

1.3 Research Background

In the industries that deal with particle conveyors, the measurement of particle flow parameters is important to control the machine productivity and efficiency. These parameters can be the amount charge on particles, velocity, mass flow rate, particle size, and humidity. For instance, in coal-fired power stations the particle size, flow velocity and mass flow rate of pulverized fuel are important to be measured and monitored in order to control the combustion quality. This control will result in better productivity of the burner and to higher efficiency in energy consumption.

Electric charge in a particle is a single parameter that can provide a great amount of information regarding the particle behavior when it is moving in a pipeline. Due to its simple and robust structure, electrostatic sensor is a good candidate for use in the dusty and harsh environment of such industries. This section is dedicated to a brief background about the electrostatic sensor application in two-phase flow measurement systems.

1.3.1 Mass Flow Rate Measurement

In powder and granule industry, mass flow rate measurement of moving particles in a pneumatic conveyor is one of the important parameters that should be measured and controlled. The amount of electric charge carried by particles in a pipeline has direct relation with the mass flow rate. Detecting and analyzing the electrostatic noise will give decent information about particle mass flow rate. The particle mass flow rate in the pneumatic conveyor can be measured in two methods, namely direct method and inferential method (indirect method).

As explained by Yan (1996) and later by Zheng and Liu (2010), in an inferential method, mass flow rate at any time is proportional to the product of instantaneous velocity and instantaneous particle volumetric concentration in the pipe cross-section. In this method, electrostatic sensor can be hired either to measure the particle velocity or to measure volumetric concentration or even both. Green *et al.* (1997a) have employed an electrostatic sensor to find both the volumetric concentration using process tomography and the velocity using cross correlation technique. Then the mass flow rate map of the pipe cross section is given by multiplying the concentration profile and the velocity for each pixel. Carter and Yong (2005) have measured the mass flow rate utilizing the indirect method where the electrostatic sensor is used to measure the velocity using cross correlation technique, and the volumetric concentration is found by digital imaging technique.

As described by Zheng and Liu (2011), in the direct method, the sensing element is compared directly with the mass flow rate. The particle mass flow rate can be compared directly with the averaged output signal level. Gajewski *et al.* (1993a) and later Gajewski (1996b) invented and developed a measurement system based on a model in which the average output voltage of the electrostatic sensor is a function of velocity and mass flow rate. If the velocity is known, the variation of output voltage of the system directly follows the variation on mass flow rate. However, when the velocity is unknown, the relation between mass flow rate and velocity with the signal quantified characteristics are complex and nonlinear. Lijun *et al.* (2005) used a novel approach by training a back-propagation neural network to establish the

relation between signal characteristics and mass flow rate and the velocity of the particles that worked with measurement error of 20%.

The direct method obviously has a simpler measurement setup than the indirect method. However, when the velocity is an unknown variable, the system shows a large measurement error. In addition, measuring the velocity using other methods eliminates the simplicity advantage of the system. The inferential method is more complex than the direct method; but it can provide useful information about mass flow rate, velocity and volumetric concentration simultaneously.

1.3.2 Velocity Measurement

Electrostatic sensor application in particle velocity measurement is the most researched and developed area for this sensor. The reason is that the velocity variation has a significant effect on the sensor output signal components both in time domain and frequency domain. There are two main techniques that use the electrostatic sensor for velocity measurement. These are the cross-correlation technique and the spatial filtering method.

Cross-correlation technique uses two identical electrodes on the pipe, aligned and installed in one line along with the flow direction over a distance from each other. The electrodes are called upstream and downstream electrodes, respectively. The cross-correlation of the sensor output gives the transit time that takes the particle to pass the distance between these two electrodes (Yan and Ma, 2000). Velocity can be easily calculated when the distance between the electrodes is known, and the particle transit time is measured. The very early application of this technique was incorporated by Gajewski *et al.* (1990) and Gajewski *et al.* (1993b) and later Gajewski (1994) for velocity measurement instrumentation in which the electrostatic sensor was used with the ring electrodes. A commercial prototype of a velocity measurement system using the cross-correlation technique was designed by Ma and Yan (2000), and the instrument performance was evaluated with different shapes of non-intrusive electrodes. The prototype is tested on the pneumatic particle conveyor

which showed a response time less than 2.5 s and repeatability better than $\pm 2\%$. The cross-correlation technique employing electrostatic sensors was used by Yan *et al.* (2010) and later by Rodrigues and Yong (2012) for strip and cable speed measurement, which is applicable in electrical cable and fiber-optic cable industries. To achieve a better accuracy, Xiangchen and Yong (2012) and Qian *et al.* (2012) instead of applying two electrodes, used an array of electrostatic sensors. In this method, the output signals from every two adjacent electrodes cross correlated, and the output resulted from a data fusion algorithm.

Spatial filtering method relates the frequency components of the electrostatic sensor output signal to the particle velocity. Hammer and Green (1983) showed that the velocity of the particles passing through a capacitive sensor, which is functionally similar to the electrostatic sensor, has direct relation with the frequency component of the output signal. Yan *et al.* (1995) and Gajewski (1996a) showed that the same relation exists for the electrostatic sensor. Zhang (2002) proposed a mathematical model that showed the velocity of a single particle passing through an electrostatic sensor has direct relation with the frequency at peak of the signal power spectrum density or PSD. In higher velocities, the particle induces the electrostatic noise to the electrode with higher frequency. The method was then developed by Xu *et al.* (2008) and Xu *et al.* (2009) to be applied to measure particle dense flow velocity in a pneumatic conveyor with particle concentration of $0.067-0.130m^3m^{-3}$. The advantage of this method was its simplicity due to using a single electrode. However, this method has a broad spectral bandwidth that reduces the frequency reading accuracy. Xu *et al.* (2012b) and Li *et al.* (2012) proposed a new method based on spatial filtering technique using two sensor arrays. The sensor arrays together, along with using a differential amplifier showed a narrow spectral bandwidth indicating that the frequency at the peak of the PSD is directly related to particle flow velocity.

1.3.3 Process Tomography System

Process tomography is an imaging technique that uses an array of sensors around the pipe circumference, and the images produced provide 2D or 3D view of

the flow parameters inside the pipe. There are different types of devices that can be used for process tomography purpose, such as capacitive sensor, optic sensor, ultrasonic sensor, CCD camera, and electrostatic sensor. Due to its non-intrusive nature and simple structure, the electrostatic sensor has been a suitable candidate in process tomography.

Green *et al.* (1997b) used 16 electrostatic sensors to calculate the concentration of the particle flow in a pneumatic pipeline. The back projection algorithm was utilized to provide a 2D image from the measured sensitivity map. The same technique was applied by Rahmat and Rahiman (2001) using two arrays of 16 electrostatic sensors to find a 3D velocity profile of moving particles. Machida and Scarlett (2005) discussed that the back projection algorithm is unable to distinguish between two adjacent point charges in the detection area. The Least-Square algorithm with the combination of the back projection algorithm is used to get a clearer image.

The researches on process tomography using electrostatic sensor follow almost the same hardware setup, and the differences come from the number of electrodes and the type of image reconstruction algorithm. For instance, neural network training has been used to find the type of the flow pattern on the pipe cross section (Rahmat and Sabit, 2004, 2007) or similarly, fuzzy logic algorithm has been applied by Rahmat *et al.* (2009b).

1.3.4 Miscellaneous Applications

Mass flow rate measurement, velocity measurement and process tomography imaging have been the most interesting areas for the electrostatic sensor application. However, the simplicity and flexibility of the sensor structure has led to some other innovative applications.

Particle mean size measurement using electrostatic sensor in a particle dense flow conveyor was investigated by Zhang and Yan (2003). Most probably the bigger

size of particles carries a larger amount of electric charge. When the particles transfer in a constant velocity and mass flow rate, the variation in particle size will change the magnitude of the sensor output signal. Zhang and Yan (2003) used different material to validate the proposed method and the results did not show a measurement error larger than 15%.

Portoghese (2005, 2008) have worked on monitoring the moisture and drying end point in bed of fluidized particles using triboelectric sensors. The triboelectric sensor is another name for the electrostatic sensor where the electric charge directly produced by particle impact and friction with the metal electrode. The moisture control is important in powder industries. The electrodes in their study were used to detect the electric charge from the liquid injection to the particles in drying the bed stages. The results showed that the magnitude of the detected electric charge (in the form of an electric current in the sensor output) follows the moisture content in the particle container. Less moisture of particles results in higher magnitudes of the output signal.

The following flow chart shows the electrostatic sensor applications and research areas which were discussed throughout this section.




Figure 1.1 Application and research areas for the electrostatic sensor

Three main areas of research for the sensor are mass flow rate, velocity, and concentration measurement. Particle size measurement and moisture analysis are categorized as miscellaneous applications because fewer research works have been done on these subjects.

1.4 Problem Statement

Instrumentation based on monitoring electric charge variations in a particle or in a particle dense flow using electrostatic sensor is well-developed and researched. The sensor has been mostly utilized for velocity measurement using the cross correlation as well as the spatial filtering techniques. Even the commercial version of the instrument is available (Yan and Ma, 2000). Other applications for the sensor so far include the measurement of mass flow rate and concentration profile using direct method and process tomography. Alongside the velocity and mass flow rate measurement, the particle size can affect the particle charging process. From that idea, there have been methods and research studies developed using the electric charge on a particle as a variable for particle size measurement.

There are two methods that employ particle charge properties to measure particle size. One of the methods is called the Differential Mobility Analyzer or DMA, which uses the electric mobility properties of charged aerosol to find size distribution of the aerosols within the air. The technique is well-researched and developed in both academic area (Guha *et al.*, 2011) and commercial models (Intra and Tippayawong, 2007). However, the method is limited to only aerosols and micron sized particles. The area for large particle size measurement based on the electric charge measurement has received less attention. The only provided method uses an electrostatic sensor and proposes that the mean size of particles in a mass flow has direct relation with the electric charge level produced by moving particles in a pipeline (Zhang and Yan, 2003). Nevertheless, the electric charge level on particle flow can be much more influenced by flow velocity and mass flow rate rather than the particle size. As a result, the small change on flow velocity and mass flow rate easily demolish the entire size measurement results.

The mean size measurement of a single particle in a magnitude dependent analysis is challenging. At first, hardly two particles with equivalent sizes, material types and densities can get an equal amount of electric charge on their surface. Therefore, they induce different levels of electrostatic noise in the sensor which produces distinct results in the measurement system for two particles of the same size. The problem mostly occurs in the measurement of the particles with dissimilar material types where each of them has specific relative permittivity and different behavior in an electric field.

Second, the electrostatic sensor basically detects the particle as a point charge not as a particle. For example, if we have two particles of different sizes with same electric charge on their surface (the case which may happen due to random processes of the particle charging), the measurement system will show that the particles are the same size which would be a wrong result.

To target the mentioned problems, this study is proposed which has pursued a new technique to deal with both challenges. In order to solve the first problem, the spatial filtering technique was employed. The method performs the measurement in the frequency domain, and it is independent from the amount of electric charge on the particle surface. However, the second problem will not be solved merely by using the spatial filtering method. A new modeling technique in conjunction with spatial filtering method is introduced to form a complete solution for the problem.

1.5 Research Objectives

The main purpose of this research work is to measure the mean size of a single particle using electrostatic sensor. To reach this purpose, the following objectives are defined for different steps of the work:

- (i) To design the electrode and the signal conditioning circuit of the sensor and to find its sensitivity properties.
- (ii) To formulate the frequency response of the electrostatic sensor.

- (iii) To formulate a mathematical model that provides the particle size information through spatial filtering method.
- (iv) To validate the proposed method via the experimental approach.

The proposed technique will be based on spatial filtering method where the measurement principles result from output signal analysis in the frequency domain. Mathematica software was used to verify modeling equations and Ansoft Maxwell software will be used for field simulation. Ansoft Maxwell uses Finite-Element Modeling method for field analysis. Excel and FindGraph software programs are utilized for plotting the graphs and for regression analysis. Multisim software is applied for design and analysis of the signal conditioning circuit. Dewetron data transfer card and DEWESoft software are employed for data collecting purposes in experimental tests.

1.6 Research Scope and Limitations

As the Project scope, the current electrostatic sensor applications and developments were studied. The information about the present technologies for particle sizing and the methods to measure the electric charge on particles were collected. The electrostatic sensor design, the methods to model the charge induced to the electrode and the spatial filtering effect of the electrode and spatial filtering method were investigated. To implement the proposed method, the sensor electrode and signal conditioning circuit were designed. The mathematical modeling of the frequency response of the sensor and relation between particle radial position and the frequency at the crest of the frequency spectrum (f_{cs}) were conducted to find the relation between particle size and f_{cs} . Finally, the proposed method was validated using the experimental tests.

However, the system is limited for the round particles. The accuracy of the measurement system is limited by the resolution frequency spectrum. The system cannot measure non-chargeable particles; nevertheless, the particle does not need to be charged by an external source. When small sized particles (<1mm) drop tangent to

the pipe wall, the electric charge on the pipe body either attracts or repels the electric field on the particle body which makes a drift on particle direction. Shielding the pipe reduces this effect, but it does not eliminate it completely. Directing the particles through the air flow would be a better solution. The particle does not need the external charger; however, the system cannot measure non-chargeable particles.

1.7 Research Contributions

The main contributions of this research are as follows:

- (i) The mathematical model is provided to explain a ring electrode behavior in the frequency domain when a particle moving in a radial direction.
- (ii) A new technique is introduced to provide the particle size information through spatial filtering method.

The design of the sensor included a ring electrode and signal conditioning circuit design. To monitor the electrode behavior in an electric field, a field simulator was employed to find the sensitivity of the sensor. The sensitivity of the sensor for different radial position of particle was recorded. The mathematical modeling was performed to find the relation between particle radial position and output frequency and in the same manner between particle size and output frequency. A gravitational test rig was designed, and the experimental results were compared with modeling results.

1.8 Thesis Outline

Chapter 1 provided a brief introduction of the entire research. Motivation and the reasons for conducting the research are explained. The structure and the modus operandi of the electrostatic sensor are described. A brief background of the electrostatic sensor application is provided. The problem statements highlighting the

existing problems in particle sizing using electrostatic sensor are given in detail. The research objective and project contribution to previous studies are provided.

Chapter 2 presents a concise literature review that consists of two main parts. In the first part, the most common methods in particle sizing are reviewed. The methods include physical methods, laser diffraction, imaging method, DMA, Coulter counter and miscellaneous methods. The second part is dedicated to everything about the electrostatic sensor in detail. The methods for electric charge measurement are described, the electrostatic sensor design is explained, and the methods to mathematically model the sensor are provided. In the last part, the spatial filtering effect and frequency properties of the electrode are described in detail.

Chapter 3 focuses on the measurement principles. Mathematical models, simulation process and hardware setup for experimental test are explained in this chapter. The mathematical models describe the spatial sensitivity of electrode, frequency response of electrode and frequency response of the sensor. The simulation procedure incorporated in finding the amount of the charge induced to electrode using the FEM field simulator is described in detail. The signal conditioning circuit design is explained. Lastly, the proposed method to find the particle size in the frequency domain is provided.

In chapter 4, the mathematical modeling results and experimental results are explained in detail. The chapter begins with the electrode design. The spatial sensitivity equation of the electrode is mathematically calculated. The spatial filtering effect and the electrode frequency response are analyzed. The sensor frequency response is calculated. The chapter closes with detailed explanation of the particle sizing algorithm, and the results from mathematical modeling and experimental tests are provided to validate the proposed method.

Chapter 5 presents the conclusion and possible future works. In this chapter, the contributions of this research work are explained. The limitations of this research as well as the suggested solutions are described. Finally, two proposals are suggested as future works which can develop the basic idea of particle size measurement used in this research.

1.9 Summary of the Introduction

Electrostatic sensor is widely researched for use in powder and particle industries. It is a robust sensor which consists of two main parts: the sensor electrode and signal conditioning circuit. The most important part of the sensor is the electrode whose geometrical shape and size determine the spatial sensitivity and spatial filtering effect of the sensor. The electrostatic sensor is studied to be used in mass flow rate measurement, velocity measurement and in process tomography. In this research, the application of electrostatic sensor was investigated in particle size measurement. A new technique which uses the spatial filtering method for particle size measurement was introduced. The proposed method is validated using mathematical modeling and experimental tests.

This chapter is started with explaining the motivations of doing the research. A brief explanation is provided about the electrostatic sensor structure. In the research background, the applications of the sensor are reviewed. These include mass flow rate measurement, velocity measurement, process tomography and two miscellaneous applications. The remaining parts of this chapter were dedicated to the research problem statement, the research objectives, the research scope and the research contributions. At the end, the thesis outline is provided.

REFERENCES

- Adi, H., Larson, I., and Stewart, P. (2007). Use of milling and wet sieving to produce narrow particle size distributions of lactose monohydrate in the sub-sieve range. *Powder Technology*. 179(1–2), 95-99.
- Allen, T. (2003a). 3 - Particle size analysis by image analysis. *Powder Sampling and Particle Size Determination*. (pp. 142-207). Amsterdam: Elsevier.
- Allen, T. (2003b). 4 - Particle size analysis by sieving. *Powder Sampling and Particle Size Determination*. (pp. 208-250). Amsterdam: Elsevier.
- Allen, T. (2003c). 7 - Gravitational sedimentation methods of particle size determination. *Powder Sampling and Particle Size Determination*. (pp. 359-391). Amsterdam: Elsevier.
- Allen, T. (2003d). 8 - Centrifugal sedimentation methods of particle size determination. *Powder Sampling and Particle Size Determination*. (pp. 392-446). Amsterdam: Elsevier.
- Analog-Devices. (2002). AMP02. USA. Available: www.analog.com 2012].
- Ansof (2005). *Ansoft Maxwell 3D Field Simulator VII User's Guide*. USA.
- Benes, P., Klusacek, S., and Pikal, P. (2004). In process measurement of particle size distribution. *Sensors, 2004. Proceedings of IEEE*. 24-27 Oct. 2004. 357-360 vol.1.
- Brunelli, N. A., Flagan, R. C., and Giapis, K. P. (2009). Radial Differential Mobility Analyzer for One Nanometer Particle Classification. *Aerosol Science and Technology*. 43, 53-59.
- Burr-Brown. (1997). OPA2604. USA. Available: www.burr-brown.com/databook/OPA2604.html 2012].
- Carter, R. M., and Yong, Y. (2005). An instrumentation system using combined sensing strategies for online mass flow rate measurement and particle sizing. *Instrumentation and Measurement, IEEE Transactions on*. 54(4), 1433-1437.

- Carter, R. M., Yong, Y., and Lee, P. (2006). On-line Nonintrusive Measurement of Particle Size Distribution Through Digital Imaging. *Instrumentation and Measurement, IEEE Transactions on*. 55(6), 2034-2038.
- Colver, G. M. (1999). 1 - Electrostatic Measurements. In SHAO LEE, S. (ed.) *Instrumentation for Fluid Particle Flow*. (pp. 47-111). Park Ridge, NJ: William Andrew Publishing.
- Dahl, C. K., and Esbensen, K. H. (2007). Image analytical determination of particle size distribution characteristics of natural and industrial bulk aggregates. *Chemometrics and Intelligent Laboratory Systems*. 89(1), 9-25.
- Dascalescu, L., Dragan, C., Bilici, M., and Chereches, R. (2010). Triboelectric Phenomena in Suction-Type Dilute-Phase Pneumatic Transportation Systems for Granular Plastics. *Industry Applications, IEEE Transactions on*. 46(4), 1570-1577.
- Fernández De La Mora, J., and Kozlowski, J. (2013). Hand-held differential mobility analyzers of high resolution for 1–30 nm particles: Design and fabrication considerations. *Journal of Aerosol Science*. 57(0), 45-53.
- Flagan, R. C. (2008). Differential Mobility Analysis of Aerosols: A Tutorial. *KONA Powder and Particle Journal*. 26, 254-268.
- Fleisch, D. (2008). *A student's Guide to Maxwell's Equations*. UK: Cambridge University Press.
- Floyd, T. L. (2012). *Electronic devices : conventional current version*. (pp. USA: Prentice Hall.
- Gajewski, J. B. (1989). Continuous non-contact measurement of electric charges of solid particles in pipes of pneumatic transport. I. Physical and mathematical models of a method. *Industry Applications Society Annual Meeting*. 1-5 Oct. 1989. 1958-1963.
- Gajewski, J. B. (1994). Measuring probes, head, and system for the non-contact, electrostatic measurements of the two-phase flow parameters in pneumatic transport of solids. *Journal of Electrostatics*. 32(3), 297-303.
- Gajewski, J. B. (1996a). Electrostatic, inductive ring probe bandwidth. *Measurement Science and Technology*. 7(12), 1766.
- Gajewski, J. B. (1996b). Monitoring electrostatic flow noise for mass flow and mean velocity measurement in pneumatic transport. *Journal of Electrostatics*. 37(4), 261-276.

- Gajewski, J. B. (1997a). Dynamic effect of charged particles on the measuring probe potential. *Journal of Electrostatics*. 40–41(0), 437-442.
- Gajewski, J. B. (1997b). Electric charge measurement in pneumatic installations. *Journal of Electrostatics*. 40-41(0), 231-236.
- Gajewski, J. B. (1999a). Electrostatic flow probe and measuring system calibration for solids mass flow rate measurement. *Journal of Electrostatics*. 45(4), 255-264.
- Gajewski, J. B. (1999b). Non-intrusive solids charge and mass flow measurements with an electrostatic flow probe. *Journal of Electrostatics*. 46(4), 271-284.
- Gajewski, J. B. (1999c). Static characteristics of an electrostatic flow probe. *Journal of Electrostatics*. 48(1), 49-64.
- Gajewski, J. B. (2000). Frequency response and bandwidth of an electrostatic flow probe. *Journal of Electrostatics*. 48(3–4), 279-294.
- Gajewski, J. B. (2008). Electrostatic Nonintrusive Method for Measuring the Electric Charge, Mass Flow Rate, and Velocity of Particulates in the Two-Phase Gas–Solid Pipe Flows—Its Only or as Many as 50 Years of Historical Evolution. *Industry Applications, IEEE Transactions on*. 44(5), 1418-1430.
- Gajewski, J. B., Glod, B., and Kala, W. (1990). Electrostatic method for measuring the two-phase pipe flow parameters. *Industry Applications Society Annual Meeting, 1990., Conference Record of the 1990 IEEE*. 7-12 Oct. 1990. 897-902 vol.1.
- Gajewski, J. B., Glod, B., Grobelny, R. A., and Kala, W. (1989). Continuous non-contact measurement of electric charges of solid particles in pipes of pneumatic transport. II. Measuring system and its application. *Industry Applications Society Annual Meeting*. 1-5 Oct. USA. 1964-1969.
- Gajewski, J. B., Glod, B. J., and Kala, W. S. (1993a). Electrostatic method for measuring the two-phase pipe flow parameters. *Industry Applications, IEEE Transactions on*. 29(3), 650-655.
- Gajewski, J. B., Kacprzyk, R., and Zuk, J. (1993b). Electrostatic, noncontact, continuous, and real-time velocity measurements in pneumatic transport pipes. *Industry Applications Society Annual Meeting, Conference Record of the IEEE*. 2-8 Oct 1993. 1709-1713.
- Gao, L., Yan, Y., Lu, G., and Carter, R. M. (2012). On-line measurement of particle size and shape distributions of pneumatically conveyed particles through

- multi-wavelength based digital imaging. *Flow Measurement and Instrumentation*. 27(0), 20-28.
- Ghazali, M. R. (2010). *Electrostatic Sensor Design and Sensor Characteristic Study for Electrostatic Charge Measurement*. Master, Universiti teknologi Malaysia.
- Graham, M. D., and Coulter, B. (2003). The Coulter Principle: Foundation of an Industry. *JALA*. 72-81.
- Green, R. G., Horbury, N. M., Rahim, R. A., Dickin, F. J., Naylor, B. D., and Pridmore, T. P. (1995). Optical fibre sensors for process tomography. *Measurement Science and Technology*. 6(12), 1699.
- Green, R. G., Rahmat, M. F., Dutton, K., Evans, K., Goude, A., and Henry, M. (1997a). Velocity and mass flow rate profiles of dry powders in a gravity drop conveyor using an electrodynamic tomography system. *Measurement Science and Technology*. 8(4), 429.
- Green, R. G., Rahmat, M. F., Evans, K., Goude, A., Henry, M., and Stone, J. a. R. (1997b). Concentration profiles of dry powders in a gravity conveyor using an electrodynamic tomography system. *Measurement Science and Technology*. 8(2), 192.
- Guha, S., Pease Iii, L. F., Brorson, K. A., Tarlov, M. J., and Zachariah, M. R. (2011). Evaluation of electrospray differential mobility analysis for virus particle analysis: Potential applications for biomanufacturing. *Journal of Virological Methods*. 178(1-2), 201-208.
- Hammer, E. A., and Green, R. G. (1983). The spatial filtering effect of capacitance transducer electrodes (flow measurement). *Journal of Physics E: Scientific Instruments*. 16(5), 438-443.
- Hancke, G. P., and Malan, R. (1997). A modal analysis technique for the on-line particle size measurement of pneumatically conveyed pulverised coal. *Instrumentation and Measurement Technology Conference, 1997. IMTC/97. Proceedings. Sensing, Processing, Networking., IEEE*. 19-21 May 1997. 848-852 vol.2.
- Hancke, G. P., and Malan, R. (1998). A modal analysis technique for the on-line particle size measurement of pneumatically conveyed pulverized coal. *Instrumentation and Measurement, IEEE Transactions on*. 47(1), 114-122.

- He, G.-C., Mao, Y.-P., and Ni, W. (2007). A new fractal modification of ultrasonic attenuation model for measuring particle size in mineral slurries. *International Journal of Mineral Processing*. 82(3), 119-125.
- Humphries, S. (2010). *Finite-element Methods for electromagnetics* USA.
- Hussain, T., Kaialy, W., Deng, T., Bradley, M. S. A., Nokhodchi, A., and Armour-Chélu, D. (2013). A novel sensing technique for measurement of magnitude and polarity of electrostatic charge distribution across individual particles. *International Journal of Pharmaceutics*. 441(1-2), 781-789.
- Idroas, M., Rahim, R. A., Green, R. G., Ibrahim, M. N., and Rahiman, M. H. F. (2010). Image Reconstruction of a Charge Coupled Device Based Optical Tomographic Instrumentation System for Particle Sizing. *Sensors*. 10, 9512-9528.
- Instrument, T. (2009). *An Applications Guide for Op Amps*. USA: National Semiconductor Corporation.
- Intra, P., and Tippayawong, N. (2006). Aerosol size distribution measurement using multi-channel electrical mobility sensor. *J. Aerosol Res*. 21(4), 329-340.
- Intra, P., and Tippayawong, N. (2007). An overview of aerosol particle sensors for size distribution measurement. *Maejo International Journal of Science and Technology*. 120-136.
- Intra, P., and Tippayawong, N. 2008a. An Electrostatic Sensor for Nanometer-Sized Aerosol Particles Detection. *Asia-Pacific Symposium on Applied Electromagnetics and Mechanics*. thailand.
- Intra, P., and Tippayawong, N. (2008b). An overview of differential mobility analyzers for size classification of nanometer-sized aerosol particles. *Songklanakarin Journal of Science and Technology*. 30, 243-256.
- Jagtiani, A. V., Carletta, J., and Zhe, J. (2011). An impedimetric approach for accurate particle sizing using a microfluidic Coulter counter. *Journal of Micromechanics and Microengineering*. 21(4), 045036.
- Jiaqing, S., Krabicka, J., and Yong, Y. (2010). Velocity Measurement of Pneumatically Conveyed Particles Using Intrusive Electrostatic Sensors. *Instrumentation and Measurement, IEEE Transactions on*. 59(5), 1477-1484.
- Jones, A. R., Parasram, N. T., and Taylor, A. M. K. P. (2002). Numerical simulation of the sizing performance of the shadow Doppler velocimeter (SDV). *Measurement Science and Technology*. 13(3), 317-330.

- Krabicka, J., and Yong, Y. (2009). Finite-Element Modeling of Electrostatic Sensors for the Flow Measurement of Particles in Pneumatic Pipelines. *Instrumentation and Measurement, IEEE Transactions on*. 58(8), 2730-2736.
- Kumar, B. R., Basheer, N. S., George, S. D., and Kurian, A. (2012). Effect of gold nanoparticle size on fluorescence quenching of rhodamine 6G measured using dual beam thermal lens technique. *Photonics (ICP), 2012 IEEE 3rd International Conference on*. 1-3 Oct. 2012. 149-153.
- Lee Black, D., Mcquay, M. Q., and Bonin, M. P. (1996). Laser-based techniques for particle-size measurement: A review of sizing methods and their industrial applications. *Progress in Energy and Combustion Science*. 22(3), 267-306.
- Li, J., Xu, C., and Wang, S. (2012). Local particle mean velocity measurement using electrostatic sensor matrix in gas–solid two-phase pipe flow. *Flow Measurement and Instrumentation*. 27(0), 104-112.
- Lijun, X., Carter, R. M., and Yong, Y. (2005). Mass Flow Measurement of Fine Particles in a Pneumatic Suspension Using Electrostatic Sensing and Neural Network Techniques. *Instrumentation and Measurement Technology Conference, Proceedings of the IEEE*. 16-19 May 2005. 1365-1368.
- Liu, K. (2009). Some factors affecting sieving performance and efficiency. *Powder Technology*. 193(2), 208-213.
- Ma, J., and Yan, Y. (2000). Design and evaluation of electrostatic sensors for the measurement of velocity of pneumatically conveyed solids. *Flow Measurement and Instrumentation*. 11(3), 195-204.
- Machida, M., and Kaminoyama, M. (2008). Sensor Design for Development of Tribo-Electric Tomography System with Increased Number of Sensors. *Journal of Visualization*. 11, 375-385.
- Machida, M., and Scarlett, B. (2005). Process tomography system by electrostatic charge carried by particles. *Sensors Journal, IEEE*. 5(2), 251-259.
- Mcclements, D. J., and Coupland, J. N. (1996). Theory of droplet size distribution measurements in emulsions using ultrasonic spectroscopy. *Colloids and Surfaces A: Physicochemical and Engineering Aspects*. 117(1–2), 161-170.
- Mcgarvey, M., Mcgregor, D., and Mckay, R. B. (1997). Particle size analysis by laser diffraction in organic pigment technology. *Progress in Organic Coatings*. 31(3), 223-228.

- Mei, F., Fu, H., and Chen, D.-R. (2011). A cost-effective differential mobility analyzer (cDMA) for multiple DMA column applications. *Journal of Aerosol Science*. 42(7), 462-473.
- Mizutani, Y., Kodama, H., and Miyasaka, K. (1982). Doppler-Mie combination technique for determination of size-velocity correlation of spray droplets. *Combustion and Flame*. 44(1-3), 85-95.
- Morikita, H., and Taylor, A. M. K. P. (1998). Application of shadow Doppler velocimetry to paint spray: potential and limitations in sizing optically inhomogeneous droplets. *Measurement Science and Technology*. 9(2), 221-231.
- Peng, L., Zhang, Y., and Yan, Y. (2008). Characterization of electrostatic sensors for flow measurement of particulate solids in square-shaped pneumatic conveying pipelines. *Sensors and Actuators A: Physical*. 141(1), 59-67.
- Pike, R. (2002). Chapter 2.4.1 - Particle Sizing by Laser Light Scattering. In ROY, P. & PIERRE, S. (eds.) *Scattering*. (pp. 895-919). London: Academic Press.
- Portoghese, F., Berruti, F., and Briens, C. (2005). Use of triboelectric probes for on-line monitoring of liquid concentration in wet gas-solid fluidized beds. *Chemical Engineering Science*. 60(22), 6043-6048.
- Portoghese, F., Berruti, F., and Briens, C. (2008). Continuous on-line measurement of solid moisture content during fluidized bed drying using triboelectric probes. *Powder Technology*. 181(2), 169-177.
- Qian, X., Yan, Y., Shao, J., Wang, L., Zhou, H., and Wang, C. (2012). Quantitative characterization of pulverized coal and biomass-coal blends in pneumatic conveying pipelines using electrostatic sensor arrays and data fusion techniques. *Measurement Science and Technology*. 23(8), 085307.
- Rahmat, M. F., and Green, R. G. (1999). Sensitivity and Spatial Filtering Effect for an Electrodynamic Sensor. *Jurnal ElektriKa*. 2, 24-29.
- Rahmat, M. F., Isa, M. D., Rahim, R. A., and Hussin, T. a. R. (2009a). Electrodynamic Sensor for the Image Reconstruction Process in an Electrical Charge Tomography System. *Sensors*. 9, 10291-10308.
- Rahmat, M. F., Kamaruddin, N. S., and Isa, M. D. (2009b). Flow Regime Identification in Pneumatic Conveyor Using Electrodynamic Transducer and Fuzzy Logic Method. *International Journal on Smart Sensing and Intelligent Systems*. 2, 396-416.

- Rahmat, M. F., and Lee, D. Y. W. (2004). Electrostatic sensor for real-time mass flow rate measurement of particle conveying in pneumatic pipeline. *Jurnal Teknologi*. 91–104.
- Rahmat, M. F., and Rahiman, M. H. F. (2001). Real-time velocity profile generation of powder conveying using electrodynamic transducer. *Jurnal Teknologi*. 35, 27-38.
- Rahmat, M. F., and Sabit, H. A. (2004). Flow regime identification using neural network based electrodynamic tomography system. *Jurnal Teknologi*. 40, 109–118.
- Rahmat, M. F., and Sabit, H. A. (2007). Application of neural network technique and electrodynamic sensors in the identification of solid flow regimes. *Jurnal Teknologi*. 46, 77-92.
- Rahmat, M. F. A., Isa, M. D., Rahim, R. A., and Hussin, T. a. R. (2009c). Electrodynamic Sensor for the Image Reconstruction Process in an Electrical Charge Tomography System. *Sensors*. 9, 10291-10308.
- Richards, A. L., Dickey, M. D., Kennedy, A. S., and Buckner, G. D. (2012). Design and demonstration of a novel micro-Coulter counter utilizing liquid metal electrodes. *Journal of Micromechanics and Microengineering*. 22(11), 115012.
- Richter, A., Babick, F., and Stintz, M. (2006). Polydisperse particle size characterization by ultrasonic attenuation spectroscopy in the micrometer range. *Ultrasonics*. 44, Supplement(0), e483-e490.
- Richter, A., Voigt, T., and Ripperger, S. (2007). Ultrasonic attenuation spectroscopy of emulsions with droplet sizes greater than 10 μm . *Journal of Colloid and Interface Science*. 315(2), 482-492.
- Rodrigues, S. J., and Yong, Y. (2012). A comparative study of rounded and strip electrostatic sensors for non-contact measurement of cable speed. *Instrumentation and Measurement Technology Conference (I2MTC), 2012 IEEE International*. 13-16 May 2012. 1159-1162.
- Rodriguez-Trujillo, R., Castillo-Fernandez, O., Garrido, M., Arundell, M., Valencia, A., and Gomila, G. (2008). High-speed particle detection in a micro-Coulter counter with two-dimensional adjustable aperture. *Biosensors and Bioelectronics*. 24(2), 290-296.

- Sandler, N. (2011). Photometric imaging in particle size measurement and surface visualization. *International Journal of Pharmaceutics*. 417(1–2), 227-234.
- Snowsill, W. L. (2010). Chapter 16 - Particle Sizing. *Instrumentation Reference Book (Fourth Edition)*. (pp. 175-189). Boston: Butterworth-Heinemann.
- Sowinski, A., Miller, L., and Mehrani, P. (2010). Investigation of electrostatic charge distribution in gas–solid fluidized beds. *Chemical Engineering Science*. 65(9), 2771-2781.
- Sowinski, A., Salama, F., and Mehrani, P. (2009). New technique for electrostatic charge measurement in gas–solid fluidized beds. *Journal of Electrostatics*. 67(4), 568-573.
- Svarovsky, L., and Friedova, J. (1972). Homogeneous sedimentation in the centrifugal field. *Powder Technology*. 5(5), 273-277.
- Takahashi, N., Aki, A., Ukai, T., Nakajima, Y., Maekawa, T., and Hanajiri, T. (2011). Proposal and experimental validation of the electrophoretic Coulter method for analyzing microparticles and biological cells. *Sensors and Actuators B: Chemical*. 151(2), 410-415.
- Tamhane, T. V., Joshi, J. B., Mudali, K., Natarajan, R., and Patil, R. N. (2012). Measurement of drop size characteristics in annular centrifugal extractors using phase Doppler particle analyzer (PDPA). *Chemical Engineering Research and Design*. 90(8), 985-997.
- Tayali, N. E., and Bates, C. J. (1990). Particle sizing techniques in multiphase flows: A review. *Flow Measurement and Instrumentation*. 1(2), 77-105.
- Taylor, D. M. (2001). Measuring techniques for electrostatics. *Journal of Electrostatics*. 51–52(0), 502-508.
- Tichenor, D. A., Mitchell, R. E., Hencken, K. R., and Niksa, S. (1985). Simultaneous in situ measurement of the size, temperature and velocity of particles in a combustion environment. *Symposium (International) on Combustion*. 20(1), 1213-1221.
- Wenbiao, Z., Chao, W., and Yulin, W. (2010). Parameter Selection in Cross-Correlation-Based Velocimetry Using Circular Electrostatic Sensors. *IEEE Transaction on Instrumentation and Measurement*. 59(5), 1268-1275.
- Woodhead, S. R., Denham, J. C., and Armour-Chelu, D. I. (2005). Electrostatic sensors applied to the measurement of electric charge transfer in gas–solids pipelines. *Journal of Physics: Conference Series*. 15(1), 108.

- Xiangchen, Q., and Yong, Y. (2012). Flow Measurement of Biomass and Blended Biomass Fuels in Pneumatic Conveying Pipelines Using Electrostatic Sensor-Arrays. *Instrumentation and Measurement, IEEE Transactions on.* 61(5), 1343-1352.
- Xiaoshu, C., Gang, Z., Naining, W., and Hongwu, Z. (1995). A new on-line measurement method for particle sizing. *Journal of Aerosol Science.* 26, Supplement 1(0), S761-S762.
- Xu, C., Li, J., Gao, H., and Wang, S. (2012a). Investigations into sensing characteristics of electrostatic sensor arrays through computational modelling and practical experimentation. *Journal of Electrostatics.* 70(1), 60-71.
- Xu, C., Li, J., and Wang, S. (2012b). A spatial filtering velocimeter for solid particle velocity measurement based on linear electrostatic sensor array. *Flow Measurement and Instrumentation.* 26(0), 68-78.
- Xu, C., Liang, C., Zhou, B., and Wang, S. (2010a). HHT analysis of electrostatic fluctuation signals in dense-phase pneumatic conveying of pulverized coal at high pressure. *Chemical Engineering Science.* 65(4), 1334-1344.
- Xu, C., Tang, G., Zhou, B., and Wang, S. (2009). The spatial filtering method for solid particle velocity measurement based on an electrostatic sensor. *Measurement Science and Technology.* 20(4), 045404.
- Xu, C., Wang, S., Tang, G., Yang, D., and Zhou, B. (2007). Sensing characteristics of electrostatic inductive sensor for flow parameters measurement of pneumatically conveyed particles. *Journal of Electrostatics.* 65(9), 582-592.
- Xu, C., Wang, S., and Yan, Y. (2013). Spatial Selectivity of Linear Electrostatic Sensor Arrays for Particle Velocity Measurement. *IEEE Transactions on Instrumentation and Measurement.* 62(1), 167-176.
- Xu, C., Zhou, B., and Wang, S. (2010b). Dense-phase pneumatically conveyed coal particle velocity measurement using electrostatic probes. *Journal of Electrostatics.* 68(1), 64-72.
- Xu, C., Zhou, B., Yang, D., Tang, G., and Wang, S. (2008). Velocity measurement of pneumatically conveyed solid particles using an electrostatic sensor. *Measurement Science and Technology.* 19(2), 024005.
- Yamamoto, T., Watanabe, N., Fukui, K., and Yoshida, H. (2009). Effect of inner structure of centrifugal separator on particle classification performance. *Powder Technology.* 192(3), 268-272.

- Yan, Y. (1996). Mass flow measurement of bulk solids in pneumatic pipelines. *Measurement Science and Technology*. 7(12), 1687.
- Yan, Y., Byrne, B., Woodhead, S., and Coulthard, J. (1995). Velocity measurement of pneumatically conveyed solids using electrodynamic sensors. *Measurement Science and Technology*. 6(5), 515.
- Yan, Y., and Ma, J. (2000). Measurement of particulate velocity under stack-flow conditions. *Measurement Science and Technology*. 11(1), 59.
- Yan, Y., Xie, Z., Krabicka, J., and Shao, J. (2010). Non-contact strip speed measurement using electrostatic sensors. *Instrumentation and Measurement Technology Conference (I2MTC), 2010 IEEE*. 3-6 May 2010. 1535-1538.
- Yoshida, H., Masuda, H., Fukui, K., and Tokunaga, Y. (2001). Particle size measurement with an improved sedimentation balance method and microscopic method together with computer simulation of necessary sample size. *Advanced Powder Technology*. 12(1), 79-94.
- Zaidi, S. H., Altunbas, A., and Azzopardi, B. J. (1998). A comparative study of phase Doppler and laser diffraction techniques to investigate drop sizes in annular two-phase flow. *Chemical Engineering Journal*. 71(2), 135-143.
- Zhang, J. (2002). *A Study of an Electrostatic Flow Meter*. PhD, University of Teesside.
- Zhang, J., and Coulthard, J. (2005). Theoretical and experimental studies of the spatial sensitivity of an electrostatic pulverised fuel meter. *Journal of Electrostatics*. 63(12), 1133-1149.
- Zhang, J., Xu, D., Coulthard, J., Xu, C., and Wang, S. (2009a). Analyses of Characteristics of Ring-Shaped Electrostatic Meter. *Chemical Engineering Communications*. 197, 192-203.
- Zhang, J., Zhou, B., and Wang, S. (2010). Effect of axial and radial velocities on solids mass flow rate measurement. *Robotics and Computer-Integrated Manufacturing*. 26(6), 576-582.
- Zhang, J., Zhou, B., Xu, C., and Wang, S. (2009b). Modelling and calibration of a ring-shaped electrostatic meter. *Journal of Physics: Conference Series*. 147(1), 012001.
- Zhang, J. Q., and Yan, Y. (2003). On-line continuous measurement of particle size using electrostatic sensors. *Powder Technology*. 135-136, 164-168.

- Zhe, K., Xiao-Lei, W., and Shu-Jiang, Z. (2013). Study on the spatial filtering and sensitivity characteristic of inserted electrostatic sensors for the measurement of gas–solid two-phase flow parameters. *Flow Measurement and Instrumentation*. 30(0), 26-33.
- Zheng, Y., and Liu, Q. (2010). Review of certain key issues in indirect measurements of the mass flow rate of solids in pneumatic conveying pipelines. *Measurement*. 43(6), 727-734.
- Zheng, Y., and Liu, Q. (2011). Review of techniques for the mass flow rate measurement of pneumatically conveyed solids. *Measurement*. 44(4), 589-604.