

COMPARISON BETWEEN A-MODE AND B-MODE ULTRASOUND IN
LOCAL HYPERTHERMIA MONITORING

MAIZATUL NADWA BINTI CHE AZIZ

A thesis submitted in fulfillment of the
requirements for the award of the degree of
Master of Philosophy (Biomedical Engineering)

Faculty of Biosciences and Medical Engineering
Universiti Teknologi Malaysia

AUGUST 2017

Dedicated, in thankful appreciation for support, encouragement and understandings to my beloved mother, father, sisters, brothers and supporters.

ACKNOWLEDGEMENT

This humble piece of research work has been completed by the grace of Almighty Allah, all glory to Him. All praise for Allah who guides us in darkness and help us in difficulties and respects are for His Holy Prophet, Muhammad (Peace Be Upon Him) Who enables us to recognize our creator.

First and foremost, I would like to express my heartily gratitude to my dedicated supervisor, Dr Maheza Irna Mohamad Salim for the guideline, endless support and enthusiasm given throughout the progress of this research. The supervision and helping hand that she gave truly boost the progression and smoothness of the research. The co-operation is much indeed appreciated.

The special thanks given to the Ministry of Higher Education (MOHE) and Universiti Teknologi Malaysia (UTM) for the awarding of the MyBrain Scholarship and the Research University Grant Scheme (4F274). Without their support, the expansion of knowledge in conducting this research could not been possible to achieved.

My appreciation also goes to my family members and beloved who has been so tolerant and supports me all these years. Special thanks also to all the members of Medical Imaging Instrumentation and Robotic (MIIR) laboratory and to all personal contributions that I could never repay. Thanks for their encouragement, love, emotional support that have been given to me. There is no meaningful word than Thank You So Much.

ABSTRACT

Hyperthermia therapy is one of the therapy methods used for cancer treatment. It has shown to be an effective way of treating the cancerous tissue when compared to surgery, chemotherapy and radiation. However, real time monitoring method is capable in delivering a consistent heat and preventing any damages to the nearby tissue. Ultrasound is among the widely used technique in clinical setting. A-Mode ultrasound involves one-dimensional (1D) signal processing which enables a quantitative measurement on different types of breast tissues to be conducted faster as it has relatively simple signal processing requirement. On the other hand, B-Mode ultrasound offers good spatial resolution for thermal monitoring. Therefore, the aim of this study is to investigate and to compare the most optimum A-Mode and B-Mode ultrasound parameters to monitor hyperthermia in normal and pathological breast tissue. A series of experiment was conducted on 40 female Sprague Dawley rats. The pathological and normal rats were dissected and exposed to hyperthermia at variation temperature of 37°C (body temperature) and 40°C, 45°C, 50°C and 55°C for hyperthermia temperatures. A-Mode and B-Mode of 7.5 Mhz and 6Mhz was used simultaneously during the experiment for collecting acoustic information and scanning purposes before and after the hyperthermia exposure. Result obtained shows that, for normal tissue condition of both A-Mode and B-Mode, the attenuation calculation to mean of pixel intensity found to be (3.59 ± 0.04) dB and 187.68 at temperature value of 50 °C. Meanwhile, in pathological tissue condition, the attenuation value with respect to pixel intensity was obtained by (3.36 ± 0.26) dB at temperature value of 45°C and 199.26 was achieved at temperature value of 40°C. For backscatter coefficient to variance analysis, the result found that, in both A-Mode and B-Mode normal tissue condition, at temperature value of 40°C, (1.81 ± 0.25) of backscatter coefficient was obtained while at 45°C, the variance value of 3298.94 was achieved. In pathological tissue, the temperature value of 40°C and 55°C was the most pronounce temperature dependent of (1.45 ± 0.28) for backscatter coefficient with respect to 3275.35 of variance analysis. The result obtained from artificial neural network have shown that, 91.67% to 87.5% of testing to validation percentage accuracy of A-Mode was achieved, while in B-Mode, 88.89% and 81.25% of testing and validation data was obtained. Therefore, it is shown that, the use of A-Mode with comparison to B-Mode ultrasound can be used as another potential approach since its attenuation to pixel intensity and backscatter coefficient with respect to variance of A-Mode and B-Mode is very sensitive to the tissue structure in monitoring hyperthermia therapy with respect to the changes of temperature.

ABSTRAK

Terapi hipertermia merupakan salah satu kaedah terapi untuk rawatan kanser. Ia telah terbukti untuk dijadikan salah satu cara yang berkesan untuk merawat tisu kanser berbanding pembedahan, kemoterapi dan radiasi. Hipertermia memerlukan kaedah pemantauan masa yang nyata bagi meningkatkan keupayaan penghantaran haba yang tetap dan mencegah sebarang kecederaan terhadap tisu yang berdekatan. Ultra bunyi adalah teknik yang digunakan secara meluas dalam bidang klinikal. Ultra bunyi A-Mod melibatkan pemprosesan isyarat satu dimensi (1D) yang membolehkan ukuran kuantitatif dapat dijalankan dengan lebih cepat terhadap pelbagai jenis tisu payudara kerana ia mempunyai keperluan pemprosesan isyarat yang agak mudah. Malah, ultra bunyi B-Mod menawarkan resolusi spatial yang baik untuk memantau haba. Tujuan kajian ini adalah untuk menyiasat dan membandingkan parameter yang paling sesuai bagi ultra bunyi A-Mod dan ultra bunyi B-Mod dalam pemantauan hipertermia terhadap tisu payudara yang normal dan patologi. Satu siri eksperimen telah dijalankan terhadap 40 ekor tikus betina jenis Spague Dawley. Tikus patologi dan normal telah dibedah dan terdedah kepada hipertermia kepada pelbagai suhu dimana 37°C (suhu badan) dan 40°C, 45°C, 50°C dan 55°C merupakan suhu hipertermia. Sebanyak 7.5Mhz dan 6Mhz frekuensi A-Mod dan B-Mod telah digunakan serentak semasa eksperimen dijalankan bagi tujuan mengumpul dan mengimbas maklumat akustik sebelum dan selepas pemantauan terapi hipertermia. Hasil kajian menunjukkan, untuk tisu normal kedua-dua A-Mod dan B-Mod, pengiraan pengecilan terhadap keamatan piksel didapati (3.59 ± 0.04) dB dan 187.68 pada suhu 50 °C. Untuk tisu patologi, nilai pengecilan terhadap keamatan piksel telah diperolehi sebanyak (3.36 ± 0.26) dB pada suhu 45°C dan 199.26 dicapai pada suhu 40 °C. Untuk penyerakan pekali terhadap varian analisis, keputusan mendapati, untuk kedua-dua ultra bunyi bagi keadaan tisu normal, pada suhu 40°C, sebanyak (1.81 ± 0.25) penyerakan pekali telah diperolehi dan pada suhu 45°C, varian sebanyak 3298.94 telah dicapai. Bagi tisu patologi, suhu 40°C dan 55°C telah dipilih sebagai suhu paling optimum dengan perolehan penyerakan pekali sebanyak (1.45 ± 0.28) terhadap analisis varian sebanyak 3275.35. Peratusan rangkaian neural A-Mod dan B-Mod menunjukkan, sebanyak 91.67 % dan 87.5%, 88.89% dan 81.25% daripada data percubaan dan data validasi dicapai. Ini menunjukkan, kegunaan ultra bunyi A-Mod dengan perbandingan B-Mod berpotensi dan sangat sensitif terhadap struktur tisu dalam memantau terapi hipertermia dengan perubahan suhu yang pelbagai.

TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	DEDICATION	iii
	ACKNOWLEDGMENTS	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	xi
	LIST OF FIGURES	xii
	LIST OF ABBREVIATIONS	xiv
	LIST OF APPENDICES	xv
1	INTRODUCTION	1
	1.1 Research Background	1
	1.2 Problem Statement	4
	1.3 Research Objectives	5
	1.4 Scope of Research	6
	1.5 Significance of Study	6
	1.6 Thesis Structure Organization	7
2	LITERATURE REVIEW	9
	2.1 Hyperthermia Therapy Monitoring	9
	2.1.1 Radio Frequency Ablation (RFA)	11
	2.1.2 Laser Ablation (LA)	12
	2.1.3 Microwave Ablation (MWA)	12
	2.1.4 High Intensity Focused Ultrasound (HIFU)	13

2.2	Gold Standard in Hyperthermia Therapy Monitoring	17
2.3	Fundamental of Ultrasound	20
2.4	Ultrasonography in Hyperthermia Therapy Monitoring	24
2.5	Parameters that are Sensitive to Temperature Changes and in A-Mode and B-Mode Ultrasound	27
2.6	Machine Learning Method for Performance Comparison	34
2.6.1	Support Vector Machine (SVM)	35
2.6.2	Artificial Neural Network (ANN)	37
2.7	Summary	41
3	METHODOLOGY	42
3.1	Introduction	42
3.2	Research Flowchart	43
3.3	Literature Review	44
3.4	Experimental Set Up and Protocol	44
3.4.1	Jig Design	44
3.4.2	Experimental Procedure	45
3.4.3	Heating Method using Water Bath	46
3.5	Animal Handling, Cancer Induction and Tissue Preparation	48
3.5.1	Animal Handling	48
3.5.2	Cancer Induction	49
3.5.3	Tissue Preparation	50
3.6	Data Collection	52
3.6.1	A-Mode Ultrasound Parameters	53
3.6.2	B-Mode Ultrasound Parameters	55
3.6.2.1	Image Processing Flowchart	55
3.6.2.2	Theoretical Approach	58
3.6.2.3	Image Segmentation Workflow	60
3.6.3	Linear Regression	61
3.6.3.1	Least Squares Estimation	62
3.7	Development of Artificial Neural Network (ANN) for Breast Cancer Classification	63
3.7.1	Data Massaging and Data Sampling	65

3.7.2	Configuration of ANN Development	66
3.7.3	Training of ANN Development	66
3.7.4	Optimization of ANN Development	69
3.7.5	The Output of the Artificial Neural Network (ANN) for Breast Cancer Classification of A-Mode and B-Mode	69
3.7.5.1	Optimization Number of Neurons in Hidden Layer for A-Mode and B-Mode Ultrasound Data	70
3.7.5.2	Optimization of Learning Rate in A-Mode and B-Mode Ultrasound Data	73
3.7.5.3	Optimization of Iteration Rate in A-Mode and B-Mode Ultrasound Data	75
3.7.5.4	Optimization of Momentum Constant in A-Mode and B-Mode Ultrasound Data	77
4	RESULTS AND DISCUSSION	79
4.1	Introduction	79
4.2	A-Mode Ultrasound Analysis	80
4.2.1	Attenuation and Backscatter Coefficient Analysis of A-Mode	80
4.3	B- Mode Ultrasound Analysis	84
4.3.1	Mean Grey Scale of Pixel Intensity and Variance of B- Mode Ultrasound Image Analysis in Normal and Pathological Tissue	84
4.4	Comparative Performance Analyses of A-Mode and B-Mode Ultrasound	88
4.5	Development of the Artificial Neural Network (ANN) for Breast Cancer Classification of A-Mode and B-Mode	94

5	CONCLUSION	98
	5.1 Conclusion	98
	5.2 Recommendation and Future Work	99
	REFERENCES	101
	Appendices A-F	126-140

LIST OF TABLES

TABLE NO.	TITLE	PAGE
2.1	Summary of Available Method for Thermal Ablation	16
2.2	A Comparison of Attenuation Coefficient at 5MHz in Canine Tissue Liver at Different Temperature	30
3.1	Training Result for A-Mode and B-Mode	67
3.2	Number of ANN Data Sampling	70
4.1	Number of ANN Data Sampling	94
4.2	Overall Classification Performance Accuracy of ANN	95

LIST OF FIGURES

FIGURE NO.	TITLE	PAGE
2.1	Preview of A-Mode Scan	21
2.2	Surface Reflection of Ultrasound Wave	23
2.3	Preview of B-Mode Scan	24
2.4	Attenuation Coefficient in Canine Liver Tissue against Temperature at 4, 5, and 6MHz	30
2.5	The Neuron Model	38
2.6	An Example of Basic Diagram of ANN	38
3.1	Flowchart of Determining A-Mode and B-Mode Ultrasound Parameter for Hyperthermia Monitoring	43
3.2	Jig Design for Experimental Set Up	45
3.3	Block Diagram of Experimental Set Up	47
3.4	Experimental Set Up for A-Mode and B-Mode Hyperthermia Monitoring	47
3.5	Normal and DMBA Induced Virgin Female Sprague Dawley	48
3.6	The Weight of Each Rat was Measured Every Week	49
3.7	Rat with Tumor after 15 Week Induction of DMBA	50
3.8	Location of Mammary Tissue Harvested from Each Rat	51
3.9	Extraction of Breast Tumor from Each Rat	52
3.10	Image Processing Flowchart	57
3.11	Active Contour Segmentation Workflow	61
3.12	Flowchart of the Development of Multilayer Feed Forward Neural Network	65
3.13	Neural Network Training Tool Window	68

3.14	Neuron Numbers in Hidden Layer against MSE and Testing Performance Accuracy (%) of A-Mode	72
3.15	Neuron Numbers in Hidden Layer against MSE and Testing Performance Accuracy (%) of B-Mode	72
3.16	Learning Rate against MSE and Testing Performance Accuracy (%) of A-Mode	74
3.17	Learning Rate against MSE and Testing Performance Accuracy (%) of B-Mode	74
3.18	Iteration Rate against MSE and Testing Performance Accuracy (%) of A-Mode	76
3.19	Iteration Rate against MSE and Testing Performance Accuracy (%) of B-Mode	76
3.20	Momentum Constant against MSE and Testing Performance Accuracy (%) of A-Mode	78
3.21	Momentum Constant against MSE and Testing Performance Accuracy (%) of B-Mode	78
4.1	Attenuation vs Temperature in Normal and Pathological Tissue	75
4.2	Backscatter coefficient vs Temperature in Normal and Pathological Tissue	81
4.3	Average Mean of Pixel Intensity of B-Mode Images in Normal and Pathological Tissue	85
4.4	Average Variance of B-Mode Images in Normal and Pathological Tissue	86
4.5	Attenuation Coefficient of A-Mode against Mean Grey Scale of B-Mode for Comparative Analysis	89
4.6	Regression Plot of A-Mode Attenuation against Mean Grey Scale of B-Mode for Comparative Analysis	89
4.7	Backscatter Coefficient of A-Mode against Variance of B-Mode for Comparative Analysis	91
4.8	Regression Plot of Backscatter Coefficient of A-Mode against Variance of B-Mode for Comparative Analysis	91

LIST OF ABBREVIATIONS

MRI	-	Magnetic Resonance Imaging
RFA	-	Radio Frequency Ablation
LA	-	Laser Ablation
MWA	-	Microwave Ablation
HIFU	-	High Intensity Focused Ultrasound
H&E	-	Hematoxylin and Eosin
PRFS	-	Proton Resonance Frequency Shift
ARFI	-	Acoustic Radiation Force Impulse
SSI	-	Supersonic Shear Imaging
ANN	-	Artificial Neural Network
SVM	-	Support Vector Machine
BN	-	Bayesian Network
st-SVM	-	Standard Support Vector Machine
PSVM	-	Proximal Support Vector Machine
NSVM	-	Newton Support Vector Machine
LPSM	-	Lagrangian Support Vector Machine
SSVM	-	Smooth Support Vector Machine
ERM	-	Empirical Risk Minimization
CAD	-	Computer-aided Diagnosis
CSV	-	Comma Separate Value
MLP	-	Multilayer Perceptron
DMBA	-	7,12-Dimethylbenz anthracene
JPEG	-	Joint Photography Experts Group
FFT	-	Fast Fourier Transform
BSC	-	Backscatter Coefficient
MSE	-	Mean Squared Error

LIST OF APPENDICES

APPENDIX	TITLE	PAGE
A	Animal Ethic Approval	119
B	Algorithm of Attenuation and Backscatter Calculation	120
C1	Source Code for Active Contour Segmentation	123
C2	Source Code for Artificial Neural Network (ANN)	126
D	A-Mode Ultrasound Specification	129
E1	B-Mode Ultrasound Specification	131
E2	Digital Oscilloscope Specification	132
F	List of Publications	133

CHAPTER 1

INTRODUCTION

1.1 Research Background

Randomized trials on the benefits of hyperthermia therapy have been demonstrated over the past 20 years [1]. Hyperthermia is a cancer therapy that elevates the tumor to cytotoxic temperatures from 41°C to 45°C to boost in their control of temperature [2, 3]. Hyperthermia has proven to be potential in order to replace the current clinical treatment including radiotherapy and chemotherapy [2, 3, 4-8]. It has been applied in several diseases including breast cancer [5], glioblastoma [9], neck and head cancer [10], hepatocellular carcinoma [11] and lung cancer [12]. However, different type of thermal therapy with high temperature value have also been searched and investigated [13-17].

In 2012, The International Agency for Research in Cancer (GLOBOCAN) reported 5,410 new breast cancer cases in Malaysia, in which 50% of Malaysian women were diagnosed with breast cancer at an early age while 20% of women in developed countries were diagnosed before the age of 50 [18, 19]. Some tumors are very aggressive, grow much faster, and require immediate treatment. The detection of breast masses of breast abnormalities at the earliest stage significantly improved the survival rate of patients as demonstrated by group of investigators [20, 21].

Hence, early treatment of breast cancer is very important at this time to get rid of the cancer and perhaps save lives.

The application of hyperthermia therapy is now growing rapidly in medical field especially for treatment of benign and malignant solid tumors [22] as an innovative and less invasive therapy method. Hyperthermia therapy is one of the cancer treatments that also include surgery, chemotherapy, radiation therapy and immunotherapy [23]. It involves the changes and increase in body temperature over the threshold temperature of an organism [23]. In other words, hyperthermia is a process of delivering heat to cancerous tissue until its durable temperature has been exceeded by expose to heating ablation, including water bath and coagulation of microwave. The treatment could be done successfully if the temperature value of targeted tissues is controlled and monitored within a suitable range [24]. Hyperthermia treatment makes use of artificial heat, usually in the range of 40°C to 60°C, from external sources to destroy cancerous cells or to prevent their further growth [25]. Also, as reported by previous study, the temperature range of hyperthermia therapy is from 40 to 48°C and is maintained at the treated site for a period of one hour or more [26-29]. In some cases, hyperthermia therapy is used as adjuvant therapy [26] with other therapy methods such as radiation to increase the tumor regression rate rather than radiation alone [30].

The lack of precise information available in guiding therapy has been reported to be a major limitation in thermal therapies [2,3,7,31-35]. In order to achieve the goals and competency of current and forthcoming heating devices and systems, the temperature distribution is routinely measured invasively. To meet the distributions of temperature satisfaction for the purpose of assessing thermal dosage appropriately, the limited number of measurement distribution could be avoided [7,32]. Multiple heating devices has boosted the demand for controlling measurement of temperature values that has potential to provide the feedback related to distribution of temperature in details. Thus, it improves the capability and consistency in delivering the heat source by achieving desired temperature value [36-40].

Hyperthermia monitoring with the guidance of Magnetic Resonance Imaging (MRI) is the current gold standard in clinical setting and it is based on the concept of shift of proton resonance frequency. Its function is primarily for surgeons to perform a real-time progression assessment of in-vivo tissue necrosis [41,42]. Although MRI has shown to be capable in visualizing very clear and anatomically correct images, it is found that its performance degrades during hyperthermia treatment due to small temperature interval changes [43]. In contrast to ultrasound system, investigation into clinical trials in hyperthermia therapy using A-Mode has not yet been tested in clinical setting although it is an attractive modality with a simple signal processing approach for data acquisition. Previous study has been focused only on applying heat to study properties of tissue. Hence a comparison between A-Mode and B-Mode ultrasound performance is not reported. However, B-Mode ultrasound is more widely accepted in clinical environment for hyperthermia therapy as compared to A-Mode. Notably, it is sensitive to temperature change, requires simple signal processing and provides good spatial resolution [44]. Additionally B-Mode ultrasound is widely used and capable to evaluate the different kind of tumors and frequently applied in guiding the biopsies especially for breast, liver, thyroid, ovarian, kidney, prostate and uterine [45,46]. Technically, the monitoring approaches of B-Mode are based on variation in acoustic speed, [47-49], energy of backscattered pulsed ultrasound and acoustic nonlinear parameter imaging [49-51]. Although B-Mode system is widely accepted among clinicians, the presence of motion artefacts and degradation of image contrast is the biggest challenge and limitation experienced by B-Mode in hyperthermia therapy, which affects the performance of treatment as a whole.

Therefore, this research was conducted to further search and analyse the potential and its effectiveness of A-Mode ultrasound in monitoring hyperthermia in comparison to B-Mode ultrasound system. Both of this ultrasound system is relevant to be use in this study specifically to monitor temperature during hyperthermia monitoring because this technique is relatively inexpensive, portable and can be easily employed in almost any current heating monitoring system with little concern about system compatibility.

Ultrasound is an attractive medical device compared with the other medical imaging modality, as it offers at low cost, non-ionizing, portable, easy to access and yet compatible [41]. This study will help to improve the overall hyperthermia therapy reliability to be accepted in clinical practice while assisting physicians and medical practitioners in monitoring hyperthermia treatment efficiently with minimal intervention.

1.2 Problem Statement

MRI is the current gold standard in hyperthermia monitoring. It is based on the quantification of shifts in the proton resonance frequency. However, this modality becomes more complex during hyperthermia monitoring and the cost itself is expensive in clinical applications [52,53]. Due to the requirement of significant facilities and capital investment with advancement of complexity heating devices, it limits the current usage though it appear to have spatial resolution satisfaction [54]. In addition, MRI requires capital investment and labour-intensive [43]. Ultrasound is a widely used modality in hyperthermia monitoring, as it is non-ionizing and has capability in giving real-time image acquisition for rapid monitoring of physiological changes and temperature values [55,56]. However, A-Mode ultrasound is not yet tested in clinical setting for hyperthermia therapy monitoring though it has simple requirement for signal processing. Additionally, B-Mode ultrasound is available at low cost, accessible, portable and compatible with other medical imaging modality [57].

Despite the wide acceptance in using B-Mode ultrasound for hyperthermia monitoring, this modality itself experiences certain limitations including insufficient image contrast and motion artefacts [58-63]. To resolve this problem, further investigation on the effectiveness of using A-Mode ultrasound performance in monitoring hyperthermia therapy in comparison to B-Mode ultrasound system was conducted.

In this study, a set of preclinical ex-vivo studies are carried out on both normal and pathological tissues that are harvested from animal models in order to observe the relationship between ultrasonic parameters of A-Mode signals and B-Mode images with the microstructural changes occurring in normal and pathological tissue samples. The subjects were dissected and exposed to hyperthermia at various temperatures of 37°C (body temperature) and 40°C, 45°C, 50°C and 55°C for hyperthermia temperatures. This indicates the progression of tissue necrosis during hyperthermia treatment. Additionally, a classification study is also conducted to compare the performance accuracy of A-Mode with the widely used B-Mode ultrasound technique via ANN in Matlab offline environment in which hyperthermia treatment method has recently displayed the potential to be promising as an adjuvant breast cancer treatment method in clinical setting.

1.3 Research Objectives

The main objective of this research is to investigate the accuracy of A-Mode ultrasound as a hyperthermia monitoring method in comparison to B-Mode ultrasound. Specifically, this study aims are as follows:

- i. To identify the A-Mode and B-Mode ultrasound parameters that are sensitive to changes in temperature and tissue structure
- ii. To compare the performance of A-Mode and B-Mode ultrasound in monitoring tissue denaturation process in hyperthermia treatment using ANN

1.4 Scope of Research

This research has been carried out to study the feasibility of A-Mode ultrasound and to compare its performance in B-Mode ultrasound as a monitoring method in hyperthermia therapy with variation of temperature value from 37°C - 55°C. Specifically, the scope of this study is to identify the main parameter of A-Mode and B-Mode ultrasound that are sensitive to tissue denaturation during hyperthermia.

It involves the measurement of the attenuation and backscatter coefficients for A-Mode and determination of pixel value and standard deviation for B-Mode. The performance percentage of A-Mode and B-Mode ultrasound in monitoring tissue denaturation process in hyperthermia treatment using ANN will then be compared. For this purpose, a series of experiments was conducted on 40 female Sprague Dawley rats in which 30 pathological rats were used as pathological study and 10 healthy rats were used for control group purposes in order to assess the potentiality and the effectiveness on the overall system. The experimental outcomes are fed to signal and image processing for performance measurements and the water bath method is used to mimic the real hyperthermia procedure.

1.5 Significance of Study

In this study, three main purposes that have been addressed. First, this study is conducted to identify the A-Mode and B-Mode parameters that are sensitive to changes in temperature and tissue structure and second, to compare the performance of A-Mode and B-Mode in monitoring tissue denaturation process in hyperthermia treatment using ANN. To date, MRI is the current gold standard in hyperthermia therapy.

Due to the requirement of significant facilities and capital investment with advancement of complexity heating devices, it limits the current usage though it appear to have spatial resolution satisfaction [54]. A-Mode ultrasound is no longer tested for hyperthermia therapy monitoring, although it provides simple signal processing; however, B-Mode ultrasound is widely used for hyperthermia therapy in clinical setting. While it has been widely accepted in clinical environment, B-Mode ultrasound has certain limitations, especially in terms of contrast and motion artefacts of image itself. Hence, further investigation on the effectiveness of using A-Mode ultrasound performance in monitoring hyperthermia therapy in comparison to B-Mode ultrasound system was conducted as solution to this issue.

The signal energy of breast tissue was monitored using A-mode and B-mode scanner before and after heating exposure of the breast at different temperature values of 37°C as body temperature and followed by 40°C, 45°C, 50°C and 55°C, as monitoring temperatures respectively. The success of this study will help to provide a simple and safe method in monitoring local hyperthermia therapy using A-Mode ultrasound.

1.6 Thesis Structure Organization

The outline of this thesis is comprised of five chapters. Chapter 1 discusses the introduction to several primary point of view including research overview, research objectives, problem statement, scope of work, significance of study and as well as thesis outline. Brief reviews and discussions relevant to this study was delivered in chapter 2, especially in breast cancer treatment, hyperthermia therapy and ultrasonic parameter in hyperthermia monitoring. Next, methodological approach that was enrolled in study is explained in detail which includes the modelling processing and designation of system, experimental procedure and ultrasonic data analysis of signal and image processing via Matlab. Results and discussions are presented and discussed

in chapter 4. Finally, summary of all the findings was delivered in chapter 5 with several recommendations that possible could be done in future work.

REFERENCES

1. Van der Zee J. Heating the Patient: A Promising Approach? *Ann Oncol*, 2002. 13:1173-1184..
2. Dewhirst, M., Prosnitz, L., Thrall, D., Prescott, D., Cleff, S., Charles, C., Macfall, J., Rosner, G., Samulski, T., Gillette, E., LaRue, S., Hyperthermic Treatment of Malignant Diseases: Current Status and a View Toward the Future. *Seminars in Oncology*, 1997. 24:616–625.
3. Myerson, R.J., Moros, E., RotiRoti, J.L., Hyperthermia. In: Perez CA, Brady LW, editors. Principles and Practice of Radiation Oncology. 3rd ed. Philadelphia: Lippincott-Raven; 1998. 637–683.
4. Overgaard, J., Gonzales, D., Hulshof, M., Arcangeli, G., Dahl, O., Mella, O., Bentzen, S., Hyperthermia as an adjuvant to radiation therapy of recurrent or metastatic malignant melanoma. A multicenter randomized trial by the European Society for Hyperthermic Oncology. *International Journal of Hyperthermia*, 1996. 12:3–20.
5. Vernon, C., Hand, J., Field, S., Machin, D., Whaley, J., Van der Zee, J., Lez, Liu, Goodman, P., Sherar M. Radiotherapy with or without hyperthermia in the treatment of superficial localized breast cancer: Results from five randomized controlled trials. *International Journal of Radiation Oncology, Biology & Physics*, 1996. 35:731–744.
6. Lee, H.K., Antell, A.G., Perez, C.A., Straube, W.L., Ramachandran, G., Myerson, R.J., Emami, B., Molmenti, E.P., Buckner, A., Lockett, M.A., Specific absorption rate as a predictor of outcome in superficial tumors treated with hyperthermia and radiation therapy. *International Journal of Radiation Oncology, Biology & Physics*, 1998. 40:365–375.
7. Myerson, R.J., Straube, W.L., Moros, E.G., Emami, B.N., Lee, H.K., Perez, C.A., Taylor, M.E., Simultaneous superficial hyperthermia and external radiotherapy: report of thermal dosimetry and tolerance to treatment. *International Journal of Hyperthermia*, 1999. 15:251–266.

8. Van der Zee, J., Gonzalez, D., Van Rhoon, G., Van Dijk, J., Van Putten, W., Hart, A., Comparison of radiotherapy alone with radiotherapy plus hyperthermia in locally advanced pelvic tumours: A prospective, randomised, multicentre trial. *The Lancet*, 2000. 355:1119–1125.
9. Sneed, P.K., Stauffer, P.R., McDermott, M.W., et al., Survival benefit of hyperthermia in a prospective randomized trial of brachytherapy boost +/- hyperthermia for glioblastoma multiforme. *Int J Radiat Oncol Biol Phys*, 1998. 40:287–295.
10. Moffat, F.L., Rotstein, L.E., Calhoun, K., et al., Palliation of advanced head and neck cancer with radiofrequency hyperthermia and cytotoxic chemotherapy. *Can J Surg*, 1984. 27:38–41.
11. Kim, S.K., Lim, H.K., Kim, Y.H., et al., Hepatocellular carcinoma treated with radio-frequency ablation: spectrum of imaging findings. *Radiographics*, 2003. 23:107–121.
12. Simon, C.J., Dupuy, D.E., DiPetrillo, T.A., et al., Pulmonary radiofrequency ablation: long-term safety and efficacy in 153 patients. *Radiology*, 2007. 243:268–275.
13. Hynynen, K., Chung, A., Fjield, T., Buchanan, M., Daum, D., Colucci, V., Lopath, P., Jolesz, F., Feasibility of using ultrasound phased arrays for MRI monitored noninvasive surgery. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, 1996. 43:1043–1053.
14. Ter Haar, G.R., High intensity focused ultrasound for the treatment of tumors. *Echocardiography*, 2001. 18:317–322.
15. Sherar, M.D., Gertner, M.R., Yue, C.K., O'Malley, M.E., Toi, A., Gladman, A.S., Davidson, S.R., Trachtenberg, J., Interstitial microwave thermal therapy for prostate cancer: Method of treatment. *Journal of Urology*, 2001. 166:1707–1714.
16. McCann, C., Kumaradas, J.C., Gertner, M.R., Davidson, S.R., Dolan, A.M., Sherar, M.D., Feasibility of salvage interstitial microwave thermal therapy for prostate. *Physics in Medicine and Biology*, 2003. 48:1041–1052.
17. Tempany, C.M., Stewart, E.A., McDannold, N., Quade, B.J., Jolesz, F.A., Hynynen, K., MR imaging-guided focused ultrasound surgery of uterine leiomyomas. *Radiology*, 2003. 226:897–905.

18. Ferlay, J., Soerjomataram, I., Ervik, M., Dikshit, R., et al., GLOBOCAN 2012 : Cancer Incidence and Mortality Worldwide. *Lyon, Fr. Int. Agency Res. Cancer*. 2013, 11.
19. Yip, C.H., Pathy, N.B., Teo, S.H., A Review of Breast Cancer Research in Malaysia, 2014. 69:8–22.
20. Tabar, L., Vitak, B., Chen, H.H.T., Yen, M.F., Duffy, S.W., and Smith, R.A., Beyond Randomized Trials: Organized Mammographic Screening Substantially Reduces Breast Carcinoma Mortality. *Cancer*, 2001. 91:1724 – 1731.
21. Andersson, I., and Ryden, S., Early Detection and Prevention: Benefits, Costs and Limitations of Screening. Breast Cancer: New Horizons in Research and Treatment, Tobias, J.S., Houghton, J., and Henderson, I.C. eds. London: *Arnold*, 2001. 105-117.
22. Kennedy, J.E., High-intensity focused ultrasound in the treatment of solid tumours. *Nature Reviews: Cancer*, March 2005. 5(4): 321–327.
23. Dewhurst, M.W., Gibbs, F.A., Jr., Roemer, R.B., Samulski, T.V., Gunderson, L.L., Tepper, J.E., Hyperthermia in cancer treatment. 1st ed. Chapter 14. New York, NY: Churchill Livingstone, *Clinical Radiation Oncology*, 2000. 256–82
24. Bailey, M.R., Khokhlova, V.A., Sapozhnikov, O.A., Kargl, S.G., Crum, L.A., Physical mechanisms of the therapeutic effect of ultrasound (a review). *Acoust Phys*, 2003. 49(4):369–388.
25. Chicheł, A., Skowronek, J., Kubaszewska, M., Kanikowski, M., Hyperthermia: Description of a method and a review of clinical applications. *Reports Pract. Oncol. Radiother*, 2007. 12:267–275.
26. J Sneed, P.K., Stauffer, P.R., Li, G.C., Stege, G.J.J., Hyperthermia. In: Leibel SA, Phillips TL (eds.). *Textbook of Radiation Oncology*. 2nd ed. Chapter 70. *Saunders*, 2004. 1569–96.
27. Wust, P., Hildebrandt, B., Sreenivasa, G., et al., Hyperthermia in combined treatment of cancer. *Lancet Oncol*, 2002. 3: 487–497.
28. Falk, M.H., Issels, R.D., Hyperthermia in oncology. *Int J Hyperthermia*, 2001. 17(1): 1–18.

29. Overgaard, J., Horsman, M.R., Hyperthermia. In: Steel, G.G., Basic clinical radiobiology. 2nd ed. Edward Arnold, 1997, 212.
30. Kim, J.H., Hahn, E.W., Ahmed, S.A., Combination hyperthermia and radiation therapy for malignant melanoma. *Cancer*, 1982. 50(3): 478–82.
31. Dewhirst, M., Sim, D., Sapareto, S., Coner, W., Importance of minimum tumor temperature in determining early and long-term responses of spontaneous canine and feline tumors to heat and radiation. *Cancer Research*, 1984. 44:43–50.
32. Myerson, R., Perez, C., Emami, B., Straube, W., Kuske, R., Leybovich, L., VonGerichten, D., Tumor control in longterm survivors following superficial hyperthermia. *International Journal of Radiation Oncology, Biology & Physics*, 1990. 18:1123–1129.
33. Leopold, K., Dewhirst, M., Samulski, T., Dodge, R., George, S., Blivin, J., Progsnitz, L., Oleson, J., Relationships among tumor temperature, treatment time, and histopathological outcome using preoperative hyperthermia with radiation in soft tissue sarcomas. *International Journal of Radiation Oncology, Biology & Physics*, 1992. 22:989–998.
34. Hand, J., Machin, D., Vernon, C., Whaley, J., Analysis of thermal parameters obtained during phase III trials of hyperthermia as an adjunct to radiotherapy in the treatment of breast carcinoma. *International Journal of Hyperthermia*, 1997. 13:343–364.
35. Dewhirst, M.W., Sneed, P.K., Those in gene therapy should pay closer attention to lessons from hyperthermia. *International Journal of Radiation Oncology, Biology & Physics*, 2003. 57:597–600.
36. Underwood, H.R., Burdette, E.C., Ocheltre, K.B., Magin, R.L., A multi element ultrasonic hyperthermia applicator with independent element control. *International Journal of Hyperthermia*, 1987. 3:257–267.
37. Samulski, T.V., Grant, W.J., Oleson, J.R., Leopold, K.A., Dewhirst, M.W., Vallario, P., Blivin, J., Clinical experience with a multi-element ultrasonic hyperthermia system: Analysis of treatment temperatures. *International Journal of Hyperthermia*, 1990. 6:909–922.
38. Stauffer, P., Rossetto, F., Leoncini, M., Gentilli, G., Radiation patterns of dual concentric conductor microstrip antennas for superficial hyperthermia. *IEEE Transactions on Biomedical Engineering*, 1998. 45:605–613.
39. Moros, E.G., Fan, X., Straube, W.L., Experimental assessment of power and temperature penetration depth control with a dual frequency ultrasonic system. *Medical Physics*, 1999. 26:810–817.

40. Novak, P., Moros, E.G., Straube, W.L., Myerson, R.J., SURLAS: a new clinical grade ultrasound system for sequential or concomitant thermos radiotherapy of superficial tumors: applicator description. *Medical Physics*, 2005. 32:230–240.
41. McDannold, N., Quantitative MRI-based temperature mapping based on the proton resonant frequency shift: review of validation studies. *Int. J. Hyperthermia*, 2005. 21: 533–546.
42. Bolomey, J.C., LeBihan, D., Mizushima, S., Recent trends in noninvasive thermal control. In: Seegenschmiedt, M.H., Fessenden, P., Vernon, C.C., (eds) Principles and practice of thermoradiotherapy and thermochemotherapy. Recent trends in noninvasive thermal control thermo-radiotherapy and thermochemotherapy. *Springer, Heidelberg*, 1995. 1:361–379.
43. Zhou, Y.F., Syed Arbab, A., Xu, R.X., High intensity focused ultrasound in clinical tumor ablation. *World J Clin Oncol*, 2011. 2:8–27.
44. Bazan, I., Vazquez, M., Ramos, A., Vera, A., Leija, L., A performance analysis of echographic ultrasonic techniques for non-invasive temperature estimation in hyperthermia range using phantoms with scatterers. *Ultrasonics*, 2009. 49:358–376.
45. Damianou, C.A., Sanghvi, N.T., Fry, F.J., MaassMoreno R. Dependence of ultrasonic attenuation and absorption in dog soft tissues on temperature and thermal dose. *Journal of the Acoustical Society of America*, 1997. 102:628–634.
46. Worthington, A.E., Trachtenberg, J., Sherar, M.D., Ultrasound properties of human prostate tissue during heating. *Ultrasound in Medicine & Biology*, 2002. 28:1311–1318.
47. Liu, X., Gong, X., Yin, C., Li, J., Zhang, D., Non-invasive estimation of temperature elevations in biological tissues using acoustic non-linearity parameter imaging. *Ultrasound Med Biol*, 2008. 34:414-424.
48. Clarke, R.L., Bush, N.L., Haar, G.R., The changes in acoustic attenuation due to in vitro heating. *Ultrasound in Medicine & Biology*, 2003. 29:127–135.
49. Techavipoo, U., Varghese, T., Chen, Q., Stiles, T.A., Zagzebski, J.A., Frank, G.R., Temperature dependence of ultrasonic propagation speed and attenuation in excised canine liver tissue measured using transmitted and reflected pulses. *Journal of the Acoustical Society of America*, 2004. 115:2859–2865.
50. Ribault, M., Chapelon, J., Cathignol, D., Gelet, A., Differential attenuation imaging for the characterization of high intensity focused ultrasound lesions. *Ultrasonic Imaging*, 1998. 20:160–177.

51. Straube, W.L., Arthur, R.M., Theoretical estimation of the temperature dependence of backscattered ultrasonic power for noninvasive thermometry. *Ultrasound in Medicine and Biology*, 1994. 20:915–922.
52. Carter, D., Mc Fall, J., Clegg, S.T., Wan, X., Prescott, D., Charles, H., Samulski, T., Magnetic resonance thermometry during hyperthermia for human high grade sarcoma. *Int J Radiat Oncol Biol Phys*, 1998. 4(4):815–822.
53. Nadobny, J., Wlodarczyk, W., Westhoff, L., Gellermann, J., Felix, R., Wust, P., A clinical water coated antenna applicator for MR-controlled deep-body hyperthermia. A comparison of calculated and measured 3D temperature data sets. *IEEE Trans Biomed Eng*, 2005. 52(3):505–519.
54. Arthur, R.M., Basu, D., Guo, Y., Trobaugh, J.W., Moros, E.G., 3D in vitro estimation of temperature using the change in backscattered ultrasonic energy. *IEEE Trans Ultrason Ferroelectr Freq Control*, 2010. 57(8):1724-33.
55. Orsi, F., Zhang, L., Arnone, P., Bonomo, G., Della Vigna P., Monfardini, L., et al., High intensity focused ultrasound (HIFU) ablation: effective and safe therapy for solid tumors at difficult locations. *AJR Am J Roentgenol*, 2010. 195(3):245-52.
56. Zhang, L., Zhu, H., Jin, C., Zhou, K., Li, K., Su, H., et al., High intensity focused ultrasound (HIFU): effective and safe therapy for hepatocellular carcinoma adjacent to major hepatic veins. *Eur Radiol*, 2009. 19:437–445.
57. Li Sheng., Pei Hong Wu., Magnetic resonance image guided versus ultrasound-guided high-intensity focused ultrasound in the treatment of breast cancer. *Chin J Cancer*, 2013. 32(8):441–452.
58. Hill, C.R., ter Haar, G.R., High intensity focused ultrasound potential for cancer treatment. *Br J Radiol*, 1995. 68:1296–1303.
59. Hsiao, Y.S., Focused ultrasound thermal therapy monitoring using ultrasound, infrared thermal, and photoacoustic imaging techniques (Ph.D. dissertation at the University of Michigan) 2013.
60. Hynynen, K., Vykhodtseva, N.I., Chung, A., Sorrentino, V., Colucci, V., Jolesz, F.A., Thermal effects of focused ultrasound on the brain: determination with MR imaging. *Radiology*, 1997. 204:247–253.
61. Rangraz, P., Behnam, H., Tayakkoli, J., Nakagami., Imaging for detecting thermal lesions induced by high-intensity focused ultrasound in tissue. *Proc Inst Mech Eng, Part H: J Eng Med*, 2014. 228(1):19-26.
62. Haar, G.R., Ultrasound focal beam surgery. *Ultrasound Med Bio*, 1995. 21:1089-1100.

63. Zhang, S., Wan, M., Zhong, H., Xu, C., Liao, Z., Liu, H., et al., Dynamic changes of integrated backscatter, attenuation coefficient and bubble activities during high intensity focused ultrasound (HIFU) treatment. *Ultrasound Med Biol*, 2009. 35(11): 1828-1844.
64. Zipes, D.P., Haissaguerre, M., (eds). Catheter Ablation of Arrhythmias. 2nd ed. New York, NY: Futura Publishing Co, 2002.
65. Manaf, NA., Aziz, MN., Ridzuan, DS., Mohamad Salim, MI., Wahab, AA., Lai, KW., Hum, YC., Feasibility of A-mode Ultrasound Attenuation as a Monitoring Method of Local Hyperthermia Treatment. *Journal of Medical and Biological Engineering and Computing*, 2016. 54:967-981.
66. Sullivan, L.D., McLoughlin, M.G., Goldenberg, L.G., Gleave, M.E., Marich, K., Early experience with high-intensity focused ultrasound for the treatment of benign prostatic hypertrophy. *Br J Urol*, 1997. 79:172-176.
67. Elham, S., Michael, K., Transmission Ultrasound Imaging to guide Thermal Therapy. *IEEE Ultrasonics Symposium*, 2007. 1812-1815.
68. Z. Zhao., F. Wu., Minimally-invasive thermal ablation of early-stage breast cancer: A systemic review. *Journal of Cancer Surgery*, 2010. 36:1149-1155.
69. Kudo, M., Local ablation therapy for hepatocellular carcinoma: current status and future perspective. *J. Gastroenterol*, 2004. 39:205-214.
70. Chiou, S.Y., Liu, J.B., Needleman, L., Current status of sonographically guided radiofrequency ablation techniques. *J Ultrasound Med*, 2007. 26:487-499.
71. Jeffrey, S.S., Birdwell, R.L., Ikeda, D.M., et al., Radiofrequency ablation of breast cancer: first report of an emerging technology. *Arch Surg*, 1999. 134:1064-8.
72. Izzo, F., Thomas, R., Delrio, P., et al., Radiofrequency ablation in patients with primary breast carcinoma: a pilot study in 26 patients. *Cancer*, 2001. 92:2036-44.
73. Burak Jr, W.E., Agnese, D.M., Povoski, S.P., et al., Radiofrequency ablation of invasive breast carcinoma followed by delayed surgical excision. *Cancer*, 2003. 98:1369-76.
74. Hayashi, A.H., Silver, S.F., van der Westhuizen, N.G., et al., Treatment of invasive breast carcinoma with ultrasound-guided radiofrequency ablation. *Am J Surg*, 2003. 185:429-35.
75. Fornage, B.D., Sneige, N., Ross, M.I., et al., Small (< or ¼ 2-cm) breast cancer treated with US-guided radiofrequency ablation: feasibility study. *Radiology*, 2004. 231:215-24.

76. Earashi, M., Noguchi, M., Motoyoshi, A., Fujii, H., Radiofrequency ablation therapy for small breast cancer followed by immediate surgical resection or delayed mammotome excision. *Breast Cancer*, 2007. 14:39–47.
77. Imoto, S., Wada, N., Sakemura, N., Hasebe, T., Murata, Y., Feasibility study on radiofrequency ablation followed by partial mastectomy for stage I breast cancer patients. *Breast*, 2009. 18:130–4.
78. Medina Franco, H., Soto Germes, S., Ulloa Go´mez, J.L., et al., Radiofrequency ablation of invasive breast carcinomas: a phase II trial. *Ann Surg Oncol*, 2008. 15:1689–95.
79. Manenti, G., Bolacchi, F., Perretta, T., et al., Small breast cancers: in vivo percutaneous US-guided radiofrequency ablation with dedicated cool tip radiofrequency system. *Radiology*, 2009. 251:339-46.
80. Susini, T., Nori, J., Olivieri, S., et al., Radiofrequency ablation for minimally invasive treatment of breast carcinoma. A pilot study in elderly inoperable patients. *Gynecol Oncol*, 2007. 104:304-10.
81. Marcy, P.Y., Magne, N., Castadot, P., Bailet, C., Namer, M., Ultrasound-guided percutaneous radiofrequency ablation in elderly breast cancer patients: preliminary institutional experience. *Br J Radiol*, 2007. 80:267–73.
82. Oura, S., Tamaki, T., Hirai, I., et al., Radiofrequency ablation therapy in patients with breast cancers two centimeters or less in size. *Breast Cancer*, 2007. 14:48-54.
83. Dowlatshahi, K., Francescatti, D.S., Bloom, K.J., Laser therapy for small breast cancers. *Am J Surg*, 2002. 184:359-63.
84. Korourian, S., Klimberg, S., Henry Tillman, R., et al., Assessment of proliferating cell nuclear antigen activity using digital image analysis in breast carcinoma following magnetic resonance-guided interstitial laser photocoagulation. *Breast J*, 2003. 9:409–13.
85. Haraldsdottir, K.H., Ivarsson, K., Gotberg, S., Ingvar, C., Stenram, U., Tranberg, K.G., Interstitial laser thermotherapy (ILT) of breast cancer. *Eur J Surg Oncol*, 2008. 34:739-45.
86. Dowlatshahi, K., Dieschbourg, J.J., Bloom, K.J., Laser therapy of breast cancer with 3-year follow-up. *Breast J*, 2004. 10:240–3.

87. Akimov, A.B., Seregin, V.E., Rusanov, K.V., et al., Nd: YAG interstitial laser thermotherapy in the treatment of breast cancer. *Lasers Surg Med*, 1998. 22:257–67.
88. Chunlan, Y., Hao, Z., Shuicai, W., Yanping, B., Hongjian, G., Correlations Between B-Mode Ultrasonic Image Texture Features and Tissue Temperature in Microwave Ablation. *J Ultrasound Med*, 2010. 29:1787–1799.
89. Gardner, R.A., Vargas, H.I., Block, J.B., et al., Focused microwave phased array thermotherapy for primary breast cancer. *Ann Surg Oncol*, 2002. 9:326–32.
90. Vargas, H.I., Dooley, W.C., Gardner, R.A., et al., Focused microwave phased array thermotherapy for ablation of early-stage breast cancer: results of thermal dose escalation. *Ann Surg Oncol*, 2004. 11:139–46.
91. Vargas, H.I., Dooley, W.C., Gardner, R.A., Gonzalez, K.D., Heywang Ko brunner, S.H., Fenn, A.J., Success of sentinel lymph node mapping after breast cancer ablation with focused microwave phased array thermotherapy. *Am J Surg*, 2003. 186:330–2.
92. A. Szent-Györgyi, Chemical and biological effects of ultra-sonic radiation. *Nature* 131, 1933. 278
93. J.G. Lynn., R.L. Zwemer., A.J. Chick., The biological application of focused ultrasonic waves. *Science*, 1942. 96(2483):119–120.
94. P.D. Wall., W.J. Fry., R. Stephens., D. Tucker., J.Y. Lettvin., Changes produced in the central nervous system by ultrasound. *Science*, 1951. 114(2974): 686–687.
95. W.J. Fry., J.W. Barnard., F.J. Fry., R.F. Krumins., J.F. Brennan., Ultrasonic lesions in the mammalian central nervous system. *Science*, 1955. 122(3168): 517–518.
96. F.J. Fry., H.W. Ades., W.J. Fry., Production of reversible changes in the central nervous system by ultrasound. *Science*, 1958. 127(3289):83–84.
97. F.A. Jolesz., N. McDannold., Current status and future potential of MRI-guided focused ultrasound surgery. *J. Magn. Reson. Imaging*, 2008. 27:391–399.

98. K. Hynynen., MRI-guided focused ultrasound treatments. *Ultrasonics*, 2010. 50:221–229.
99. P.S. Yarmolenko., E.J. Moon., C. Landon., A. Manzoor., D.W. Hochman., B.L. Viglianti., M.W. Dewhirst., Thresholds for thermal damage to normal tissues: an update. *Int. J. Hyperthermia*, 2011. 27:320–343.
100. G. Ter Haar., C. Coussios., High intensity focused ultrasound: physical principles and devices. *Int. J. Hyperthermia*, 2007 23:89–104.
101. N.M. Hijnen., E. Heijman., M.O. Köhler., M. Ylihautala., G.J. Ehnholm., A.W. Simonetti., H. Grüll., Tumour hyperthermia and ablation in rats using a clinical MR-HIFU system equipped with a dedicated small animal set-up. *Int. J. Hyperthermia*, 2012. 28(2):141–155.
102. M. de Smet., E. Heijman., S. Langereis., N.M. Hijnen., H. Grüll., Magnetic resonance imaging of high intensity focused ultrasound mediated drug delivery from temperature-sensitive liposomes: an in vivo proof-of-concept study. *J. Control. Release*, 2011. 150:102–110.
103. A.H. Negussie., P.S. Yarmolenko., A. Partanen., A. Ranjan., G. Jacobs., D. Woods., H. Bryant., D. Thomasson., M.W. Dewhirst., B.J. Wood., M.R. Dreher., Formulation and characterization of magnetic resonance imageable thermally sensitive liposomes for use with magnetic resonance-guided high intensity focused ultrasound. *Int. J. Hyperthermia*, 2011. 27:140–155.
104. Wu, F., Wang, Z.B., Cao, Y.D., et al., A randomised clinical trial of high intensity focused ultrasound ablation for the treatment of patients with localized breast cancer. *Br J Cancer*, 2003. 89:2227–33.
105. Gianfelice, D., Khiat, A., Amara, M., Belblidia, A., Boulanger, Y., MR imaging- guided focused ultrasound surgery of breast cancer: correlation of dynamic contrast-enhanced MRI with histopathologic findings. *Breast Cancer Res Treat*, 2003. 82:93–101.
106. Furusawa, H., Namba, K., Thomsen, S., et al., Magnetic resonance-guided focused ultrasound surgery of breast cancer: reliability and effectiveness. *J Am Coll Surg*, 2006. 203:54–63.
107. Gianfelice, D., Khiat, A., Boulanger, Y., Amara, M., Belblidia, A., Feasibility of magnetic resonance imaging-guided focused ultrasound surgery as an adjunct to tamoxifen therapy in high-risk surgical patients with breast carcinoma. *J Vasc Interv Radiol*, 2003. 14:1275–82.

108. Furusawa, H., Namba, K., Nakahara, H., et al., The evolving non-surgical ablation of breast cancer: MR guided focused ultrasound (MRgFUS). *Breast Cancer*, 2007. 14:55–8.
109. Cline, H.E., Hynynen, K., Schneider, E., et al., Simultaneous magnetic resonance phase and magnitude temperature maps in muscle. *Magn Reson Med*, 1996. 35(3):309–315.
110. Hutchinson, E., Dahleh, M., Hynynen, K., The feasibility of MRI feedback control for intra cavitory phased array hyperthermia treatments. *Int J Hyperthermia*, 1998. 14(1):39-56.
111. James, R., MacFall Brian, J., Soher., MR Imaging in Hyperthermia. *Radio Graphics*, 2007. 27:1809–1818
112. Delannoy, J., LeBihan, D., Hoult, D.I., Levin, R.L., Hyperthermia system combined with a magnetic resonance imaging unit. *Med Phys*, 1990. 17(5): 855–860.
113. Jolesz, F.A., Bleier, A.R., Jakab, P., Ruenzel, P.W., Huttli, K., Jako, G.J., MR imaging of laser-tissue interactions. *Radiology*, 1988. 168(1):249–253.
114. Clegg, S.T., Das, S.K., Zhang, Y., Macfall, J., Fullar, E., Samulski, T.V., Verification of hyperthermia model method using MR thermometry. *Int J Hyperthermia*, 1995. 11(3):409–424.
115. Craciunescu, O.I., Samulski, T.V., MacFall, J.R., Clegg, S.T., Perturbations in hyperthermia temperature distributions associated with counter-current flow: numerical simulations and empirical verification. *IEEE Trans Biomed Eng*, 2000. 47(4):435-443.
116. Craciunescu, O.I., Das, S.K., McCauley, R.L., MacFall, J.R., Samulski, T.V., 3D numerical reconstruction of the hyperthermia induced temperature distribution in human sarcomas using DE-MRI measured tissue perfusion: validation against non-invasive MR temperature measurements. *Int J Hyperthermia*, 2001. 17(3):221-239.
117. MacFall, J.R., Prescott, D.M., Charles, H.C., Samulski, T.V., 1H MRI phase thermometry in vivo in canine brain, muscle, and tumor tissue. *Med Phys*, 1996. 23(10):1775–1782.

118. Craciunescu, O.I., Raaymakers, B.W., Kotte, A.N., Das, S.K., Samulski, T.V., Lagendijk, J.J., Discretizing large traceable vessels and using DE-MRI perfusion maps yields numerical temperature contours that match the MR noninvasive measurements. *Med Phys*, 2001. 28(11):2289–2296.
119. Samulski, T.V., Clegg, S.T., Das, S., MacFall, J., Prescott, D.M., Application of new technology in clinical hyperthermia. *Int J Hyperthermia*, 1994. 10(3):389–394.
120. Samulski, T.V., MacFall, J., Zhang, Y., Grant, W., Charles, C.. Non-invasive thermometry using magnetic resonance diffusion imaging: potential for application in hyperthermic oncology. *Int J Hyperthermia*, 1992. 8(6):819-829.
121. Samulski, T.V., Fessenden, P., Lee, E.R., Kapp, D.S., Tanabe, E., McEuen, A., Spiral microstrip hyperthermia applicators: technical design and clinical performance. *Int J Radiat Oncol Biol Phys*, 1990. 18(1):233-242.
122. Zhang, Y., Joines, W.T., Oleson, J.R., Prediction of heating patterns of a microwave interstitial antenna array at various insertion depths. *Int J Hyperthermia*, 1991, 7(1):197-207.
123. Prescott, D.M., Charles, H.C., Sostman, H.D., et al., Therapy monitoring in human and canine soft tissue sarcomas using magnetic resonance imaging and spectroscopy. *Int J Radiat Oncol Biol Phys*, 1994. 28(2):415-423.
124. Das, S.K., Clegg, S.T., Samulski, T.V., Computational techniques for fast hyperthermia temperature optimization. *Med Phys*, 1999. 26(2):319-328.
125. Sheng Li and Pei Hong Wu. Magnetic resonance image-guided versus ultrasound-guided high-intensity focused ultrasound in the treatment of breast cancer. *Chines Journal of Cancer*, 2013. 32(8): 441–452.
126. Chen, J.W., Huang, T.Y., Peng, H.H., et al., Proton resonance frequency shift-weighted imaging for monitoring MR-guided high-intensity focused ultrasound transmissions. *J Magn Reson Imaging*, 2011. 33:1474–1481.
127. U. Techavipoo, T. Varghese, J. A. Zagzebski, T. Stiles, and G. Frank., Temperature dependence of ultrasonic propagation speed and attenuation in canine tissue. *Ultrason. Imaging*, 2002. 24:246–260

128. Wu, S.C., Bai, Y.P., Ren, X.Y., Current situation home and abroad in the research of noninvasive temperature estimation by ultrasound in cancer hyperthermia. *Chin J Rehabil*, 2004. 17:3424–3425.
129. Jignesh, S., Salavar, R.A., Konstantin, S., et al., Ultrasound imaging to monitor photothermal therapy: feasibility study. *Opt Express*, 2008. 16:3776–3785.
130. Liu, H.L., Li, M.L., Shih, T.C., et al., A novel ultrasonic-imaging based temperature estimation approach by instantaneous frequency detection. In: *Proceedings of the 2008 IEEE International Ultrasonics Symposium. Piscataway, NJ:Institute of Electrical and Electronics Engineers*, 2008. 2984–2087.
131. Zhu, N., Jiang, Y., Kato, S., Ultrasonic computerized tomography (CT) for temperature measurements with limited projection data based on an extrapolated filtered back projection (FBP) method. *Energy*, 2005. 30:509–522.
132. Techavipoo, U., Varghese, T., Zagzebske, J.A., Stiles, T., Frank, G., Temperature dependence of ultrasonic propagation speed and attenuation in canine tissue. *Ultrason Imaging*, 2002. 24:246–260.
133. Anand, A., Kaczkowski, P.J., Noninvasive measurement of local thermal diffusivity using backscattered ultrasound and focused ultrasound heating. *Ultrasound Med Biol* 2008; 34:1449–1464.
134. Trobaugh, J.W., Arthur, R.M., Straube, W.L., Moros, E.G., A simulation model for ultrasonic temperature imaging using change in backscattered energy. *Ultrasound Med Biol*, 2008. 34:289-298.
135. Arthur, R.M., Trobaugh, J.W., Straube, W.L., Moros, E.G., Temperature dependence of ultrasonic backscattered energy in motion compensated images. *IEEE Trans Ultrason Ferroelectr Freq Control*, 2005. 52:1644-1652.
136. Arthur, R.M., Straube, W.L., Trobaugh, J.W., Moros, E.G., In Vivo Change in Ultrasonic Backscattered Energy with Temperature in Motion-Compensated Images”. *Int. J. Hyperthermia*, August 2008. 24(5): 389–398.
137. Lu, Y., Liu, X.Z., Gong, X.F., et al., Relationship between the temperature and the acoustic nonlinearity parameter in biological tissues. *Chin Sci Bull*, 2004. 49:2360–2363.

138. R. Seip., and E. Ebbini., Noninvasive estimation of tissue temperature response to heating fields using diagnostic ultrasound. *IEEE Trans. Biomed. Eng.* 1995. 42:828–839
139. C. Simon., P. VanBaren., and E. Ebbini., Two-dimensional temperature estimation using diagnostic ultrasound, *IEEE Trans. Ultrason., Ferroelect., Freq. Contr.* 1998. 45:1088–1099, 1998.
140. Abolohassani, M., Alirezaie, J., Norouzi, A., Study of fast temperature estimation during hyperthermia using ultrasound digital images. *Signal Process Appl.* 2007.
141. Constant, A., Wright, W.M.D., Estimation of tissue elasticity by image processing of simulated B-Mode ultrasound images. In: *IER Irish signal and systems conference, 2009.* 5.
142. Pouch, A.M., Theodore, W.C., Shultz, S.M., Seghal, C.M., In vivo noninvasive temperature measurement by B-mode ultrasound imaging. *J Ultrasound Med.* 2010. 29(11):1595-1606.
143. Zhang, J., Zhou, S., Brunke, S., Comaniciu, D., Databaseguided breast tumor detection and segmentation in 2D ultrasound images. In: *Proceedings of the SPIE 7624 medical imaging computer aided diagnosis*, 2010.
144. Haar, G.R., Ultrasound focal beam surgery. *Ultrasound Med Bio.* 1995. 21:1089–1100
145. Goksel, O., Salcudean, S.E., B-mode ultrasound image simulation in deformable 3D medium. *IEEE Trans Med Imaging*, 2009. 28(11):1657–1669
146. Rawat, V., Jain, A., Shrimali V., Investigation and assessment of disorder of ultrasound B-mode images. *Int J Comput Sci Inf Secur*, 2010. 7(2).
147. Frizzell, L.A., Carstensen, E.L., Shear properties of mammalian tissues at low megahertz frequencies. *J Acoust Soc Am*, 1976. 60:1409–1411.
148. Gheonea, D.I., Saftoiu, A., New trends in elasticity imaging. *Ultraschall in Med.* 2010. 31(5):525–527.
149. Madsen, E.L., Sathoff, H.J., Zagzebski, J.A., Ultrasonic shear wave properties of soft tissues and tissue like materials. *J Acoust Soc Am*, 1983. 74:1346-1355.

150. C. Le Floch., M. Tanter., and M. Fink., Self-defocusing in Hyperthermia: Experiments and simulations. *Appl. Phys. Lett* ,1999. 74:3062–3064.
151. Arthur, R.M., Trobaugh, J.W., Straube, W.L., Moros, E.G., Sangkatumvong, S., Non-invasive Estimation of Hyperthermia Temperatures with Ultrasound. *Int. J. Hyperthermia*, September 2005. 21(6): 589–600.
152. Mladina, R., Risavi, R., Branica, S., Heinzl, B., A-mode diagnostic ultrasound of maxillary sinuses: possibilities and limitations. *Rhinology*, 1994. 32(4):179–183.
153. Craig C.F., How Ultrasounds Work, Available at: http://www.physics.utoronto.ca/~jharlow/teaching/phy138_0708/lec04/ultrasoundx.htm; Accessed on: September 2015.
154. Gallotti, A., D’Onofrio, M., Mucelli, R.P., Acoustic Radiation Force Impulse (ARFI) technique in ultrasound with Virtual Touch tissue quantification of the upper abdomen. *Radiol Med (Torino)*, 2010. 115(6):889–897.
155. Bota, S., Sporea, I., Sirli, R., Popescu, A., Danila, M., Costachescu, D., Intra- and interoperator reproducibility of acoustic radiation force impulse (ARFI) elastography-preliminary results. *Ultrasound Med Biol*, 2012. 38(7):1103–1108.
156. Lee, S., Kim, D.Y., Non-invasive diagnosis of hepatitis B virus-related cirrhosis. *World J Gastroenterol*, 2004. 20(2):445–459.
157. Towa, R.T., Miller, R.J., Frizzell, L.A., Zachary, J.F., O’Brien, W.D., Jr., Attenuation coefficient and propagation speed estimates of rat and pig intercostal tissue as a function of temperature. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control* 2002. 49:1411–1420.
158. Robinson, T.C., Lele, P.P., An analysis of lesion development in the brain and in plastics by high intensity focused ultrasound at low-megahertz frequencies. *Journal of the Acoustical Society of America*, 1972. 5:1333–1351.
159. Bamber, J.C., Hill, C.R., Ultrasonic attenuation and propagation speed in mammalian tissues as a function of temperature. *Ultrasound in Medicine and Biology*, 1979. 5:149-157.

160. J. Fessler., Reflection-mode ultrasound imaging, 2009. 11:18. <http://web.eecs.umich.edu/~fessler/course/516/l/cu-ultra.pdf>, accessed June 2015.
161. Getner, Mr., Wilson, B.C., Sherar, M.D., Ultrasound imaging of thermal therapy in in-vitro liver. *Ultrasound Med Biol*, 1998. 24:1023-1032.
162. Jeremy, P.K., and Michael L.O., Ultrasonic Assessment of Thermal Therapy in rat liver. *Electrical and Computer Engineering*, 2012. 38: 2130 -2137.
163. Xu. T., Ye, Z.Q., Cai, W.M., Texture analysis methods used in B scan liver images. *Foreign Med Sci Fascicle Biomed Eng*, 2006. 29:228–230.
164. Udomchai, T., Quan, C., Varghese, T., Ultrasonic noninvasive temperature estimation using echo shift gradient maps: simulation results. *Ultrason Imaging*, 2005. 27:166–180.
165. Pousek, L., Jelinek, M., Storkova, B., et al., Noninvasive temperature monitoring using ultrasound tissue characterization method. In: *Proceedings of the 2006 Information Technology Interfaces Conference*. Piscataway, NJ: Institute of Electrical and Electronics Engineers, 2006. 219–224.
166. Ren, X.Y., Wu, S.C., Zeng, Y., Experimental study for noninvasive monitoring of hyperthermia based on B-mode ultrasonic tissue characterization. *J Beijing Univ Technol*, 2008. 34: 90–96.
167. Wu, X., Qian, S.Y., Sun, F.C., et al., Study of noninvasive temperature measurement based on B-mode ultrasonic image processing. *Comput Eng Appl*, 2007. 43:178–179.
168. Wu, N., Yang, W., Chen, L., et al., Texture analysis of B-mode ultrasound images on tissue in radiofrequency ablation. *Chin J Med Imaging Technol*, 2007. 23:140–143.
169. Ren X., Wu S., Zeng Y., Noninvasive Monitoring for Hyperthermia Based on Ultrasonic Tissue Characterization of B-mode. *IEEE*, 2007. 1173-1176
170. S. Siebers., M. Schwabe., U. Scheipers., C. Welp., J. Werner., H. Ermert., Evaluation of Ultrasonic Texture and Spectral Parameters for Coagulated Tissue Characterization. *IEEE International Ultrasonics, Ferroelectrics, and Frequency Control Joint 50th Anniversary Conference*, 2004. 1804-1807

171. T. Douglas Mast., Daniel, P., Pucke, Swetha, E., Subramanian, William J. Bowlus, Steven M. Rudich, Joseph F. Buell. Ultrasound Monitoring of In Vitro Radio Frequency Ablation by Echo Decorrelation Imaging. *J Ultrasound Med*, 2008. 27:1685-1697.
172. Woo Kyung Moon., Chiun-Sheng Huang., Wei-Chih Shen., Etsuo Takada., Ruey-Feng Chang., Juliwati Joe, Michiko Nakajima, and Masayuki Kobayashi, Analysis of Elastographic and B-mode Features at Sonoelastography for Breast Tumor Classification. *Ultrasound in Med. & Biol*, 2009. 35:1794–1802.
173. Haralick, R. M., Shanmugam, K. and Dinstein, I., Textural features for image classification, *IEEE Trans. Syst. Man Cyber*, 1973. 3:610-621.
174. Basset, O., Sun, Z., Mestas, J. L., and Gimenez, G., Texture analysis of ultrasonic images of the prostate by means of co-occurrence matrices, *Ultrasonic Imaging*, 1993. 15:218-237.
175. Nicholas, D., Nassiri, D., Garbutt, P. and Hill, C., Tissue characterization from ultrasound B-scan data, *Ultrasound Med. Biol*, 1986. 12:135-143.
176. Morris, D. T., An evaluation of the use of texture measurements for the tissue characterization of ultrasonic images of in vivo human placentae, *Ultrasound Med. Biol*, 1988. 14:387-395.
177. McPherson, D., Aylward, P., Knosp, B., et al., Ultrasound characterization of acute myocardial ischemia by quantitative texture analysis, *Ultrasonic Imaging*, 1986. 8:227-240.
178. Wu, C., Chen, Y. and Hsieh, K., Texture features for classifications of ultrasonic liver images, *IEEE Trans. Med. Imag*, 1992. 11:141-152.
179. Layer, G., Zuna, I., Lorentz, A., et al., Computerized ultrasound B-scan texture analysis of experimental fatty liver disease: influence of total lipid content and fat deposit distribution, *Ultrasonic Imaging*, 1990. 12:171-188.
180. Haberkorn, U., Zuna, I., Lorentz, A., et al., Echographic tissue characterization in diffuse parenchymal liver disease: correlation of image structure with histology, *Ultrasonic Imaging*, 1990. 12:155-170 (1990).
181. Parkkinen, J., and Oja, E., Detecting texture periodicity from the co-occurrence matrix, *Patt. Recog. Letts*, 1990. 11:43-50.

182. Konstantina Kourou , Themis P. Exarchos , Konstantinos P. Exarchos , Michalis V. Karamouzis, Dimitrios I. Fotiadis, Machine learning applications in cancer prognosis and prediction. *Computational and Structural Biotechnology Journal*, 2015. 13:8-17.
183. Ahmad Taher Azar., Shaimaa Ahmed El-Said., Performance analysis of support vector machines classifiers in breast cancer mammography recognition. *Neural Comput & Applic*, 2014. 24:1163–1177.
184. Blanz, V., Scholkopf, B., Bulthoff, H., et al., Comparison of view based object recognition algorithms using realistic 3d models. In: von der Malsburg, C., von Seelen, W., Vorbruggen, J.C., Sendhoff, B., (eds) *Artificial Neural Networks-ICANN'96, Springer Lecture Notes in Computer Science, Berlin*, 1996. 1112: 251–256
185. Joachims, T., Nedellec, C., Rouveirol, C., Text categorization with support vector machines: learning with many relevant. *Springer, Springer-Verlag GmbH, Berlin*, 1998.
186. Osuna, E., Freund, R., Girosit, F., Training support vector machines: an application to face detection. *Proceedings of IEEE Computer Society Conference on Computer Vision and Pattern Recognition*, June 17–19, 1997. 130–136.
187. Burges, C.J.C, Scholkopf, B., Improving the accuracy and speed of support vector learning machines. In: Mozer, M., Jordan, M., Petsche, T., (eds) *Advances in neural information processing systems 9. MIT Press, Cambridge*, 1997. 375–381
188. Schmidt., Identifying speaker with support vector networks. *Interface '96 Proceedings, Sydney*, 1996.
189. Cortes, C., Vapnik, V., Support vector network. *Mach Learn*, 1995. 20:273–297.
190. Scholkopf, B., Burges, C., Vapnik, V., Extracting support data for a given task. In: Fayyad, U.M., Uthurusamy, R., (eds) *Proceedings, first international conference on knowledge discovery & data mining. AAAI Press, Menlo Park*, 1995.
191. Scholkopf, B., Burges, C., Vapnik, V., (1996) Incorporating invariances in support vector learning machines. In: von der Malsburg, C., von Seelen, W.,

- Vorbruggen, J.C., Sendhoff, B., (eds) *Artificial neural networks- ICANN'96. Springer Lecture Notes in Computer Science, Berlin, 1996.* 1112:47–52.
192. Vapnik, V., Golowich, S., Smola, A., Support vector method for function approximation, regression estimation, and signal processing. In: Mozer M, Jordan M, Petsche T (eds) *Advances in neural information processing systems 9. MIT Press, Cambridge, 1997.* 281–287.
 193. Mangasarian, O.L., Generalized support vector machines. In: Smola, A., Bartlett, P., Scholkopf, B., Schuurmans, D., *Advances in large margin classifiers. MIT Press, Cambridge, 2000.* 135–146.
 194. Vapnik, V.N., *The nature of statistical learning theory, 2nd edn.* New York, Springer, 1999.
 195. Fung, G., Mangasarian, O.L., Proximal support vector machine classifiers. *International Conference on Knowledge Discovery and Data Mining, San Francisco, 2001.* 77-86.
 196. Fung, G., Mangasarian, O.L., Finite Newton method for Lagrangian support vector machine classification. *Neuro computing, 2003.* 55(1-2):39-55.
 197. Fung, G., Mangasarian, O.L., A feature selection Newton method for support vector machine classification. *Comput Optim Appl, 2004.* 28(2):185-202.
 198. Lee, Y.J., Mangasarian, O.L., SSVM : a smooth support vector machine. *Comput Optim Appl, 2001.* 20:5-22.
 199. Gunn, S.R., Support vector machines for classification and regression. *Technical Report, Faculty of Engineering, University of Southampton, 1998.*
 200. Bishop, C., *Neural networks for pattern recognition. Clarendon Press, Oxford, 1997.*
 201. Burges, C.J.C., A tutorial on support vector machines for pattern recognition. *Data Min Knowl Discov, 1998.* 2(2):121–167
 202. Chang, R.F., Wu, W.J., Moon, W.K., et al., Support vector machines for diagnosis of breast tumors on US images. *Acad Radiol, 2003.* 10(2):189–197

203. Liu, H.X., Zhang, R.S., Luan, F., et al., Diagnosing breast cancer based on support vector machines. *J Chem Inf Comput Sci*, 2003. 43(3):900–907.
204. Vapnik, V.N., The nature of statistical learning theory. *Springer*, New York, 1995
205. Chen, H.L., Yang, B., Wang, G., et al., Support vector machine based diagnostic system for breast cancer using swarm intelligence. *J Med Syst*, 2012. 36(4):2505-2519
206. Cristianini, N., Taylor, J.S., An introduction to support Vector Machines: and other kernel-based learning methods. *Cambridge University Press, Cambridge*, 2000.
207. Taylor, J.S., Cristianini, N., Kernel methods for pattern analysis. *Cambridge University Press, Cambridge*, 2004.
208. Ng, E.Y.K., Ung, L.N., Ng, F.C., Sim, L.S.J., Statistical analysis of healthy and malignant breast thermography. *J. Med. Eng. Technol*, 2001. 25:253-263.
209. Jiang, J., Trundle, P., Ren, J., Medical image analysis with artificial neural networks. *Comput. Med. Imaging Graph*, 2010. 34:617-631.
210. Al-Shayea, Q., Artificial Neural Networks in Medical Diagnosis. *J. Comput. Sci. Issues*, 2011. 8:150-154.
211. Whitley, D., Starkweather, T., Bogart, C., Genetic algorithms and neural networks: optimizing connections and connectivity. *Parallel Comput*, 1990. 14:347-361.
212. Lisboa, P.J., Taktak, A.F.G., The use of artificial neural networks in decision support in cancer: a systematic review. *Neural Netw*, 2006. 19:408-415.
213. Amato, F., López, A., Peña-Méndez, E.M., Vañhara, P., et al., Artificial neural networks in medical diagnosis. *J. Appl. Biomed*, 2013. 11:47–58.
214. Ibrahim, F., Faisal, T., Salim, M.I., Taib, M.N., Non-invasive diagnosis of risk in dengue patients using bioelectrical impedance analysis and artificial neural network. *Med Biol Eng Comput*, 2010. 48:1141–1148.

215. U Beyli, D.E., Implementing Automated Diagnostic System for Breast Cancer Detection. *Expert System with Application*, 2007. 33:1054-1062.
216. Kubat, M., Neural networks: a comprehensive foundation by Simon Haykin, Macmillan, 1994. *Knowl. Eng. Rev*, 1999. 13:409-412.
217. Wei, J.T., Zhang, Z., Barnhill, S.D., Madyastha, K.R., et al., Understanding artificial neural networks and exploring their potential applications for the practicing urologist. *Urology*, 1998. 52:161-172.
218. Haykin, S.S., Neural Networks: A Comprehensive Foundation, New York; Prentice Hall, 1999.
219. Sun, M., and Scwabasi, R.J., The Forward EEG Solutions can be compared using Artificial Neural Networks. *IEEE Transactions on Biomedical Engineering*, 2000. 47: 1044-1050.
220. Lisboa, P.J.G., A review of evidence of health benefit from artificial neural networks in medical intervention. *Neural Networks*, 2002. 15:11-39.
221. Salim, M.I., Ahmad, A.H., Ariffin, I., Rosidi, B., Supriyanto, E., Development of breast cancer diagnosis tool using hybrid magnetoacoustic method and artificial neural network. *Int J Biol Biomed Eng*, 2012. 6:61-68.
222. Matlab. *Matlab ANN Toolbox*, 2010.
223. Boreham, D.R., Gasmann, H.C., Mitchel, R.E., Water bath hyperthermia is a simple therapy for psoriasis and also stimulates skin tanning in response to sunlight. *Int J Hyperthermia*, 1995. 11(6):745-54.
224. Gao, F., Ye, Y., Yang, J., Water bath hyperthermia reduces stemness of colon cancer cells. *Clin Biochem*, Nov 2013. 46(16-17):1747-54.
225. R.B. Gear., M. Yan., J. Schneider., P. Succop., S.C. Heffelfinger., and D.J. Clegg., Charles River Sprague Dawley Rats Lack Early Age-Dependent Susceptibility to DMBA-Induced Mammary Carcinogenesis. *International Journal of Biological Sciences*, 2007. 3(7):408-416.

226. Russo, I.H., Russo, J.. Developmental stage of the rat mammary gland as determinant of its susceptibility to 7,12-dimethylbenz(a)anthracene. *J Natl Cancer Inst*, 1978. 61:1439-1442.
227. Russo, I., Russo, J., Mammary gland neoplasia in long-term rodent studies. *Environmental Health Perspectives*, 1996. 104:938-966.
228. Isaacs, J., Genetic control of resistance to chemically induced mammary adeno carcinogenesis in the rat. *Cancer Res*, 1986. 46:3958-3963.
229. Miyata, M., Furukawa, M., Takahashi, K., Gonzalez, F.J., Yamazoe, Y., Mechanism of 7,12-Dimethylbenz[a]anthracene Induced Immuno-toxicity : Role of Metabolic Activation at the Target Organ. *Jpn J Pharmacol*, 2001. 86:302-309.
230. M. F. Dias., E. Sousa., S. Cabrita., J. Patrício., and C. F. Oliveira., Chemoprevention of DMBA-induced mammary tumors in rats by a combined regimen of alpha-tocopherol, selenium, and ascorbic acid, *Breast Journal*, 2000. 6(1):14–19.
231. Muggins, C., Briziarelli, G., and Button, H., Jr. Rapid induction of mammary carcinoma in the rat and the influence of hormones on the tumors. *J. Exp. Med*, 1959. 709:25-42.
232. Landini, L., Sarnelli, R., Evaluation of the attenuation coefficients in normal and pathological breast tissue. *Med. Biol. Eng. Comput*, 1986. 24:243–247.
233. Mortensen, C.L., Edmonds, P.D., Gorf, Y., Hill, J.R., et al., Ultrasound tissue characterization of breast biopsy specimens: expanded study. *Ultrason. Imaging*, 1996. 18:215–3011.
234. Chen, X., Phillips, D., Schwarz, K.Q., Mottley, J.G., Parker, K.J., The measurement of backscatter coefficient from a broadband pulse-echo system: A

- new formulation. *IEEE Trans. Ultrason. Ferroelectr. Freq. Control*, 1997. 44:515–525.
235. F. Leymarie., and M. D. Levine., Tracking deformable objects in the plane using an active contour model. *IEEE Trans. Pattern Anal. Machine Intell*, 1993. 15:617-634.
236. R. Durikovic., K. Kaneda., and H. Yamashita., Dynamic contour: A texture approach and contour operations. *Vis. Comput*, 1995. 11: 277-289.
237. M. Kass., A. Witkin., and D. Terzopoulos., Snakes: Active contour models. *Int. J. Comput. Vis*, 1987. 1:321–331.
238. D. Terzopoulos., and R. Szeliski., Tracking with Kalman snakes in Active Vision, A. Blake., and A. Yuille., Eds. *Cambridge, MA: MIT Press*, 1992. 3-20.
239. D. Terzopoulos., and K. Fleischer., Deformable models. *Vis. Comput*, 1988. 4:306-331.
240. T. McInerney., and D. Terzopoulos., A dynamic finite element surface model for segmentation and tracking in multidimensional medical images with application to cardiac 4D image analysis. *Comput. Med. Imag. Graph*, 1995. 19:69–83.
241. L. D. Cohen., On active contour models and balloons. *CVGIP: Image Understand*, Mar 1991. 53:211–218.
242. C. Davatzikos., and J. L. Prince., An active contour model for mapping the cortex. *IEEE Trans. Med. Image*, Mar 1995. 14:65–80.

243. A.J. Abrantes., and J.S. Marques., A class of constrained clustering algorithms for object boundary extraction. *IEEE Trans. Image Processing*, Nov 1996. 5:1507-1521.
244. J. L. Prince., and C. Xu., A new external force model for snakes. *Proceeding Image and Multidimensional Signal Processing Workshop*, 1996. 30–31.
245. C. Davatzikos., and J. L. Prince., Convexity analysis of active contour models. *Proc. Conf. Information Science and Systems*, 1994. 581–587.
246. L. D. Cohen., and I. Cohen., Finite-element methods for active contour models and balloons for 2-D and 3-D images. *IEEE Trans. Pattern Anal. Machine Intell*, Nov 1993. 15:1131-1147.
247. B. Leroy., I. Herlin., and L. D. Cohen., Multi-resolution algorithms for active contour models. *12th Int. Conf. Analysis and Optimization of Systems*, 1996. 58-65.
248. Abdul Kadir Jumaat et al., Segmentation of Masses from Breast Ultrasound Images using Parametric Active Contour Algorithm. *Procedia Social and Behavioral Sciences*, 2010. 8:640-647.
249. Xu, C., & Prince, J.L., Gradient Vector Flow: A New External Force for Snakes. *Proc. IEE Conf. on Computer Vision*, 1997. 66-71.
250. Uppu, S.K., A Study of Parametric Active Contours for the Application of MR Axial Abdominal Images. *Madurai Kamaraj University, Madurai, India*, August 2006.
251. Tabrizi, J.H., Using Active Contours for Segmentation of Middle-Ear Images. *McGill University, Quebec*, Jan 2003.

252. John, O., Rawlings., Sastry G. Pentula., David A., Dickey., Applied regression analysis: a research tool, *Springer-Verlag, 2nd edition*, 1998.
253. Ballabio, D., Vasighi, M., A MATLAB toolbox for Self-Organizing Maps and supervised neural network learning strategies. *Chemom. Intell. Lab. Syst*, 2012. 118:24-32.
254. Demuth, H., Neural Network Toolbox. *Networks* 2002. 24:1-8.
255. Li, C., Duric, N., Huang, L., Breast imaging using transmission ultrasound: reconstructing tissue parameters of sound speed and attenuation. In: Peng, Y., Zhang, Y., (eds) international conference on biomedical engineering and informatics. Piscataway. *IEEE; Sanya, China, May 27-30, 2008*. 708–712.
256. Hunt, J.W., Arditi, M., Foster, F.S., Ultrasound Transducers for Pulse-Echo Medical Imaging. *IEEE Trans Biomed Eng*, 1983. 30:453–481.
257. Parmar, N., Kolios M.C., An Investigation of the use of Transmission Ultrasound to Measure Acoustic Attenuation Changes in Thermal Therapy. *Med Biol Eng Comput*, 2006. 44(7):583–591.
258. C. R. Hill, J. C. Bamber, G. R. ter Haar., Physical Principles of Medical Ultrasonics. *John Wiley & Son, Ltd, 2nd edition*, 1981.
259. Liang, X.G., Ge, X.S., Zhang, Y.P., Wang, G.J., A convenient method of measuring the thermal conductivity of biological tissue. *Phys Med Biol*, 1991. 36(12):1599–605.
260. Bhavaraju, N.C., Valvano, J.W., Thermophysical properties of swine myocardium. *International Journal of Thermophysics*, 1999. 20(2):665–76.
261. Guntur, S.R., Lee, K.I., Paeng, D.G., Coleman, A.J., Choi, M.J., Temperature-dependent thermal properties of ex vivo liver undergoing thermal ablation. *Ultrasound Med Biol*, 2013. 39(10):1771–84.
262. Choi, J., Morrissey, M., Bischof, J.C., Thermal Processing of Biological Tissue at High Temperatures: Impact of Protein Denaturation and Water Loss on the Thermal Properties of Human and Porcine Liver in the Range 25–80 °C. *Journal of Heat Transfer*, 2013. 135(6):061302.