

STABILIZE MACHINING CONDITION OF ELECTRICAL DISCHARGE
MACHINING TO MINIMIZE FRICTION OF HIP IMPLANT

NAZRIAH BINTI MAHMUD

A thesis submitted in fulfilment of the
requirements for the award of the degree of
Doctor of Philosophy (Biomedical Engineering)

Faculty of Biosciences and Medical Engineering
Universiti Teknologi Malaysia

JANUARY 2018

To my beloved father and mother (Mahmud Man and Fatimah Jaafar)
My beloved husband, Mohd Zulhairi Daud
My beloved daughter, Nur Alyaa' Qaisarah Mohd Zulhairi
My beloved siblings (Siti Zahrah, Azizah and Maimunah)
My beloved father and mother in law (Daud and Che Rohani)

ACKNOWLEDGEMENT

Alhamdulillah, praise to Allah whose his blessings and guidance have helped me to be able to finish this thesis.

I wish to express my sincere appreciation to my supervisor, PM Dr Azli Yahya and my Co-supervisor, Prof Dato' Ir. Dr. Mohammed Rafiq Dato' Abdul Kadir for their guidance, advices and motivation during my PhD study.

I am also indebted to Malaysian Ministry of Higher Education for funding my PhD study through MyBrain15(MyPhD) program.

I am also very thankful to Dr Trias Andromeda for helping me in this study. I wish to appreciate my friends, Liyana, Ade, Kartiko and Nabil who helped me with their own way to support, provide critical help and assistance at various occasions.

I am grateful to all my family members who have sacrifice a lot to help me finalize my thesis.

ABSTRACT

Electrical Discharge Machining, EDM is one of the technologies used for surface modification such as the embedded micro-dimples on the metallic acetabular cup. During the machining process, changes in the gap distance may lead to load changes from open to short circuit. Limiting the load current under short circuit conditions and load voltage under open circuit conditions is the requirement in this system. Power supply is one of the elements that controls the process parameters which is related to improve the machining condition as well as Material Removal Rate (MRR). This research proposes a Switch Mode Power Supply method implementing new design of Flyback power supply which can stabilize the voltage during open circuit condition as well as during discharge condition. A model of Flyback power supply was designed and simulated using MATLAB/SIMULINK to investigate the response of the model by measuring the output voltage. The design was then fabricated and an experiment was conducted to validate the simulation. Eight micro-dimples in lower position and twelve micro-dimples in upper position, both in circular arrangement were machined on metallic acetabular cup using Flyback power supply and Linear power supply. Each micro-dimples were then further investigated to find the MRR. The experimental studies then followed by tribological test to screen the friction and wear rate. This experimental work is conducted using four ball tester to describe the friction from the coefficient of friction value measured. Research conducted shows that the Flyback power supply improve the consistency of MRR by 4.2 % for electrode diameter 1000 μm and 21 % for electrode diameter 500 μm . This consistency may help to predict the machining time, thus improving the production of micro-dimples in required time. The quantitative results of the friction test concluded that metallic acetabular cup with 8 micro-dimples has reduce the friction and delayed severe damage, which may then prolong the survival rate of hip implant.

ABSTRAK

Pemesinan Nyahcas Elektrik (EDM) adalah salah satu teknologi yang digunakan dalam pengubahsuaian permukaan seperti penghasilan lubang mikro di atas permukaan melengkung implan pinggul. Semasa proses penghasilan lubang mikro, perubahan jarak di antara elektrod dan bahan kerja membolehkan perubahan beban berlaku dari litar terbuka kepada litar pintas. Mengehendkan arus beban ketika keadaan litar pintas dan voltan beban ketika litar terbuka adalah keperluan dalam sistem ini. Bekalan kuasa adalah salah satu elemen yang mengawal proses tersebut bagi memperbaiki keadaan semasa pemesinan dan meningkatkan Kadar Hakisan Bahan (MRR). Kajian ini mencadangkan satu reka bentuk penjana kuasa baru dikenali sebagai bekalan kuasa *Flyback* yang boleh menstabilkan voltan semasa keadaan litar buka serta semasa keadaan nyahcas. Model *Flyback* direka dan disimulasi menggunakan MATLAB/SIMULINK untuk menyiasat tindak balas model dengan mengukur voltan keluaran. Model fizikal kemudian dilaksanakan dan kajian eksperimen dilakukan untuk mengesahkan simulasi. Lapan dan dua puluh lubang mikro pada permukaan logam melengkung dalam susunan membulat dimesin menggunakan bekalan kuasa *Flyback* dan bekalan kuasa *Linear*. Setiap lubang mikro kemudiannya dianalisa untuk mendapatkan kadar hakisan bahan. Ujikaji diteruskan dengan ujian tribologi untuk menyaring geseran dan kadar kehausan dalam logam ke logam permukaan melengkung ditentukan melalui Nilai Pekali Geseran (CoF) yang diukur. Keputusan menunjukkan bahawa konsistensi MRR adalah lebih tinggi ketika bekalan kuasa *Flyback* digunakan dengan peningkatan 4.2 % untuk elektrod diameter 1000 μm dan 21 % untuk elektrod 500 μm . Konsistensi ini boleh membantu untuk meramalkan masa pemesinan, sekaligus meningkatkan pengeluaran lubang mikro dalam masa yang diperlukan. Analisa kuantitatif daripada ujikaji geseran menyimpulkan lapan lubang mikro dapat melambatkan kerosakan kepada permukaan melengkung, di mana seterusnya boleh memanjangkan hayat implan.

TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	DEDICATION	iii
	ACKNOWLEDGEMENT	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	xi
	LIST OF FIGURES	xiii
	LIST OF ABBREVIATIONS	xvii
	LIST OF SYMBOLS	xviii
	LIST OF APPENDICES	xix
1	INTRODUCTION	1
	1.1 Background of the Study	1
	1.2 Problem Statement	5
	1.3 Objectives	6
	1.4 Scope of Study	6
	1.5 Limitations	7
	1.6 Assumptions	8
	1.7 Significant of study	8
	1.8 Organization of the thesis	9

2	LITERATURE REVIEW	10
	2.1 Introduction	10
	2.2 Electrical Discharge Machining (EDM)	10
	2.2.1 Die-Sinking EDM system	13
	2.2.2 Die-Sinking EDM process	14
	2.2.3 EDM parameters	15
	2.3 Power Generator Configuration in EDM system	17
	2.3.1 RC type generator	18
	2.3.2 Controlled pulse generator	19
	2.3.3 Development on EDM power supplies	20
	2.4 Metal-on-Metal Hip Replacement	25
	2.4.1 Lubrication in Hip Implant	27
	2.4.2 Surface Texturing in Hip Implant	32
	2.5 Summary	34
3	RESEARCH DESIGN AND METHODOLOGY	36
	3.1 Introduction	36
	3.2 Development of Power Supply for Micro-EDM	38
	3.2.1 Implementation and verification of the Flyback converter	40
	3.3 Preparation for machining micro-dimples on hip implant	42
	3.3.1 Experimental Setup	43
	3.3.2 Lab scale Micro-EDM	45
	3.3.3 Material Removal Rate	47
	3.4 Preparation for tribology test	48
	3.4.1 Modified Four Ball tester as Friction Screening for Hip Implant	48
	3.5 Summary	52

4	SIMULATION OF FLYBACK CONVERTER AND CONTROL DESIGN	53
	4.1 Introduction	53
	4.2 Flyback Converter Topology	53
	4.3 Designing of Flyback Converter	55
	4.4 Flyback Dynamic Behavior Model	57
	4.5 Controller design	59
	4.6 Dynamic simulation results	61
	4.7 Summary	71
5	EXPERIMENTAL RESULTS	72
	5.1 Introduction	72
	5.2 Experimental verification and performance comparisons	73
	5.3 Micro-dimples production using EDM	79
	5.4 Material Removal Rate (MRR) Results	83
	5.5 MRR Results Analysis	95
	5.5.1 Material Removal Rate of Alloy Tool Steel workpiece with Tungsten electrode	95
	5.5.2 Material Removal Rate of Mild Steel workpiece with Tungsten electrode	97
	5.5.3 Material Removal Rate of Mild Steel workpiece material with High Steel electrode	100
	5.6 Friction screening	103
	5.6.1 Effect of micro-dimples on the Friction of Coefficient of the micro-dimpled Acetabular Cup	104
	5.7 Summary	107
6	CONCLUSION AND FUTURE WORKS	108

6.1 Summary of Research and Conclusion	108
6.2 Future works	109
REFERENCES	111
Appendices	119-123

LIST OF TABLES

TABLE NO.	TITLE	PAGE
2.1	Major differences between conventional and Micro-EDM [24]	12
2.2	Previous study on Flyback power supply and the research gap	24
2.3	Previous surface texturing works on artificial hip joint	34
3.1	Parameters setup for Lab scale EDM system	44
3.2	Properties of RBD Palm Olein	50
5.1	MRR of alloy tool steel workpiece with tungsten electrode for micro-dimples in lower position	84
5.2	MRR of mild steel workpiece with tungsten electrode for micro-dimples in lower position	85
5.3	MRR of mild steel workpiece with high steel electrode for micro-dimples in lower position	87
5.4	MRR of alloy tool steel workpiece with tungsten electrode for micro-dimples in upper position	89
5.5	MRR of mild steel workpiece with tungsten electrode for micro-dimples in upper position	91
5.6	MRR of mild steel workpiece with high steel electrode for micro-dimples in upper position	93
5.7	MRR of alloy tool steel workpiece material with tungsten electrode using Flyback and Linear Power Supply	95
5.8	MRR of mild steel workpiece material with tungsten electrode using Flyback and Linear Power Supply	98

5.9	MRR of mild steel workpiece material with high steel electrode using Flyback and Linear Power Supply	101
5.10	Coefficient of friction and frictional torque for different speed	105

LIST OF FIGURES

FIGURE NO.	TITLE	PAGE
2.1	Die-Sinking Electrical Discharge Machining system	13
2.2	Die-Sinking EDM machining process	15
2.3	Resistance-capacitance RC pulse generators [39]	18
2.4	Capacitor charging and discharging for RC pulse generator [40]	18
2.5	EDM configurations with pulse control circuit [47]	19
2.6	Voltage and current waveform for transistor type pulse generator [48]	20
2.7	Basic block diagram of the current controlled EDM system [51]	22
2.8	Basic block diagram of LCC resonant converter for EDM power supply [52]	22
2.9	Flyback converter with regenerative clamp [11]	23
2.10	Hip prosthetics: resurfacing hip replacement and total hip replacement [56]	25
2.11	Lubrication regime with their respective lambda ratio[65]	27
2.12	Stribeck curve plot [68]	29
2.13	Schematic diagram of a hip implant with radii of the femoral head (R_1) and outside of the acetabular cup (R_2) [70]	31
2.14	Particles from sliding contact trapped within the dimples [79]	33
3.1	Research design and methodology	37
3.2	Block diagram of power generator connected to electrode and workpiece	38
3.3	Transformer based AC to DC converter [85]	39

3.4	Flyback converter with simple block diagram	39
3.5	Flow chart for Flyback design and development	41
3.6	Laboratory setup	42
3.7	Laboratory setup for machining micro-dimple	43
3.8	Dimple position and distance for each micro-dimple	45
3.9	Lab scale of Micro-EDM system	46
3.10	Novel workpiece positioning system mounted on EDM tank	47
3.11	Four ball tester with its working area and assembly pot for acetabular cup	49
3.12	Experimental procedures	51
4.1	Flyback converter when (a) Switch is ON, (b) Switch is OFF	54
4.2	Schematic diagram of Flyback converter with a simple control block	58
4.3	UC3842 block diagram	60
4.4	Current mode control schematic diagram	62
4.5	SIMULINK model of Flyback converter using UC3842	64
4.6	Simulink model for Flyback converter	65
4.7	Simulink model of UC3842 current mode control	66
4.8	Error Amplifier with (a) resistive feedback, (b) lead compensation feedback	67
4.9	Bode diagram for closed-loop with resistive feedback	68
4.10	Bode diagram for closed-loop with lead compensation feedback	68
4.11	Comparison of bode diagram for uncompensated and compensated system with capacitor of 3.5 pF	69
4.12	Comparison of bode diagram for uncompensated and compensated system	70
4.13	Comparison of output voltage for uncompensated and compensated system	71
5.1	Simulated and experimental output voltage	73
5.2	Experimental setup to measure gap voltage and gap current	74
5.3	Gap voltage and gap current for Linear power supply	75
5.4	Gap voltage and gap current for Flyback power supply	76

5.5	EDM gap voltage and gap current profile using Linear power supply	77
5.6	EDM gap voltage and gap current profile using Flyback power supply	78
5.7	Comparison of discharge error profile for Flyback power supply and Linear power supply	79
5.8	Position of micro-dimple (a) Eight micro-dimples, (b) Twenty micro-dimples	80
5.9	Example of micro-dimples diameter of 500 μm machined using (a) Linear power supply (b) Flyback power supply	80
5.10	Example of micro-dimples machined using (a) Linear power supply (b) Flyback power supply	81
5.11	Close up image of micro-dimples machined using Linear power supply	82
5.12	Close up image of micro-dimples machined using Flyback power supply	83
5.13	Weight of material removed for alloy tool steel workpiece with tungsten electrode in lower position	85
5.14	Weight of material removed for mild steel workpiece with tungsten electrode in lower position	86
5.15	Weight of material removed for mild steel workpiece with high steel electrode in lower position	88
5.16	Weight of material removed for alloy tool steel workpiece with tungsten electrode in upper position	90
5.17	Weight of material removed for mild steel workpiece with tungsten electrode in upper position	92
5.18	Weight of material removed for mild steel workpiece with high steel electrode in upper position	94
5.19	Material Removal Rate for Alloy Tool Steel workpiece material and tungsten electrode	96
5.20	Mean for both power supply with their respective standard deviation	97
5.21	Material Removal Rate for Mild Steel workpiece material and tungsten electrode	99

5.22	Mean for both power supply with their respective standard deviation	100
5.23	Material Removal Rate for Mild Steel workpiece material and High steel electrode	102
5.24	Mean for both power supply with their respective standard deviation	102
5.25	Friction coefficient produced for non-dimpled surface	105
5.26	Friction coefficient produced for micro-dimpled surface using Linear power supply and flyback power supply	106

LIST OF ABBREVIATIONS

BJT	-	Bipolar Junction Transistor
CoC	-	Ceramic-on-Ceramic
CoF	-	Coefficient of friction
CoP	-	Ceramic-on-Polymer
EDM	-	Electrical Discharge Machining
MoM	-	Metal-on-Metal
MoP	-	Metal-on-Plastic
MRR	-	Material Removal Rate
PO	-	Palm Olein
PRC	-	Parallel resonant converter
RC	-	Resistance-capacitance
SMPS	-	Switch Mode Power Supply
SR	-	Surface roughness
SRC	-	Series resonant converter
TWR	-	Tool Wear Rate

LIST OF SYMBOLS

W_b	-	Weight before
W_a	-	Weight after
I_p	-	Current pulse
t_{off}	-	Pulse off-time
t_{on}	-	Pulse on-time
V_p	-	Voltage pulse
α	-	Removal constant of material
L	-	Normal applied load
η	-	Lubricant viscosity

LIST OF APPENDICES

APPENDIX	TITLE	PAGE
A	Flyback power supply circuit	119
B	UC3842 control circuit	120
C	List of Author's Publication	121

CHAPTER 1

INTRODUCTION

1.1 Background of the Study

Hip joint is one of the most important joint in human body that help in our daily movement. It composed of a femoral stem (thigh bone) with a femoral head on top of it. The femoral head articulates against the acetabular cup in the acetabulum [1]. The femoral head and the acetabular cup are coated by a cartilage that lubricates their movement and facilitate the articulation. However, this joint may prone to degenerate due to any causes such as arthritis, damage from injury, or in aged people. As an example, in an arthritic hip joint, the cartilage has been damaged due to inflammation or by degenerative process. If only a part of the joint is damaged, the surgeon may be able to repair the damaged part. However, if the entire joint is damaged, Total Hip Arthroplasty (THA) is suggested in which the damaged part is removed and replaced by an implant [2]. Materials combination between femoral head and acetabular cup is one of the factors affecting the longevity of the hip implant. There are four broad material combinations which are metallic femoral head in a polymeric cup (MoP), ceramic head in a polymeric cup (CoP), ceramic head in a ceramic cup (CoC) and metallic head in a metallic cup (MoM). Severe wear and aseptic loosening of polymer material are the main reasons in failure of MoP and CoP hip implant. Thus, hard-on-hard materials have been introduced which is (CoC) and (MoM). Major problem with CoC implant is squeaking which found to have significant influence on wear mechanism [3]. Although many first

generation of MoM failed to severe wear, the second generation exhibits encouraging wear resistance due to development in material choice, design and manufacturing. Wear rate and friction are known as the major concerns for determining the lifespan of MoM hip implant. From engineering point of view, wear and friction could be minimized through improving lubrication, thus similar principal can be applied to MoM hip implant by surface texturing technique.

Surface texturing technique was previously used on golf ball surfaces to improve their aerodynamic characteristic. Surface texturing is now employed in implant design due to the improved lubrication performance as well as reducing wear and friction. Numerical studies on surface texturing have indicated micro-texture may significantly affect the tribological performance of contact surface and improved lubrication performance [4]. However, there are few surface texture experimental studies in the metallic hip implant, especially in surface dimpling. This approach of surface texturing by applying micro-dimples on the metallic acetabular help to promote in reservation and distribution of lubricant between the contact surfaces thus give better tribological performance [5]. There are many surface texture technologies used for surface modification such as CNC micro drilling, Solid State Laser, Abrasive waterjet and Electrical Discharge Machining (EDM). Due to stamping and drilling process in conventional machining technique, the micro-dimples may suffer from micro cracks. Non-contact process which is the characteristic of EDM would be the best solution to replace drilling and stamping process. While considering the micro-dimples quality, EDM is suggested since it results in no burr formation. Since EDM process is of stochastic nature, it requires a lot of improvements on it especially while considering machining of micro-dimples on metallic acetabular cup. Most of the literatures have discussed about relating of machining parameters to improve performance measures [6], modelling and optimization of the process variables [7,8], and monitoring and control of the sparking process [9]. It has been observed that less attention has been given for enhancing electrical process parameters of EDM process in terms of controlling the process variable such as voltage and current. These two variables highly influence the EDM performance measurement such as Tool Wear Rate (TWR), Material Removal Rate (MRR), and Surface Roughness (SR). Among the three, MRR is the

most dominant performance measure in this study since it affects the quality of micro-dimples in term of consistency of material removal for each micro-dimple.

MRR is defined as the volume of material removed per minutes. Material removal in EDM is realized by thermal action of electrical discharge between the tool and the workpiece, which are connected to a DC power supply. The discharge energy released by this power supply is responsible for melting small quantity of material of both electrode and workpiece. This power supply generates high enough voltage to breakdown the dielectric at a very small gap (10-50 μ m). Before the striking of the spark, the power supply operated at no load condition as the output sees an open circuit. The EDM servo system then adjusts the position of the electrode to breakdown the dielectric. The power supply then sees a negative resistance until the voltage drops to the working gap which is normally ranges from 10-25 V. The current is maintained during this time until the pulse is terminated. The process is repeated at the next cycle. With the working principle stated, enhancing the power supply design that meets the EDM system working principle is necessary.

The conventional power supplies applied in EDM are typically traditional Linear converter, which is act as current source. The major problems of this power supply include large volume and high power losses. In order to solve these problems, resonant-switching power converter is introduced. However, the drawback is when the load resistor of the resonant tank is in high impedance state, the output voltage of the resonant tank is fairly high. The power supply for EDM can be implemented using conventional converter topologies such as buck, buck-boost, single-ended primary-inductance converter (SEPIC) and Flyback converter. The major concern in buck-boost is it inverted polarity in the output terminal while control complexity and components counts are the disadvantages of SEPIC converter. Buck based topology is preferred for implementing the power supply [10], however due to desirable of low output power for EDM application, Flyback converter is suggested. Several studies implementing Flyback converter only discuss on the design of the power supply by evaluating the gap voltage, gap current and

surface images after machining process [11–13]. Therefore, this study provides an intensive knowledge of using Flyback converter in EDM system by analyzing the gap voltage, gap current, surface images and MRR. Experimental results of using Flyback converter comparing with Linear converter are provided by machining micro-dimples on metallic acetabular cup. This novel study is to highlight the design of power supply which can stabilize the voltage during open circuit condition as well as during discharge condition which in turn ensure stable energy levels provided to the gap during machining process. Stable energy level during machining process will lead to high material removal. This power supply will be applied between the electrode and workpiece (acetabular cup) for machining micro-dimples. The machined micro-dimples are expected to improve surface quality as well as consistency of material being removed.

This research seeks to focus on the design and development of power supply utilizing Flyback converter. This approach is compared with Linear power supply by evaluating the consistency of MRR for each micro-dimples machined on metallic acetabular cup of hip implant. Since concern on surface texturing to the hip implant has been rising, this study provided a significance idea on surface texturing by embedded the acetabular cup with micro-dimples. The friction between the femoral head and micro-dimpled acetabular cup is analyzed using four ball tester. The result is compared with non-dimpled acetabular cup. This study will add knowledge to the tribological study of hip implant, as this method expected to minimize friction on hip implant.

1.2 Problem Statement

Enhancing the performance measure of EDM process such as Material Removal Rate (MRR) is widely done on several studies. There are, study on the effect of machining parameters to MRR, modelling and optimization of the process parameters during the discharge process as well as monitoring and control of the spark in order to improve the MRR. Less attention given on the development of the power supply that controls the process parameters itself. Regulation is important for the output voltage to be relatively independent of line or load variations. The regulator circuit compares the output with a reference and adjusts the current flow to make the output as nearly equal to the desired voltage as possible. However, a significant amount of power may be lost in the regulator especially under high line voltage/high load conditions. Therefore the efficiency of Linear power supplies is usually quite low. This lead to less material removal throughout the EDM process. Thus, a Switch Mode Power Supply (SMPS) has been introduced in EDM which high frequency switching devices such as Bipolar Junction Transistors (BJTs) and MOSFETs, are used to directly rectified line voltage and convert it to a pulsed waveform. Since the switching devices are fully on or fully off, there is relatively little power loss. Therefore, the efficiency can be much higher for SMPSs. Due to this advantage, SMPS is proposed that supply the power for erosion to take place. This SMPS consists of feedback which helps to stabilize the power supplied with changing load. Stabilize machining condition will lead to more consistent material removal during the EDM process. This stabilize machining condition is required to machine micro-dimples on acetabular cup of hip implant in order to promote lubrication activities that lead to minimize friction of hip implant.

1.3 Objectives

Based on the issue regarding EDM and MoM hip implant, there is a demand for researcher to develop power supply which will be used to machine micro-dimples on a model of metallic acetabular cup. Thus, this study embarks on the following objectives:

1. To design and develop Switch Mode Power Supply (SMPS) using Flyback converter tailored for machining micro-dimples on hip implant
2. To analyze the consistency of Material Removal Rate (MRR) by comparing two types of power supply namely Linear power supply and Flyback power supply.
3. To analyze friction between femoral head and acetabular cup by conducting tribology test using modified four ball machine.

1.4 Scope of Study

The following scopes of work are covered as follows:

1. A broad but critical literature review of Electrical Discharge Machining. The review covered on power supply, machining parameter and machining performance.
2. A simulation using MATLAB is conducted to predict the response of Switch Mode Power Supply before implement it into hardware.
3. The implementation of the SMPS is applied to the laboratory scaled EDM machine.
4. Machining of micro-dimples in circular arrangement with total of eight micro-dimples and twenty micro-dimples are conducted.

5. An experiment of machining process is conducted in order to measure the consistency of material removal rate for Linear power supply as well as SMPS.
6. The result of machining micro-dimples on metallic acetabular cup hip implant is then tested using modified four ball tester to observe the friction when micro-dimples are applied using different power supply.

1.5 Limitations

The limitations of this study are as follows:

1. Experiment for tribology test was limited to Palm Olein as the lubricant with load of 40N.
2. The diameter of the dimples was limited according to the electrode size of 500 μm and 1000 μm .
3. Experiment for machining micro-dimples on model of hip implant was limited to metal-on-metal material bearing combinations.

1.6 Assumptions

The following assumptions are made in this study:

1. Temperature and pressure of dielectric fluid were constant during the experiment for machining micro-dimples
2. Current consumption was constant throughout the experiment for machining micro-dimples.
3. The temperature was assumed to be 37 ± 3 °C throughout the tribology test.

1.7 Significant of study

MRR is one of the performance measures which are widely emphasized by other researchers in order to improve the capability of EDM process. The study of consistency of MRR is important as it directly affect the machining time as well as the surface quality of each micro-dimples machined on acetabular cup of hip implant. More consistent value of MRR will lead to nearly equal quality of micro-dimples produced such as depth and diameter of the micro-dimples. This help in reserving the same amount of lubricant for each micro-dimple that lead to wear and friction reduction, thus prolonging the lifespan of the hip implant. The value of MRR is essential to the selection of power supply used for machining. Therefore the study of designing power supply using switch mode technology is necessary for replacing the existing Linear converter power supply. This adds value to the surface texturing industry by using EDM with switch mode technology. This research also aims to observe the effectiveness of micro-dimples embedded in the metallic acetabular cup used for lubrication purpose. For biomedical field, this study add knowledge to the design of hip implant with textured with micro-dimples.

1.8 Organization of the thesis

Chapter 1 introduces this study including problem statement, objectives, scope, limitations, assumptions made during this study and significant of the study. Chapter 2 describes a comprehensive literature on EDM system including the development of power supplies. This chapter also include literature on Metal-on-Metal hip replacement detailed on lubrication and surface texturing on hip implant. Chapter 3 explains the research methodology including the mathematical equation used for simulation of power supply. This chapter also describes the experimental setup for machining micro-dimples on hip implant and tribology experiments to analyze friction on hip implant. Chapter 4 presents the simulation of Flyback converter and control design. Chapter 5 discusses the results obtained from machining micro-dimples on metallic acetabular cup in term of gap voltage and gap current profile, surface images and MRR. This chapter also includes the experimental results obtained from friction test, comparison between non-dimpled and micro-dimpled hip implant. Chapter 6 concludes the study and gives suggestion for further work.

REFERENCES

1. S.J.Hall, Lower extremity, In: James M. Smith, *Basic Biomech.*, 1995.
2. Rabiei, A., *Hip Prosthesis*, 2009.
3. Currier, J.H., Anderson, D.E., Citters, D.W. Van, A proposed mechanism for squeaking of ceramic-on-ceramic hips. *Wear*, 2010. 269: 782–789.
4. Gao, L., Yang, P., Dymond, I., Fisher, J., Jin, Z., Effect of surface texturing on the elastohydrodynamic lubrication analysis of metal-on-metal hip implants. *Tribol. Int.*, 2010. 43: 1851–1860.
5. Choudhury, D., Walker, R., Roy, T., Paul, S., Mootanah, R., Performance of honed surface profiles to artificial hip joints: An experimental investigation. *Int. J. Precis. Eng. Manuf*, 2013. 14: 1847–1853.
6. Singh, A., Singh, R.P., To Determine the Effect of Machining Parameters on Surface Roughness of EN 31 Stainless steel using EDM, 2016. 4.
7. Amorim, F.L., Weingaertner, W.L., Bassani, I.A., Aspects on the optimization of die-sinking EDM of tungsten carbide-cobalt. *J. Brazilian Soc. Mech. Sci. Eng*, 2010. 32: 496–502.
8. Sameh, S.H., Parameter optimization of electrical discharge machining process by using Taguchi approach. *J Eng. Technol. Res*, 2014. 6: 27–42.
9. Muthuramalingam, T., Mohan, B., Rajadurai, A., Saravanakumar, D., Monitoring and fuzzy control approach for efficient electrical discharge machining process. *Mater. Manuf, Process*, 2014. 29: 281–286.
10. Tastekin, D., Krötz, H., Gerlach, C., Roth-Stielow, J., A Novel Electrical Power Supply for Electrothermal and Electrochemical Removal Machining Methods, 2009: 2682–2688.
11. Odulio, C.M.F., Sison, L.G., Escoto, M.T., Regenerative clamp as reset

- winding in flyback converters for EDM applications. *2004 IEEE Int. Conf. Ind. Technol. 2004. IEEE ICIT '04*, 2004.1: 0–3.
12. Odulio, C.M.F., Sison, L.G., Ph, D., Escoto, M.T., Energy-saving Flyback Converter for EDM Applications, 2005.
 13. Mysi, W., Power supply unit for an electric discharge machine. *13th Eur. Conf. Power Electron. Appl.* 2009.
 14. Bojorquez, B., Marloth, R.T., Es-Said, O.S., Formation of a crater in the workpiece on an electrical discharge machine. *Eng. Fail. Anal.*, 2002. 9: 93–97.
 15. Luis, C.J., Puertas, I., Villa, G., Material removal rate and electrode wear study on the EDM of silicon carbide. *Mater. Process. Technol.*, 2005. 164–165: 889–896.
 16. Kiyak, M., O.C,akır, Examination of machining parameters on surface roughness in EDM of tool steel. *Mater. Process. Technol.*, 2007. 191: 141–144.
 17. Ho, K.H., Newman, S.T., State of the art electrical discharge machining (EDM). *Int. J. Mach. Tools Manuf.* 2003, 43, 1287–1300.
 18. Wansheng, Z., Zhenlong, W., Shichun, D., Guanxin, C., Hongyu, W., Ultrasonic and electric discharge machining to deep and small hole on titanium alloy. *Mater. Process. Techno*, 2002. 120: 101–106.
 19. Janeček, M., Nový, F., Stráský, J., Harcuba, P., Wagner, L., Fatigue endurance of Ti-6Al-4V alloy with electro-eroded surface for improved bone in-growth. *J. Mech. Behav. Biomed. Mater*, 2011. 4: 417–422.
 20. Stráský, J., Janeček, M., Harcuba, P., Bukovina, M., Wagner, L., The effect of microstructure on fatigue performance of Ti-6Al-4V alloy after EDM surface treatment for application in orthopaedics. *J. Mech. Behav. Biomed. Mater*, 2011. 4: 1955–1962.
 21. Tamaki, Y., Kataoka, Y., Miyazaki, T., Bone regenerative potential of mesenchymal stem cells on a micro-structured titanium processed by wire-type electric discharge machining. 2010, 6.
 22. Moses, M.D., Jahan, M.P., Micro-EDM machinability of difficult-to-cut Ti-

- 6Al-4V against soft brass. *Int. J. Adv. Manuf. Technol*, 2015. 81: 1345–1361.
23. Masuzawa, T., State of the Art of Micromachining. *CIRP Ann. - Manuf. Technol*, 2000. 49: 473–488.
 24. Raju, L., Hiremath, S.S., A State-of-the-art Review on Micro Electro-discharge Machining. *Procedia Technol*, 2016. 25: 1281–1288.
 25. Gurule, N.B., Pansare, S.A., Potentials of Micro-EDM n.d., 50–55.
 26. Pandey, A., Singh, S., Current research trends in variants of Electrical Discharge Machining : A review. *Int. J. Eng. Sci. Technol*, 2010. 2: 2172–2191.
 27. Jahan, M.P., Rahman, M., Wong, Y.S., A review on the conventional and micro-electrodischarge machining of tungsten carbide. *Int. J. Mach. Tools Manuf.* 2011. 51: 837–858.
 28. Ojha, K., Garg, R.K., Singh, K.K., MRR Improvement in Sinking Electrical Discharge Machining : A Review, 2010. 9: 709–739.
 29. Gostimirovic, M., Kovac, P., Sekulic, M., Skoric, B., Influence of discharge energy on machining characteristics in EDM. *J. Mech. Sci. Technol*, 2012. 26: 173–179.
 30. Son, S., Lim, H., Kumar, A.S., Rahman, M., Influences of pulsed power condition on the machining properties in micro EDM. *J. Mater. Process. Technol*, 2007. 190: 73–76.
 31. Kansal, H.K., Singh, S., Kumar, P., Parametric optimization of powder mixed electrical discharge machining by response surface methodology, 2005. 169: 427–436.
 32. Hascalik, A., Caydas, U., Electrical discharge machining of titanium alloy (Ti – 6Al – 4V) 2007, 253, 9007–9016.
 33. Kumar, S., Singh, R., Singh, T.P., Sethi, B.L., Surface modification by electrical discharge machining : A review. *J. Mater. Process. Technol*, 2009. 209: 3675–3687.
 34. Lee, S.H., Li, X.P., Study of the effect of machining parameters on the machining characteristics in electrical discharge machining of tungsten carbide. *J. Mater. Process. Technol*, 2001. 115: 344–358.

35. Pradhan, M.K., Biswas, C.K., Modelling of machining parameters for MRR in EDM using response surface methodology. *Proc. NCMSTA'08 Conf.*, 2008. 535–542.
36. Lin, Y., Chen, Y., Wang, D., Lee, H., Optimization of machining parameters in magnetic force assisted EDM based on Taguchi method. *J Mater. Process. Technol.*, 2009. 9: 3374–3383.
37. Lin, Y.C., Yan, B.H., Chang, Y.S., Machining characteristics of titanium alloy (Ti ± 6Al ± 4V) using a combination process of EDM with USM. *J. Mater. Process. Technol.*, 2000. 104: 171–177.
38. Haron, C.H.C., Deros, B., Ginting, A., Fauziah, M., Investigation on the influence of machining parameters when machining tool steel using EDM. *J Mater. Process. Technol.*, 2001. 116: 84–87.
39. Huang, H., Bai, J., Lu, Z., Guo, Y., A Novel Half-Bridge Power Supply for High Speed Drilling Electrical Discharge Machining. *J. Electromagn. Anal. Appl.*, 2009. 2: 108–113.
40. Sen, B., Kiyawat, N., Singh, P., Mitral, S., et al., Developments in Electric Power Supply Configurations for Electrical Discharge-Machining (EDM). *Power Electron. Drive Syst.*, 2003. 659–664.
41. Wong, Y.S., Rahman, M., Lim, H.S., Han, H., Ravi, N., Investigation of micro-EDM material removal characteristics using single RC -pulse discharges, 2003. 140: 303–307.
42. Jahan, M.P., Wong, Y.S., Rahman, M., A study on the quality micro-hole machining of tungsten carbide by micro-EDM process using transistor and RC-type pulse generator. *J. Mater. Process. Technol.*, 2009. 209: 1706–1716.
43. Kunieda, M., Challenges to miniaturization in micro edm
44. Singh, S., Bhardwaj, A., Review to EDM by Using Water and Powder-Mixed Dielectric Fluid, 2011. 10: 199–230.
45. Han, F., Wachi, S., Kunieda, M., Improvement of machining characteristics of micro-EDM using transistor type isopulse generator and servo feed control, 2004. 28; 378–385.
46. Han, F., Chen, L., Yu, D., Zhou, X., Basic study on pulse generator for micro-

- EDM. *Int. J. Adv. Manuf Technol.*, 2007. 33: 474–479.
47. Shah, A., Prajapati, V., Patel, P., Pandey, A., et al., Development of pulsed power dc supply for micro edm, *UGC National Conference on Advances in Computer Integrated Manufacturing*, 2007.
 48. Yahya, A., Manning, C.D., Determination of material removal rate of an electro-discharge machine using dimensional analysis. *J. Phys. D. Appl. Phys.*, 2004. 37: 1467–1471.
 49. Peng, Z.L., Li, Y.N., Wang, Z.L., Study on the Characteristics of Discharge Waveform in Micro EDM Deposition Process. *Mater. Sci. Forum*, 2011. 697–698: 187–191.
 50. Lin, R., Hsu, C., Changchien, S., Interleaved Four-Phase Buck-Based Current Source With Center-Tapped Energy-Recovery Scheme for Electrical Discharge Machining, *IEEE Transactions on power electronics*, 2011. 26: 110–118.
 51. Casanueva, R., Ochoa, M., Azcondo, F.J., Current Mode Controlled LCC Resonant Converter for Electrical 2000, 505–510.
 52. Casanueva, R., Chiquito, L.A., Azcondo, F.J., Bracho, S., Current source LCC Resonant Converter for an EDM Power Supply. *The 27th Annual Conference of the IEEE Industrial Electronics Society*, 2001. 1027-1032.
 53. Bhat, A.K.S, Analysis and Design of Series-Parallel Resonant Power Supply *IEEE Transactions of Aerospace and Electronic System*, 1992. 28: 249–258.
 54. Batarseh, I., Resonant Converter Topologies with Three and Four Energy Storage Elements, *IEEE Transactions on Power Electronics*, 1994. 9(1):64-73
 55. Mattei, L., Puccio, F. Di, Piccigallo, B., Ciulli, E., Lubrication and wear modelling of artificial hip joints : A review. *Tribology Int*, 2011. 44: 532–549.
 56. Di Puccio, F., Mattei, L., Biotribology of artificial hip joints. *World J. Orthop.* 2015. 6: 77–94.
 57. Roberts, P., Grigoris, P., Bosch, H., Talwaker, N., et al., (iii) Resurfacing arthroplasty of the hip. *Curr. Orthop.*, 2005. 263–279.
 58. Hernandez-Rodriguez, M.A.L., Mercado-Solis, R.D., Perez-Unzueta, A.J., Martinez-Delgado, D., Cantu-Sifuentes, M., Wear of cast metal – metal pairs

- for total replacement hip prostheses. *Wear* , 2005. 259: 958–963.
59. Katti, K.S., Biomaterials in total joint replacement. *Colloids and Surfaces*, 2004. 39: 133–142.
 60. Dumbleton, J.H., Manley, M.T., Metal-on-Metal Total Hip Replacement What Does the Literature Say ? *J. Arthroplasty* ,2005. 20: 174–188.
 61. Haynes, J.A., Stambough, J.B., Barrack, R.L., Nam, D., Conversion of a failed hip resurfacing arthroplasty to total hip arthroplasty : pearls and pitfalls. *Curr Rev Musculoskelet Med* , 2016. 9: 103–111.
 62. Cuckler, J.M., Moore, K.D., Lombardi, A. V, Mcpherson, E., Emerson, R., Large Versus Small Femoral Heads in Metal-on-Metal Total Hip Arthroplasty. *J. Arthroplasty*, 2004. 19: 41–44.
 63. Ghosh, S., Abanteriba, S., Status of surface modification techniques for artificial hip implants. *Sci. Technol. Adv. Mater.*, 2016. 17: 1–21.
 64. Zhou, Z.R., Jin, Z.M., Biotribology: Recent Progresses and Future Perspectives. *Biosurface and Biotribology*, 2015. 1: 3–24.
 65. Rieger, J.S., Heitzmann, D.W.W., Kretzer, J.P., Sonntag, R., Hard-on-Hard Lubrication in the Artificial Hip under Dynamic Loading Conditions, 2013. 8(8): 1–8.
 66. Sagbas, B., Biotribology of Artificial hip Joints. In: *Adv. Tribol.*, 2016.
 67. Hamrock, B., Dowson, D., Isothermal Elastohydrodynamic Lubrication of Point Contacts: Part 1—Theoretical Formulation. *J Lubr. Technol.* 1976. 98: 223–228.
 68. Maru, M.M., Consideration of Stribeck Diagram Parameters in the Investigation on Wear and Friction Behavior in Lubricated Sliding 2007, XXIX, 55–62.
 69. Stachowiak, G.W., Batchelor, A., Engineering Tribology, 2006.
 70. Jin, Z.M., Stone, M., Ingham, E., Fisher, J., (v) Biotribology. *Curr. Orthop.* 2006. 20: 32–40.
 71. Affatato, S., Spinelli, M., Zavalloni, M., Mazzega-Fabbro, C., Viceconti, M., Tribology and total hip joint replacement: Current concepts in mechanical simulation. *Med. Eng. Phys.*, 2008. 30: 1305–1317.

72. Hsu, S.M., Jing, Y., Hua, D., Zhang, H., Friction reduction using discrete surface textures : principle and design. *J. Phys. D. Appl. Phys.*, 2014. 47: 1–12.
73. Hart, A.J., Sabah, S., Henckel, J., Lewis, A., et al., The painful metal-on-metal hip resurfacing. *J. Bone Jt. Surg.* 2009. 91: 738–744.
74. Nine, M.J., Choudhury, D., Hee, A.C., Mootanah, R., Osman, N.A.A., Wear debris characterization and corresponding biological response: Artificial hip and knee joints. *Materials (Basel)*, 2014. 7: 980–1016.
75. Hasegawa, M., Yoshida, K., Wakabayashi, H., Sudo, A., Cobalt and Chromium Ion Release After Total Hip Arthroplasty. *J. Arthroplasty*, 2012. 27: 990–996.
76. Geiger, M., Roth, S., Becker, W., Influence of laser-produced microstructures on the tribological behaviour of ceramics. *Surf. Coatings Technol.*, 1998. 100–101: 17–22.
77. Ito, H., Kaneda, K., Yuhta, T., Nishimura, I., et al., Reduction of Polyethylene wear by concave dimples on the frictional surface in artificial hip joints. *J Arthroplasty*, 2000. 15:332–338.
78. Roy, T., Choudhury, D., Ghosh, S., Bin, A., Pinguan-murphy, B., Improved friction and wear performance of micro dimpled ceramic-on-ceramic interface for hip joint arthroplasty, *Ceramic International*, 2015. 41: 681–690.
79. Sawano, H., Ishihara, S., Study on long life of artificial joints by investigating optimal sliding surface geometry for improvement in wear resistance, 2009. 33: 492–498.
80. Zhou, X., Galvin, A.L., Jin, Z., Yan, X., The influence of concave dimples on the metallic counterface on the wear of ultra-high molecular weight polyethylene. *Proc IMechE Part JJ Eng. Tribol.*, 2016. 0(0): 1–8.
81. Chyr, A., Qiu, M., Speltz, J.W., Jacobsen, R.L., et al., A patterned microtexture to reduce friction and increase longevity of prosthetic hip joints. *Wear*, 2014. 315: 51–57.
82. Choudhury, D., Urban, F., Vrbka, M., Hartl, M., Krupka, I., A novel tribological study on DLC-coated micro-dimpled orthopedics implant

- interface. *J. Mech. Behav. Biomed. Mater.*, 2015. 45: 121–131.
83. Razak, D.M., Syahrullail, S., Sapawe, N., et al., A new tribological approach on metal cup with optimized pits model using spark discharge machine. *Part. Sci. Technol.*, 2016. 00(00): 1–8.
 84. Moylan, S.P., Chandrasekar, S., Benavides, G.L., High-Speed Micro-Electro-Discharge Machining, 2005.
 85. Fleming, B., Build a Pulse EDM Machine, 2011.
 86. Kartiko, N., Azli, Y., Nor Liyana, S.H., Syahrullail, S., Razak, M.D., Development of Computer-Aided EDM for Machining Micropits on Spherical Surface of Hip Implant. *Appl. Mech. Mater.*, 2014. 554: 541–545.
 87. Syahrullail, S., Sapawe, N., Razak, M.D., Azli, Y., Effect of Surface Modification of Acetabular Cup with Embedded Micro-Pits on Friction Properties. *American J. of Mech. Eng.*, 2014. 2(5): 125–129.
 88. Razak, D., Syahrullail, S., Yahya, A., Mahmud, N., et al., Lubrication on the Curve Surface Structure Using Palm oil and Mineral oil. *Procedia Eng.*, 2013. 68: 607–612.
 89. Kumar Pandey, S., Patil, S., S.Rajguru, V., Isolated Flyback converter designing, modeling and suitable control strategies. *Proc. Int. Conf. Adv. Power Electron. Instrum. Eng.*, 2014. 47–58.
 90. Mclyman, C.W.M.T., *Transformer and inductor design handbook*. (3rd ed.), 2004.
 91. Mahmud, N., Yahya, A., Andromeda, T., Hashim, N.L.S., Design and Implementation of Flyback Converter for Electrical Discharge Machining Power Generator. *J. Teknol.* 2014. 67: 65–70.
 92. Liao Wei-Hsin, Wang Shun-Chung, L.Y.-H., Generalized Simulation Model for a Switched-Mode Power Supply Design Course Using MATLAB/SIMULINK. *IEEE Trans. Educ.* 2012. 55: 36–47.