

2 MHz ELECTRICAL RESISTANCE TOMOGRAPHY FOR STATIC LIQUID-
SOLID PROFILE MEASUREMENT

YASMIN BINTI ABDUL WAHAB

UNIVERSITI TEKNOLOGI MALAYSIA

2 MHz ELECTRICAL RESISTANCE TOMOGRAPHY FOR STATIC LIQUID-
SOLID PROFILE MEASUREMENT

YASMIN BINTI ABDUL WAHAB

A thesis submitted in fulfilment of the
requirements for the award of the degree of
Doctor of Philosophy (Electrical Engineering)

Faculty of Electrical Engineering
Universiti Teknologi Malaysia

MAY 2017

In the name of Allah, the most Gracious and the most Merciful.

To my beloved and supportive parents,

husband, brothers, sisters

and

my lovely children

ACKNOWLEDGEMENT

I would like to dedicate my deepest gratitude to my supervisor Prof. Dr. Ruzairi Abdul Rahim for his outstanding support and excellent supervisions. This research would not have been successful without his valuable guidance, enthusiastic help as well as constructive criticisms throughout the research. I would also like to give my sincere thanks to Dr. Mohd Hafiz Fazalul Rahiman, Dr. Leow Pei Ling, and Assoc. Prof. Dr. Herlina Abdul Rahim as my co-supervisors for their valuable suggestions and constructive criticisms.

My whole appreciation to PROTOM-*i* research group members at Faculty of Electrical Engineering, Universiti Teknologi Malaysia especially Suzanna Ridzuan Aw, Fazlul Rahman Mohd Yunus, Bro Jaysuman, Juliza Jamaludin, Helen Goh, Nor Muzakkir, Naizatul shima, Mohd Fadzli, Saiful Badri, my friend Dr. Nurul Adila and process control laboratory technician Mr. Md Fadzli Bin Sahril for your helps and supports during my research. Also, thanks to my friends and all those whom had helped me in one-way or other during my research.

Special thanks to my parents for sharing their wisdoms and continued guidance during my study. To my lovely husband, Ahmad Syamrim, you are my better half, thank you for your constant encouragement and infinitive support from the beginning of my research. For my beloved daughters Iffah Humaira and Izzah Humaira, thank you for always cheering up for me.

I would like to thank Universiti Malaysia Pahang and Ministry of Higher Education for granting my scholarship. Last but not least, to Universiti Teknologi Malaysia for allowing me to use the facilities during my research is greatly appreciated and without it, this research could not have been carried out.

ABSTRACT

Tomography is a technique used to reconstruct cross-sectional image of a pipeline for flow monitoring applications. There are several types of tomography system such as X-ray tomography, ultrasonic tomography, and electrical resistance tomography (ERT). ERT has many advantages compared to other types of tomography such as low cost, robust and no radiation. Thus, it becomes particularly suitable for industrial applications. However, it has been observed that the conventional practice of ERT through invasive sensing technique has exposed the ERT metal sensor to corrosion and limited its application because of inaccurate measurement of the data. Consequently, non-invasive ERT has also been introduced in low frequency (in kHz) applied to the ERT system. The low frequency ERT makes use of the phase-sensitive demodulation (PSD) approach and is a complicated technique to implement. Hence, the goal of this research is to design and develop a non-invasive ERT system with a high frequency (2 MHz) source. A total impedance of coupling capacitances (between metal electrode and conductive medium) series with resistance (conductive medium) for each pair of electrodes was assumed in the research. Based on the mathematical equation of the total impedance, the real part is the resistance (conductive medium) must be larger than the imaginary part (capacitances), so that it can easily detect the concentration profile of the conductive medium. Therefore, the minimum frequency needed to ensure that the real part is bigger than the imaginary one is 2 MHz. Simultaneously, the independent and flexible sixteen ERT electrodes designed for the system make it easier to replace and troubleshoot any problems with the sensor. In addition, the experiment was carried out on a two-phase static liquid–solid regime for a linear back-projection algorithm using online configuration, with MATLAB as a software platform. It was also able to detect and visualize the non-homogenous system of the two-phase regime. Later, the reconstructed image was improved using a global threshold technique through offline configuration. The experiment results indicate that it could detect obstacles in a vertical pipe with minimum 12 mm in diameter and 4.5 cm in height.

ABSTRAK

Tomografi merupakan satu teknik yang digunakan untuk menggambarkan keratan rentas bagi saluran paip dalam aplikasi-aplikasi pemantauan aliran. Terdapat beberapa jenis sistem tomografi seperti tomografi X-ray, tomografi ultrasonik, dan tomografi rintangan elektrik (ERT). ERT mempunyai banyak kelebihan jika dibandingkan dengan jenis-jenis tomografi yang lain seperti kos yang rendah, kukuh dan tiada radiasi. Maka, ia sangat sesuai untuk aplikasi industri. Tetapi, konvensional ERT melalui teknik penderia invasif menyebabkan penderia logam ERT terdedah kepada kesan kakisan dan ia menghadkan penggunaannya kerana pengukuran data yang tidak tepat. Maka, teknik penderia bukan invasif juga telah diperkenalkan dengan menggunakan frekuensi yang rendah (dalam kHz). Frekuensi rendah memerlukan kaedah penyahmodulatan peralihan fasa (PSD) dan ianya merupakan teknik yang rumit untuk dilaksanakan. Oleh itu, matlamat kajian ini adalah untuk mereka bentuk dan membangunkan sistem ERT tidak invasif dengan menggunakan sumber frekuensi yang tinggi (2 MHz). Anggaran jumlah galangan bagi setiap pasangan elektrod dengan mengambil kira gandingan kemuatan (antara elektrod logam dan bahan konduktif) sesiri dengan rintangan (bahan konduktif) digunakan dalam kajian ini. Berdasarkan persamaan matematik bagi jumlah galangan tersebut, bahagian nyata mesti lebih besar daripada bahagian khayalan supaya lebih mudah untuk mengesan profil kepekatan bahan konduktif. Maka, frekuensi minimum yang diperlukan untuk membolehkan bahagian nyata lebih besar daripada bahagian khayalan ialah 2 MHz. Pada masa yang sama, enam belas elektrod ERT yang telah direka secara individu dan fleksibel membolehkan penderia lebih mudah diperiksa dan ditukar. Sebagai tambahan, eksperimen telah dijalankan terhadap dua fasa cecair-pepejal rejim yang statik untuk algoritma unjuran kembali linear menggunakan konfigurasi terus, dengan MATLAB sebagai platform perisian. Ia juga telah dapat mengesan dan memberi gambaran bagi sistem dua fasa yang bukan homogen. Kemudiannya, kaedah ambang global melalui konfigurasi tidak terus untuk penambahbaikan gambaran tersebut telah digunakan. Keputusan eksperimen juga telah menunjukkan sistem ini boleh mengesan objek dalam paip menegak dengan ukuran diameter minimum ialah 12 mm dan tinggi sekurang-kurangnya 4.5 cm.

TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	DEDICATION	iii
	ACKNOWLEDGEMENT	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	xii
	LIST OF FIGURES	xiv
	LIST OF ABBREVIATIONS	xix
	LIST OF SYMBOLS	xxi
	LIST OF APPENDICES	xxiii
1	INTRODUCTION	1
	1.1 Introduction	1
	1.2 Sensing Technique of Process Tomography	3
	1.3 Research Background	4
	1.4 Problem Statements	5
	1.5 Aim and Research Objectives	6
	1.6 Research Scopes	6
	1.7 Motivation and Contribution	7
	1.8 Structure of Thesis	8
2	LITERATURE REVIEW	9
	2.1 Introduction	9
	2.2 Current Research on Types of Non-Invasive Industrial Tomography	9

2.2.1	X-ray Tomography	9
2.2.2	Ultrasonic Tomography	11
2.2.3	Optical Tomography	13
2.2.4	Electrical Capacitance Tomography	14
2.3	Recent Works Related to Electrical Resistance Tomography	17
2.4	Basic Principles of Non-Invasive ERT System	21
2.4.1	Resistance and Conductivity	22
2.4.2	Quasi-Static Electric Field	23
2.4.3	Measurement Strategy	26
2.5	Image Reconstruction in Process Tomography	30
2.5.1	Forward Problem	30
2.5.2	Inverse Problem	31
2.6	Summary	36
3	MODELLING AND SIMULATION	39
3.1	Introduction	39
3.2	General Set-up of the Model in COMSOL Multiphysics Software	39
3.3	Determination of Compatible Frequency	42
3.3.1	Optimizing a Suitable Frequency for Non-Invasive ERT	46
3.3.2	Limitation of Main Medium Applied with the 2 MHz Frequency	51
3.4	Modelling for Electrode Dimension	52
3.4.1	Optimizing the Electrode Dimension of Non-Invasive ERT Electrode	53
3.5	Summary	64
4	IMAGE RECONSTRUCTION	66
4.1	Introduction	66

4.2	Forward Problem Solving	66
4.2.1	Generating Map from COMSOL Multiphysics Software	67
4.2.2	Masking Data for Better Sensitivity Map	68
4.3	Inverse Problem Solving	76
4.3.1	Linear Back-Projection Algorithm (LBP)	76
4.4	Image Quality Assessment	77
4.4.1	Multi Scale Structural Similarity (MSSIM)	77
4.4.2	Area Error, AE	78
4.4.3	Solid Concentration	79
4.5	Thresholding technique	79
4.6	Summary	80

5 **HARDWARE AND SOFTWARE**

	DEVELOPMENT	82
5.1	Introduction	82
5.2	Non-Invasive ERT System-An Overview	82
5.2.1	Sensor Design	84
5.2.2	Sensor Switching	87
5.2.3	Selection of the Types of Source Signal	87
5.2.4	Signal Generator Circuit	92
5.2.4.1	DDS Circuit	92
5.2.4.2	Demultiplexer	95
5.2.4.3	Amplifier Circuit	96
5.2.5	Signal Conditioning Circuit	98
5.2.5.1	Current-to-Voltage Amplifier Circuit	98
5.2.5.2	Peak Detector Circuit	100
5.2.6	Microcontroller unit	

	(dsPIC30F6010A)	102
	5.2.6.1 Analogue-to-Digital Conversion (ADC)	104
	5.2.7 Printed Circuit Board (PCB)	106
5.3	Software Development	107
5.4	ANOVA for sensor validation	110
5.5	Summary	111
6	RESULTS AND DISCUSSION	112
6.1	Introduction	112
6.2	Sensor Reading Analysis and Validation	113
	6.2.1 Homogeneity of Variance Test	114
	6.2.2 Analysis using ANOVA	118
6.3	Reconstruction Image Analysis and Validation	123
	6.3.1 Weakness of LBP Algorithm and Threshold Pre-Set Value Approach	126
	6.3.2 Limitation of the Image Reconstructed	128
	6.3.2.1 Blind Spot Experiment	128
	6.3.2.1.1 Analysis and Discussion for Blind Spot Experiment	131
	6.3.2.2 Height Limitation of Phantoms	134
6.3.3	Single Phantom	136
	6.3.3.1 Experimental versus Simulation Results for Single Phantom	136
	6.3.3.2 Analysis and Discussion for Single Phantom	138
6.3.4	Multiple Phantoms	142

6.3.4.1	Experimental versus Simulation Results for Double Phantoms	142
6.3.4.2	Analysis and Discussion for Double Phantoms	143
6.4	Summary	146
7	CONCLUSIONS AND SUGGESTIONS FOR FUTURE WORK	148
7.1	Conclusions	148
7.2	Contribution of the Research	149
7.3	Recommendations for Future Work	150
	REFERENCES	151
	Appendices A–D	170-178

LIST OF TABLES

TABLE NO.	TITLES	PAGE
2.1	Summary of conventional ERT for two-phase mixtures	17
2.2	Current measurement strategy applied in conventional ERT system [100], [104]-[107]	28
2.3	Current measurement strategy applied in conventional ERT system [100], [104]-[107] (continued)	29
2.4	Comparison sensitivity distribution between hard-field and soft-field tomography	34
2.5	Summary of medium tested for research on conventional ERT	37
2.6	Summary research on non-invasive industrial tomography	38
3.1	Parameters material defined in COMSOL Multiphysics	41
3.2	Range of conductivity of tap water to fix with real part bigger than imaginary part	52
3.3	Properties and specific dimension	53
3.4	Example of surface current distribution at height 130 mm with increment of width	56
3.5	Comparison between homogenous and non-homogenous system	61
3.6	Comparison between homogenous and non-homogenous system (continued)	62
3.7	Surface current distribution for the increment of	

	obstacle at the centre of the pipe	64
4.1	Parameters for the system	67
5.1	Initialization sequence setting	93
6.1	P-Value of homogenous variance test for each set group of each transmitte	115
6.2	Parameter for material of phantom	123
6.3	Pth based on AE value for all simulations and experiments (single phantom)	127
6.4	Pth based on AE value for all simulations and experiments (double phantoms)	128
6.5	Pth based on AE value for all simulations and experiments (blind spot)	128
6.6	Tomograms for blind spot experiment at different positions (22 mm)	129
6.7	Tomograms for blind spot experiment at different positions (22 mm) (continued)	130
6.8	Tomograms for blind spot experiment at different positions (12 mm)	131
6.9	MSSIM indexed measured on tomogram for blind spot	132
6.10	Tomograms of single phantom for diameter 12 mm	137
6.11	Tomograms of single phantom for diameter 22 mm	138
6.12	Tomograms of double phantoms for diameter 12 mm	142
6.13	Tomograms of double phantoms for diameter 22 mm	143

LIST OF FIGURES

FIGURE NO.	TITLES	PAGE
1.1	General system configuration of process tomography	2
1.2	Types of sensing techniques	4
2.1	ERT system using ECT sensor [76]; A. ECT sensor, B. Switching unit, C. Impedance analyser (Agilent 4294A)	19
2.2	(a–b) Segmented non-invasive ERT system; (c) example of reconstructed image for annular flow [20]	21
2.3	Example of comparison between reconstructed image using LBP versus FBP	35
3.1	Circle drawing for non-invasive ERT system in COMSOL Multiphysics	40
3.2	Extra fine meshing model	42
3.3	Non-invasive ERT sensor and its equivalent circuit	43
3.4	One pair of the measurement electrodes for non-invasive ERT	45
3.5	Schematic diagram for one pair of electrode measurement	46
3.6	Current distribution with different frequencies for: (a) 100 kHz; (b) 500 kHz; (c) 1 MHz; and (d) 2 MHz	47
3.7	Voltage distribution with different frequencies for: (a) 100 kHz; (b) 500 kHz; (c) 1 MHz; and (d) 2 MHz	48
3.8	Absolute voltage versus location of fifteen points measured from excitation electrode (point 1) to the detection electrode (point 15)	49
3.9	Normalised current distribution for the increment of	

	frequencies	50
3.10	Surface current distribution and electric field distribution with different frequencies for: (a) 100 kHz; (b) 500 kHz; (c) 1 MHz; and (d) 2 MHz	51
3.11	Normalised current distribution for a different width at a different height (90 mm to 130 mm): (a) E1-E2; (b) E1-E9	54
3.12	Normalised current distribution for a different height at a different width (9 mm to 16 mm): (a) E1-E2; (b) E1-E9	55
3.13	(a) Normalized current distribution versus detection electrode; (b) Normalized current distribution for the nearest and the furthest pair of measurement electrodes. For width 16 mm at different height (100–500 mm)	57
3.14	Normalized current distribution between homogenous and non-homogenous system for varying electrode height at different phantom placed at origin: (a) 10 mm; (b) 20 mm; (c) 30 mm	59
3.15	Sensitivity distribution at detection electrode position with varying obstacle size at origin	63
4.1	Basic drawing of 136×136 pixels to 128×128 pixels	69
4.2	Coding for eliminating thickness of pipe written in MATLAB	69
4.3	Example of illustration to eliminate pipe thickness	70
4.4	Example pairing projection between E_1 and E_9 : (a) before; and (b) after eliminating the thickness of the pipe	71
4.5	Total map of the system before and after eliminating the thickness of the pipe: (a) before; (b) after	72
4.6	Sensitivity distribution for transmitter 1 with receiver 2 to receiver 9)	74

4.7	Sensitivity distribution for transmitter 1 with receiver 10 to receiver 16	75
5.1	Experimental setup for non-invasive ERT system	84
5.2	Designed non-invasive ERT sensor	85
5.3	(a) Ring holder with sensor jigs; (b) example of a sensor jig from the side; and (c) inner views	86
5.4	Sensor jig attached to the pipe (height of pipe is 500 mm)	86
5.5	Analogue switch circuit	87
5.6	Tested square waveform at 500 kHz	88
5.7	Tested square waveform at 2 MHz	89
5.8	Tested sinusoidal waveform at 500 kHz	90
5.9	Tested sinusoidal waveform at 2MHz	91
5.10	Connection of 1pF capacitor at the sensor connector	92
5.11	Connection pin of AD9833	94
5.12	Output of 2 MHz sinusoidal waveform from AD9833	95
5.13	Connection of DG406B	96
5.14	Schematic diagram of amplifier circuit	97
5.15	Signal of amplifier circuit	98
5.16	Schematic diagram of I-to-V converter amplifier circuit	99
5.17	Example signal of the converted output voltage	100
5.18	Schematic diagram of peak detector circuit	101
5.19	Example signal of peak value from the input AC signal	102
5.20	Flow chart for measurement process	103
5.21	Timing diagram for one frame	105
5.22	PCB for non-invasive ERT system	106
5.23	Front panel of online non-invasive ERT GUI using MATLAB software	107
5.24	Online main program flowchart	108
5.25	Linear Back-Projection Algorithm	109

5.26	Front panel of offline analysis of non-invasive ERT GUI using MATLAB software	110
6.1	Measurement data in sixteen transmitter groups collected via a non-invasive ERT system in a homogenous field: (a) simulation versus (b) experiment	114
6.2	Homogeneity variance test results for each transmitter group (transmitter 1 until transmitter 8) between experiment and simulation	116
6.3	Homogeneity variance test results for each transmitter group (transmitter 9 till transmitter 16) between experiment and simulation	117
6.4	One-way ANOVA test for each source of channel 1 till channel 4	119
6.5	One-way ANOVA test for each source of channel 5 till channel 8	120
6.6	One-way ANOVA test for each source of channel 9 till channel 12	121
6.7	One-way ANOVA test for each source of channel 13 till channel 16	122
6.8	Sample of wooden rod used (left side); example tomogram obtained in online system (right side)	124
6.9	Tap water measured using a conductor meter	124
6.10	Measuring the electrical conductivity of cooking oil using conductor meter and testing for image reconstruction	125
6.11	Jig for placing the phantom at the bottom of pipe	125
6.12	Example of tomogram reconstructed from simulation and experiment	126
6.13	Example of getting threshold value from AE versus range of threshold value graph	127
6.14	Percentage of AE of different positions for blind spot test (simulation versus experiment)	132

6.15	MSSIM indexed versus different positions of phantom for blind spot	133
6.16	Concentration of solid obtained from simulation and experiment (Blind Spot): (a) 12 mm; (b) 22 mm	133
6.17	Different heights of phantoms applied. From left: 1 cm, 1.5 cm, 2 cm, 2.5 cm, 3 cm, 3.5 cm, 4 cm, and 4.5 cm	135
6.18	(a) No image detected for height 1 cm till 4 cm; (b) image detected when height was 4.5 cm; (c) reference image	136
6.19	Percentage of AE for same sizes of single phantom at different positions (simulation versus experiment)	139
6.20	MSSIM versus different positions of phantom for single phantom	139
6.21	Concentration of solid obtained from simulation and experiment (single 12 mm)	140
6.22	Concentration of solid obtained from simulation and experiment (single 22 mm)	141
6.23	Percentage of AE for same sizes of double phantoms at different positions (simulation versus experiment)	144
6.24	MSSIM versus different positions of phantom for double phantoms	144
6.25	Concentration of solid (double 12 mm)	145
6.26	Concentration of solid (double 22 mm)	145

LIST OF ABBREVIATIONS

ERT	—	Electrical resistance tomography
ECT	—	Electrical capacitance tomography
kHz	—	Kilo hertz
PSD	—	Phase-sensitive demodulation
MHz	—	Mega hertz
PT	—	Process tomography
DAS	—	Data acquisition system
EIT	—	Electrical impedance tomography
UT	—	Ultrasonic tomography
LBP	—	Linear back-projection
FEM	—	Finite element model
PVC	—	Plasticized polyvinyl chloride
OT	—	Optical tomography
EQS	—	Electro quasi-static
MQS	—	Magneto quasi-static
2D	—	Two-dimensional
PDE	—	Partial differential equation
k Ω	—	Kilo ohm
pF	—	Pico farad
mA	—	Mili ampere
3D	—	Three-dimensional
MSSIM	—	Multi scale structural similarity
AE	—	Area error
P_{Th}	—	Threshold pre-set value
I-to-V	—	Current-to-voltage
DDS	—	Direct digital synthesizer

AC	—	Alternate-Current
DC	—	Direct Current
GBP	—	Gain Bandwidth Product
ADC	—	Analogue-To-Digital Conversion
PCB	—	Printed Circuit Board
GUI	—	Graphical User Interface
ANOVA	—	Analysis Of Variance
V_{pp}	—	Peak-to-peak voltage

LIST OF SYMBOLS

R	—	Resistance
V	—	Voltage
I	—	Current
σ	—	Electrical conductivity
L	—	Outer diameter pipe
A	—	Area of electrode
G	—	Conductance
D	—	Electric flux density
E	—	Electric field intensity
J	—	Current density
ρ	—	Free charge density
B	—	Magnetic flux density
H	—	Magnetic field intensity
ω	—	Angular frequency
ϵ	—	Permittivity
μ	—	Permeability
Z	—	Impedance
C	—	Capacitance
f	—	Frequency
IM	—	Independent measurement
N	—	Total sensors
m	—	Mili
d	—	Thickness of non-conducting pipe
d	—	Outer plane thickness
M	—	Sensitivity map
G_T	—	Threshold image

\times	—	multiplication
\cdot	—	Scalar multiplication
π	—	Pi (3.142)
+/-	—	Plus or minus sign

LIST OF APPENDICES

APPENDIX	TITLE	PAGE
A	Publications	170
B	Programming code for waveform generator	174
C	Part of programming codes for DSPIC30F6010A	175
D	Part of programming codes for MATLAB	176

CHAPTER 1

INTRODUCTION

1.1 Introduction

The word ‘tomography’ comes from the Greek: the term *tomos* means to slice, and *graphein* means to write [1]. The *Oxford English Dictionary* [2] defines tomography as:

Radiography in which an image of a predetermined plane in the body or other object is obtained by rotating the detector and the source of radiation in such a way that points outside the plane give a blurred image. Also in extended use, any analogous technique using other forms of radiation.

Tomography’s introduction into the medical field started in the 1950s and led to the possibility of scanning the human body for diagnostic purposes. In medical fields, X-ray tomography was implemented firstly to image the internal human structure (bones) based on the attenuation of X-ray. This radiation-based method allows the medical staff to investigate the internal human structure or the object of interest non-invasively. As a result of this concept of tomography, the technique has become a pioneer for subsequent industrial applications.

Nevertheless, tomography used in the medical field is different from that used in industry, due to the different aims of the applications. Normally, medical tomography is used to determine a specific object in the space, whereas industrial tomography focuses on measuring phase proportions; for instance, the concentration

of mediums and the velocity of movement. The development of tomography for industrial applications evolved in the 1980s. Process Industrial Tomography (PIT), known simply as Process Tomography (PT), is a term used for industrial applications. Process tomography has become a promising technique for visualizing and analysing the internal characteristics of process plants in industry applications, such as for two-phase/multiphase flow in pipelines. This PT consists of has many advantages: low cost, non-invasive, non-intrusive, no radiation hazards, and it is suitable for different sized vessels. Thus, process tomography is one of the most important techniques in industrial processes nowadays.

There are three main parts of process tomography: the sensing system; interfacing; and an image reconstruction algorithm for displaying the tomogram, as shown in Figure 1.1. The sensing system includes the sensor, measurement circuits for transmitting and receiving a signal, and signal conditioning circuits to amplify and filter the signal before they are used for the interfacing part. The interfacing part refers to the data acquisition system (DAS). Then, the information from the sensing system about the medium of interest will be used via the DAS in the image reconstruction algorithm for getting and analysing the image of interest, or the ‘tomogram’. As a result, the different types of process tomography work based on the different principles of the sensors involved and implemented in the applications.




Figure 1.1: General system configuration of process tomography

Tomography can be divided into two fields: hard-field tomography; and soft-field tomography [3]. These classifications of tomography are referred to as the natural behaviour of sensors. Hard-field tomography refers to a condition where the sensitivity of the medium imaged is independent of the distribution of the measured parameters in its whole volume [4]. X-ray tomography, optical tomography, and ultrasonic tomography are examples of hard-field tomography. Soft-field tomography means that the sensitivity of the medium imaged depends on the distribution of the measured parameters in its whole volume [5]. Alternatively, this will cause a challenge in solving the inverse problem of the medium of interest. Electrical tomography can be categorized as a soft-field tomography that includes Electrical Capacitance Tomography (ECT), Electrical Impedance Tomography (EIT) and Electrical Resistance Tomography (ERT).

1.2 Sensing Technique of Process Tomography

Process tomography offers a unique opportunity to reveal the complexities of the internal structure of an object without the need to invade it. There are four types of sensing techniques for tomography: intrusive, non-intrusive, invasive, and non-invasive. The word ‘intrusive’ is related to how the sensor protrudes into the medium of interest, and ‘invasive’ means the sensor is applied to the inner surface of the wall of the pipeline. Additionally, the sensing techniques can be combined so that they can be intrusive and invasive, intrusive and non-invasive, non-intrusive and invasive, and non-intrusive and non-invasive as in [6,7]. These concepts are represented in detail in Figure 1.2. However, the non-intrusive and non-invasive is a well-known method applied in industry as it has several advantages. For example, it can avoid the contamination of pure or sterile materials, minimizing the hazards of working with poisonous, radioactive, explosive, flammable or corrosive materials, and decreasing safety and accountancy difficulties with valuable process materials. For this reason, the non-invasive and non-intrusive method is one of the favourite methods applied in process plants compared to other sensing techniques.

Sensing Technique	Invasive	Non-Invasive
Intrusive		
Non-Intrusive		

Figure 1.2: Types of sensing techniques

1.3 Research Background

A mixture such as liquid–liquid, liquid–solid and liquid–gas two-phase regime is the main concern in process industry applications. The industrial applications involving mixtures, for example, are bubble column, fluidized bed reactor, stirred tank reactor and vertical or horizontal vessel. Visualization of the mixture at an early stage in an industry application may promise good performance and prevent any unwanted conditions during the process. One of the methods that can do this prior visualization is process tomography. Process tomography is used to reconstruct the cross-sectional image of the medium of interest. As it is a non-destructive technique, process tomography has gained wide interest amongst researchers.

Electrical resistance tomography (ERT) is one of the most extensive modalities of process tomography being applied, and has been used in many

applications such as in geological surface [8–11], agriculture processes [12,13] and also in industrial processes [14]. The advantages of the ERT application are that it is relatively safe to use and it provides fast response for online and real-time monitoring of the process plant. There are many examples of ERT systems that have been studied, focusing on imaging technique of the liquid–gas, liquid–liquid or liquid–solid mixtures. However, only a few researchers have considered ERT with a non-invasive sensing technique [15–21]. At present, non-invasive process tomography has become a promising imaging technique for monitoring, with potentially enormous applications for mixtures analysis. Therefore, the non-invasive ERT system for measuring and imaging mixtures of a static two-phase regime in a pipeline is proposed for this research.

1.4 Problem Statements

The following are the problems that need to be considered in the design of a non-invasive electrical resistance tomography system for liquid–solid profile measurement.

- i. The conventional technique of ERT is applied invasively to the pipeline and causes the metal electrodes to have direct contact with the conductive liquid. The contact between the electrodes and the conductive liquid, such as electrolyte, causes an oxidation to the electrode, also known as an electrochemical erosion effect. This situation leads to electrode corrosion [16].
- ii. The current researches into ERT systems produces inconsistency in measurements with unpredictable error [21]. Improper, invasive installation of the electrodes enables the electrodes to produce small bubbles around it when energized by the current signal. Thus, the signal cannot be transmitted and received appropriately.
- iii. The use of an invasive ERT system in industrial applications is limited [19]. Limitation of conventional ERT is because the electrodes might affect the

nature of the process flow, because the contact between metal electrode and conductive liquid can produce a chemical reaction.

Consequently, the solution to these problems is to use a non-invasive approach, whereby the sensor is clamped to the outside of the pipe wall.

1.5 Aim and Research Objectives

The aim of this research is to design and develop a non-invasive ERT system for a liquid–solid two–phase regime. The objectives of this research are as follows:

- i. To design and develop a suitable ERT measurement section using a non-invasive sensing technique, including proposing the suitable excitation frequency and its electronic measurement systems.
- ii. To interface the hardware and software for system validation. This validation is demonstrated by the concentration profile computations using online configuration.
- iii. To reconstruct the image of a phantom position, using non-invasive ERT system.

1.6 Research Scopes

In developing the non-invasive electrical resistance tomography system to monitor the two-phase regimes of liquid–solid; the scopes of the research will comprise of:

- i. The sensor consists of sixteen electrodes that perform as transceiver-sensing operation and clamped on an experimental pipe non-invasively.
- ii. Non-invasive ERT development and analysis is conducted on a vertical non-conducting (acrylic) pipe with 100 mm outer diameter and 2 mm wall

thickness. The specification of the pipe used in this research does not reflect a vessel applied in industrial applications.

- iii. The experimental will be only tested for static flow with tap water and a wooden rod as the conductive liquid and phantom, respectively.
- iv. Only one set of measurement data is collected for every experiment conducted in the research. The collection of data is through online configuration.
- v. The image reconstruction application is based on existing back-projection algorithms, with no new derivation of an image reconstruction algorithm in this research. The tomogram is reconstructed based on a linear back-projection algorithm through online configuration, and improved using a global thresholding technique through offline configuration.

1.7 Motivation and Contribution

The development of ERT systems in existing studies indicates that most studies have focused on the invasive ERT system. Only several papers have studied the non-invasive ERT system technique [15–21]. However, the presented papers did not particularly discuss hardware designation and development. Therefore, the work in this thesis has thus designed and developed a non-invasive ERT system using 2 MHz frequency that can produce an analysis for multiphase mixtures.

In addition, most of the presented works in the cited literature for ECT and ERT develop the system with a frequency in kHz. At this point, a phase-sensitive demodulation (PSD) method using sophisticated circuit design also needs to be developed in conjunction with the ECT or ERT system, to measure the internal permittivity or internal conductivity. Thus, in this work, it is proven that by developing the non-invasive ERT system with a high frequency level, the PSD technique is not required. The utilization of a high frequency level enables the system to detect the concentration profile of the medium of interest.

Furthermore, corrosion-free electrodes and no contact with the flow in the medium in the non-invasive ERT system introduced a non-invasive conductivity conductor for industry. Thus, it can provide the system with better sensing accuracy.

Moreover, any damage of the sensor occurring in the system can easily be replaced and dealt with in non-invasive ERT. The independent and flexible design of the research enables it to reduce the duration of the troubleshooting process.

1.8 Structure of Thesis

The thesis consists of seven chapters as summarized below:

- i. Chapter 1 briefly describes the background of tomography, types of sensing techniques, research background, problem statement, the aim and research objectives, scopes and the study's contribution.
- ii. Chapter 2 is the literature review of non-invasive process tomography, current works related the electrical resistance tomography (ERT), the basic principles of non-invasive ERT and also image reconstruction applied in industrial tomography.
- iii. Chapter 3 presents the modelling of the non-invasive ERT system that focuses on the optimum frequency source and optimum electrode dimension using COMSOL software.
- iv. Chapter 4 discusses solving the forward problem of the system, back-projection algorithm for image reconstruction, the image quality assessment and improving the tomogram using a thresholding technique.
- v. Chapter 5 provides details of the hardware and measurement systems, such as the design of electronic measurement circuits, hardware jig and metal sensor, together with the software part of the developed system.
- vi. Chapter 6 presents the experiments, results and discussion and compares these with the simulation results. It involves two parts; sensor performance analysis and image reconstructed analysis.
- vii. Chapter 7 provides the conclusion to the study and suggestions for future work.

REFERENCES

- [1] R. B. Northrop, *Noninvasive Instrumentation and Measurement in medical Diagnosis*. New York: CRC Press, 2002.
- [2] R. A. Williams and M. S. Beck, *Process Tomography: Principles, Techniques and Applications*, 1 st. ed. Oxford: Butterworth-Heinemann, 1995.
- [3] T. York, H. McCann, and K. Ozanyan, “Agile Sensing Systems for Tomography,” *IEEE Sensors Journal*, vol. 11, no. 12, pp. 3086–3105, 2011.
- [4] U. Z. Ijaz, J. Kim, A. K. Khambampati, M. Kim, S. Kim, and K. Kim, “Concentration Distribution Estimation of Fluid through Electrical Impedance Tomography based on Interacting Multiple Model Scheme,” *Flow Measurement and Instrumentation*, vol. 18, no. 1, pp. 47–56, 2007.
- [5] S. C. Murphy and T. A. York, “Electrical Impedance Tomography with Non-Stationary Electrodes,” *Measurement Science and Technology*, vol. 17, no. 11, pp. 3042–3052, 2006.
- [6] R. C. Asher, “Ultrasonic Sensors in the Chemical and Process Industries,” *Journal of Physics E: Scientific Instruments*, vol. 16, no. 10, pp. 959–963, 1983.
- [7] G. Steiner and F. Podd, “A Non-Invasive and Non-Intrusive Ultrasonic Transducer Array For Process Tomography,” in *Proceedings of the XVIII IMEKO World Congress Metrology for a Sustainable Development*, September 17-22, 2006, Rio de Janeiro, Brazil, 2006, pp. 17–22.
- [8] E. Auken, J. Doetsch, G. Fiandaca, A. V. Christiansen, A. Gazoty, A. G. Cahill, and R. Jakobsen, “Imaging Subsurface Migration of Dissolved CO₂ in A Shallow Aquifer using 3-D Time-Lapse Electrical Resistivity Tomography,” *Journal of Applied Geophysics*, vol. 101, pp. 31–41, Feb. 2014.
- [9] S. D. Carrière, K. Chalikakis, G. Sénéchal, C. Danquigny, and C. Emblanch,

- “Combining Electrical Resistivity Tomography and Ground Penetrating Radar to study geological structuring of karst Unsaturated Zone,” *Journal of Applied Geophysics*, vol. 94, pp. 31–41, Jul. 2013.
- [10] J. Qiyun, H. Jishan, S. Ping, and P. Jing, “The Hardware Design of A Mine Resistivity Tomography Instrument,” in *Proceedings of the 2009 IEEE International Workshop on Imaging Systems and Technique*, May 11-12, 2009, Shenzhen, China, 2009, pp. 235–238.
- [11] R. L. Newmark, A. L. Ramirez, and W. D. Daily, “Monitoring Carbon Dioxide Sequestration using Electrical Resistance Tomography (ERT): A Minimally Invasive Method,” *Greenhouse Gas Control Technologies*, vol. I, pp. 353–358, 2003.
- [12] S. Garré, I. Coteur, C. Wongleecharoen, K. Hussain, W. Omsunarn, T. Kongkaew, T. Hilger, J. Diels, and J. Vanderborght, “Can We Use Electrical Resistivity Tomography to Measure Root Zone Dynamics in Fields with Multiple Crops?,” *Procedia Environmental Sciences*, vol. 19, pp. 403–410, 2013.
- [13] J. Boaga, M. Rossi, and G. Cassiani, “Monitoring Soil-plant Interactions in an Apple Orchard Using 3D Electrical Resistivity Tomography,” *Procedia Environmental Sciences*, vol. 19, pp. 394–402, 2013.
- [14] M. Sharifi and B. Young, “Electrical Resistance Tomography (ERT) Applications to Chemical Engineering,” *Chemical Engineering Research and Design*, vol. 91, pp. 1625–1645, 2013.
- [15] Z. Cao, L. Xu, C. Xu, and H. Wang, “Electrical Resistance Tomography (ERT) by using An ECT Sensor,” in *Proceedings of the 2010 IEEE International Conference on Imaging Systems and Techniques*, July 1-2, 2010, Thessaloniki, Greece, 2010, pp. 63–66.
- [16] B. L. Wang, Z. Y. Huang, H. F. Ji, and H. Q. Li, “Towards Capacitively Coupled Electrical Resistance Tomography,” in *Proceedings of the 6th World Congress on Industrial Process Tomography*, September 6-9, 2010, Beijing, China, 2010, pp. 1574–1577.
- [17] B. Wang, Y. Hu, H. Ji, Z. Huang, and H. Li, “A Novel Electrical Resistance Tomography System Based on C4D Technique,” in *Proceedings of the 2012 IEEE International Conference on Instrumentation and Measurement Technology*, May 13-16, 2012, Graz, Austria, 2012, pp. 1929–1932.

- [18] Y. Li and M. Soleimani, "Imaging Conductive Materials with High Frequency Electrical Capacitance Tomography," *Measurement*, vol. 46, no. 9, pp. 3355–3361, 2013.
- [19] B. Wang, Y. Hu, H. Ji, Z. Huang, and H. Li, "A Novel Electrical Resistance Tomography System Based on C4D Technique Technique," *IEEE Transactions on Instrumentation and Measurement*, vol. 62, no. 5, pp. 1017–1024, 2013.
- [20] B. Wang, W. Zhang, Z. Huang, H. Ji, and H. Li, "Modeling and Optimal Design of Sensor for Capacitively Coupled Electrical Resistance Tomography System," *Flow Measurement and Instrumentation*, vol. 31, pp. 3–9, 2013.
- [21] B. Wang, W. Tan, Z. Huang, H. Ji, and H. Li, "Image Reconstruction Algorithm for Capacitively Coupled Electrical Resistance Tomography," *Flow Measurement and Instrumentation*, vol. 40, pp. 216–222, 2014.
- [22] J. L. Hubers, A. C. Striegel, T. J. Heindel, J. N. Gray, and T. C. Jensen, "X-Ray Computed Tomography in Large Bubble Columns," *Chemical Engineering Science*, vol. 60, pp. 6124–6133, 2005.
- [23] T. J. Heindel, J. N. Gray, and T. C. Jensen, "An X-Ray System for Visualizing Fluid Flows," *Flow Measurement and Instrumentation*, vol. 19, pp. 67–78, 2008.
- [24] A. Shaikh and M. Al-Dahhan, "Characterization of The Hydrodynamic Flow Regime in Bubble Columns via Computed Tomography," *Flow Measurement and Instrumentation*, vol. 16, pp. 91–98, 2005.
- [25] C. Wu, Y. Cheng, Y. Ding, F. Wei, and Y. Jin, "A Novel X-Ray Computed Tomography Method for Fast Measurement of Multiphase Flow," *Chemical Engineering Science*, vol. 62, pp. 4325–4335, 2007.
- [26] Z. Zhang, M. Bieberle, F. Barthel, L. Szalinski, and U. Hampel, "Investigation of Upward Cocurrent Gas–Liquid Pipe Flow using Ultrafast X-Ray Tomography and Wire-Mesh," *Flow Measurement and Instrumentation*, vol. 32, pp. 111–118, 2013.
- [27] H. Prasser, M. Misawa, and I. Tiseanu, "Comparison Between Wire-Mesh Sensor and Ultra-Fast X-Ray Tomograph for An Air–Water Flow in A Vertical Pipe," *Flow Measurement and Instrumentation*, vol. 16, pp. 73–83, 2005.
- [28] S. Boden, M. Bieberle, and U. Hampel, "Quantitative Measurement of Gas

- Hold-Up Distribution in a Stirred Chemical Reactor using X-Ray Cone-Beam Computed Tomography,” *Chemical Engineering Journal*, vol. 139, no. 2, pp. 351–362, 2008.
- [29] R. Abdul Rahim, N. W. Nyap, M. H. Fazalul Rahiman, and C. K. San, “Determination of Water and Oil Flow Composition Using Ultrasonic Tomography,” *Jurnal Teknologi*, vol. 9, no. 1, pp. 19–23, 2007.
- [30] M. H. Fazalul Rahiman, Z. Zakaria, R. Abdul Rahim, and W. N. Ng, “Ultrasonic Tomography Imaging Simulation of Two-Phase Homogeneous Flow,” *Sensor Review*, vol. 29, no. 3, pp. 266–276, 2009.
- [31] G. Steiner and C. Deinhammer, “Ultrasonic Time-of-Flight Techniques for Monitoring Multi-Component Processes,” *Elektrotechnik und Informationstechnik*, vol. 126, no. 5, pp. 200–205, 2009.
- [32] S. Wöckel, U. Hempel, and J. Auge, “Phase Boundary Characterization in Liquids by Acoustic Waves,” *Measurement Science and Technology*, vol. 20, no. 12, pp. 1–6, 2009.
- [33] M. H. Fazalul Rahiman and R. Abdul Rahim, “Development of Ultrasonic Transmission-Mode Tomography for Water-Particles Flow,” *Sensors & Transducer*, vol. 117, no. 6, pp. 99–105, 2010.
- [34] I. R. Muhamad, Y. A. Wahab, and S. Saat, “Identification of Water/Solid Flow Regime Using Ultrasonic Tomography,” in *Proceedings of the 2012 IEEE International Conference on System Engineering and Technology*, September 11-12, 2012, Bandung, Indonesia, 2012, pp. 1–5.
- [35] Z. Zakaria, M. H. Fazalul Rahiman, and R. Abdul Rahim, “Simulation of the Two-Phase Liquid–Gas Flow through Ultrasonic Transceivers Application in Ultrasonic Tomography,” *Sensors & Transducer*, vol. 112, no. 1, pp. 24–38, 2010.
- [36] M. D. Supardan, Y. Masuda, A. Maezawa, and S. Uchida, “The Investigation of Gas Holdup Distribution in A Two-Phase Bubble Column using Ultrasonic Computed Tomography,” *Chemical Engineering Journal*, vol. 130, no. 2–3, pp. 125–133, 2007.
- [37] M. H. Fazalul Rahiman, Z. Zakaria, and R. Abdul Rahim, “Ultrasonic Process Tomographic Imaging Sensor: An Approach Utilising Transceivers Method,” in *Proceedings of the 2008 IEEE International Conference on Computer and Communication Engineering*, May 13-15, 2008, Kuala Lumpur, Malaysia,

- 2008, pp. 1147–1150.
- [38] M. H. Fazalul Rahiman, R. Abdul Rahim, H. Abdul Rahim, and N. M. Nor Ayob, “Novel Adjacent Criterion Method for Improving Ultrasonic Imaging Spatial Resolution,” *IEEE Sensors Journal*, vol. 12, no. 6, pp. 1746–1747, 2012.
- [39] M. H. F. Rahiman, R. A. Rahim, H. A. Rahim, N. M. N. Ayob, E. J. Mohamad, and Z. Zakaria, “Modelling Ultrasonic Sensor for Gas Bubble Profiles Characterization of Chemical Column,” *Sensors and Actuators B: Chemical*, vol. 184, pp. 100–105, 2013.
- [40] M. H. F. Rahiman, R. A. Rahim, H. A. Rahim, E. J. Mohamad, Z. Zakaria, and S. Z. M. Muji, “An Investigation on Chemical Bubble Column using Ultrasonic Tomography for Imaging of Gas Profiles,” *Sensors and Actuators B: Chemical*, vol. 202, pp. 46–52, 2014.
- [41] N. M. Nor Ayob, S. Yaacob, Z. Zakaria, M. H. Fazalul Rahiman, R. Abdul Rahim, and M. R. Manan, “Improving Gas Component Detection of An Ultrasonic Tomography System for Monitoring Liquid/Gas Flow,” in *Proceedings of the 2010 6th IEEE International Colloquium on Signal Processing & Its Applications*, May 21-23, 2010, Melaka, Malaysia, 2010, pp. 278–282.
- [42] N. M. Nor Ayob, M. J. Puspanathan, R. Abdul Rahim, M. H. Fazalul Rahiman, F. R. Mohd Yunus, S. Buyamin, I. M. Abd Rahim, and Y. Md. Yunos, “Design Consideration for Front-End System in Ultrasonic Tomography,” *Jurnal Teknologi*, vol. 64, no. 5, pp. 53–58, 2013.
- [43] N. M. Nor Ayob, M. H. Fazalul Rahiman, Z. Zakaria, S. Yaacob, R. Abdul Rahim, and M. R. Manan, “Simulative Study in Liquid/Gas Two-Phase Flow Measurement for Dual-Plane Ultrasonic Transmission-Mode Tomography,” *Jurnal Teknologi*, vol. 54, pp. 79–94, 2011.
- [44] R. Abdul Rahim, J. F. Pang, and K. S. Chan, “Optical Tomography Sensor Configuration using Two Orthogonal and Two Rectilinear Projection Arrays,” *Flow Measurement and Instrumentation*, vol. 16, no. 5, pp. 327–340, 2005.
- [45] E. Schleicher, M. J. Da Silva, S. Thiele, A. Li, E. Wollrab, and U. Hampel, “Design of An Optical Tomograph for the Investigation of Single- and Two-Phase Pipe Flows,” *Measurement Science and Technology*, vol. 19, no. 9, pp. 1–14, 2008.

- [46] Y. Md. Yunus, R. Abdul Rahim, and R. G. Green, "Initial Result on Measurement of Gas Volumetric Flow Rate in Gas/Liquid Mixtures using Linear CCD," *Jurnal Teknologi*, vol. 48, no. D, pp. 1–12, 2008.
- [47] S. Ibrahim, M. A. M. Yunus, R. G. Green, and K. Dutton, "Concentration Measurements of Bubbles in A Water Column using An Optical Tomography System," *ISA Transactions*, vol. 51, no. 6, pp. 821–826, 2012.
- [48] N. S. Mohd Fadzil, R. Abdul Rahim, M. S. Karis, S. Z. Mohd Muji, M. F. Abdul Sahib, M. S. B. Mansor, N. M. Nor Ayob, M. F. Jumaah, and M. Z. Zawahir, "Hardware Design of Laser Optical Tomography System for Detection of Bubbles Column," *Jurnal Teknologi*, vol. 64, no. 5, pp. 69–73, 2013.
- [49] W. Wang and H. Yang, "Imaging A Rectangular Fluidised Bed by Electrical Capacitance Tomography: Preliminary Results," *Sensor Review*, vol. 31, no. 4, pp. 315–320, 2011.
- [50] W. Yang, "Key Issues in Designing Capacitance Tomography Sensors," in *Proceedings of the IEEE Sensors*, October 22-25, 2006, Daegu, Korea, 2006, pp. 497–505.
- [51] D. Chen and Y. Han, "An Image Data Capture System for Electrical Capacitance Tomography of Oil/Water Two-Phase Flow," in *Proceedings of the 2006 IEEE International Conference on Information Acquisition*, August 20 - 23, 2006, Shandong, China, 2006, pp. 722–726.
- [52] H. Y. Cao, X. M. Duan, and H. X. Wang, "Design of Electrical Capacitance Tomography Hardware System," *Advanced Materials Research*, vol. 753–755, pp. 2311–2315, 2013.
- [53] N. T. Ali Othman, H. Obara, and M. Takei, "Cross-Sectional Capacitance Measurement of Particle Concentration in A Microchannel with Multi-Layered Electrodes," *Flow Measurement and Instrumentation*, vol. 31, pp. 47–54, 2013.
- [54] W. Q. Yang, A. Chondronasios, S. Natrass, V. T. Nguyen, M. Betting, I. Ismail, and H. McCann, "Adaptive Calibration of A Capacitance Tomography System for Imaging Water Droplet Distribution," *Flow Measurement and Instrumentation*, vol. 15, no. 5–6, pp. 249–258, 2004.
- [55] C. Yu, Z. Jian, and C. Deyun, "Two-Phase Flow Parameters Measurement and Gauss-Newton Image Reconstruction Algorithm for Electrical Capacitance

- Tomography,” in *Proceedings of the 2009 IEEE International Conference on Industrial Mechatronics and Automation*, December 9-11, 2009, Christchurch, New Zealand, 2009, pp. 192–195.
- [56] F. Teixeira and L. Fan, “Advances in Electrical Capacitance Tomography”, Ph.D. Thesis, The Ohio State University, 2006.
- [57] W. A. Al-Masry, E. M. Ali, S. A. Alshebeili, and F. M. Mousa, “Non-Invasive Imaging of Shallow Bubble Columns using Electrical Capacitance Tomography,” *Journal of Saudi Chemical Society*, vol. 14, no. 3, pp. 269–280, 2010.
- [58] R. Banasiak, R. Wajman, T. Jaworski, P. Fiderek, H. Fidos, J. Nowakowski, and D. Sankowski, “Study on Two-Phase Flow Regime Visualization and Identification using 3D Electrical Capacitance Tomography and Fuzzy-Logic Classification,” *International Journal of Multiphase Flow*, vol. 58, pp. 1–14, 2014.
- [59] J. C. Gamio, J. Castro, L. Rivera, J. Alamilla, F. Garcia-Nocetti, and L. Aguilar, “Visualisation of Gas–Oil Two-Phase Flows in Pressurised Pipes using Electrical Capacitance Tomography,” *Flow Measurement and Instrumentation*, vol. 16, no. 2–3, pp. 129–134, 2005.
- [60] Z. Huang, D. Xie, H. Zhang, and H. Li, “Gas–Oil Two-phase Flow Measurement using An Electrical Capacitance Tomography System and A Venturi Meter,” *Flow Measurement and Instrumentation*, vol. 16, no. 2–3, pp. 177–182, 2005.
- [61] C. Ortiz-Alemán and R. Martin, “Inversion of Electrical Capacitance Tomography Data By Simulated Annealing: Application to Real Two-Phase Gas–Oil Flow Imaging,” *Flow Measurement and Instrumentation*, vol. 16, no. 2–3, pp. 157–162, 2005.
- [62] D. Xie, Z. Huang, H. Ji, and H. Li, “An Online Flow Pattern Identification System for Gas–Oil Two-Phase Flow Using Electrical Capacitance Tomography,” *IEEE Transactions on Instrumentation and Measurement*, vol. 55, no. 5, pp. 1833–1838, 2006.
- [63] E. J. Mohamad, R. Abdul Rahim, L. P. Ling, M. H. Fazalul Rahiman, O. M. F. Marwah, and N. M. Nor Ayob, “Segmented Capacitance Tomography Electrodes : A Design and Experimental Verifications,” *IEEE Sensors Journal*, vol. 12, no. 5, pp. 1589–1598, 2012.

- [64] Y. Zhao, H. Yeung, E. E. Zorgani, A. E. Archibong, and L. Lao, "High Viscosity Effects on Characteristics of Oil and Gas Two-Phase Flow in Horizontal Pipes," *Chemical Engineering Science*, vol. 95, pp. 343–352, 2013.
- [65] R. Zhang, Q. Wang, H. Wang, M. Zhang, and H. Li, "Data Fusion in Dual-Mode Tomography for Imaging Oil–Gas Two-Phase Flow," *Flow Measurement and Instrumentation*, vol. 37, pp. 1–11, 2014.
- [66] F. Dong, C. Xu, Z. Zhang, and S. Ren, "Design of Parallel Electrical Resistance Tomography System for Measuring Multiphase Flow," *Chinese Journal of Chemical Engineering*, vol. 20, no. 2, pp. 368–379, 2012.
- [67] H. Zhou, L. Xu, Z. Cao, J. Hu, and X. Liu, "Image Reconstruction for Invasive ERT in Vertical Oil Well Logging," *Chinese Journal of Chemical Engineering*, vol. 20, no. 2, pp. 319–328, 2012.
- [68] S. Kim, A. N. Nkaya, and T. Dyakowski, "Measurement of Mixing of two Miscible Liquids in A Stirred Vessel with Electrical Resistance Tomography," *International Communications in Heat and Mass Transfer*, vol. 33, no. 9, pp. 1088–1095, 2006.
- [69] F. Ricard, C. Brechtelsbauer, X. Y. Xu, and C. J. Lawrence, "Monitoring of Multiphase Pharmaceutical Processes using Electrical Resistance Tomography," *Chemical Engineering Research and Design*, vol. 83, no. 7, pp. 794–805, 2005.
- [70] Y. H. Liu, N. Wang, H. Y. Sun, and W. Chen, "Solid-Liquid Two-Phase Flow Image Reconstruction in Microchannel Based on Tikhonov Regularization Method," *Applied Mechanics and Materials*, vol. 333–335, pp. 1013–1019, 2013.
- [71] M. Sharifi and B. Young, "Towards An Online Milk Concentration Sensor using ERT: Correlation of Conductivity, Temperature and Composition," *Journal of Food Engineering*, vol. 116, no. 1, pp. 86–96, 2013.
- [72] M. Sharifi and B. Young, "The Potential Utilisation of Electrical Resistance Tomography (ERT) in Milk Powder Processing for Monitoring and Control," in *Proceedings of the 6th World Congress on Industrial Process Tomography*, September 6-9, 2010, Beijing, China, 2005, pp. 1–14.
- [73] M. Sharifi and B. Young, "3-Dimensional Spatial Monitoring of Tanks for The Milk Processing Industry using Electrical Resistance Tomography,"

- Journal of Food Engineering*, vol. 105, no. 2, pp. 312–319, 2011.
- [74] M. Sharifi and B. Young, “Qualitative Visualization and Quantitative Analysis of Milk Flow using Electrical Resistance Tomography,” *Journal of Food Engineering*, vol. 112, no. 3, pp. 227–242, 2012.
- [75] S. Hosseini, D. Patel, F. Ein-Mozaffari, and M. Mehrvar, “Study of Solid–Liquid Mixing in Agitated Tanks through Electrical Resistance Tomography,” *Chemical Engineering Science*, vol. 65, no. 4, pp. 1374–1384, 2010.
- [76] S. A. Razzak, S. Barghi, and J.-X. Zhu, “Application of Electrical Resistance Tomography on Liquid–Solid Two-Phase Flow Characterization in An LSCFB Riser,” *Chemical Engineering Science*, vol. 64, no. 12, pp. 2851–2858, 2009.
- [77] S. J. Stanley and G. T. Bolton, “A Review of Recent Electrical Resistance Tomography (ERT) Applications for Wet Particulate Processing,” *Particle & Particle Systems Characterization*, vol. 25, no. 3, pp. 207–215, Sep. 2008.
- [78] A. Madupu, A. Mazumdar, J. Zhang, D. Roelant, and R. Srivastava, “Electrical Resistance Tomography for Real-Time Mapping of The Solid-Liquid Interface in Tanks Containing Optically Opaque Fluids,” in *Proceedings of the SPIE-IS&T Electronic Imaging - Computational Imaging III*, January 17-18, 2005, California, USA, 2005, vol. 5674, pp. 36–46.
- [79] M. Sobri, A. Ahmad, M. Irwan, and S. Jantan, “ERT Visualization of Gas Dispersion Performance of Aerofoil and Radial Impellers in An Agitated Vessel,” *Jurnal Teknologi*, vol. 64, no. 5, pp. 75–78, 2013.
- [80] A. D. Okonkwo, M. Wang, and B. Azzopardi, “Characterisation of A High Concentration Ionic Bubble Column using Electrical Resistance Tomography,” *Flow Measurement and Instrumentation*, vol. 31, pp. 69–76, 2013.
- [81] W. Yenjaichon, J. R. Grace, C. Jim Lim, and C. P. J. Bennington, “Characterisation of Gas Mixing in Water and Pulp-Suspension Flow Based on Electrical Resistance Tomography,” *Chemical Engineering Journal*, vol. 214, pp. 285–297, 2013.
- [82] C. Yang, H. Wang, and Z. Cui, “Application of Electrical Resistance Tomography in Bubble Columns for Volume Fraction Measurement,” in *Proceedings of the 2012 IEEE International Conference on Instrumentation and Measurement Technology*, May 13-16, 2012, Graz, Austria, 2012, pp.

- 1199–1203.
- [83] X. Deng, G. Li, Z. Wei, Z. Yan, and W. Yang, “Theoretical Study of Vertical Slug Flow Measurement by Data Fusion from Electromagnetic Flowmeter and Electrical Resistance Tomography,” *Flow Measurement and Instrumentation*, vol. 22, no. 4, pp. 272–278, 2011.
- [84] J. Kourunen, T. Niitti, and L. M. Heikkinen, “Application of Three-Dimensional Electrical Resistance Tomography to Characterize Gas Holdup Distribution in Laboratory Flotation Cell,” *Minerals Engineering*, vol. 24, pp. 1677–1686, 2011.
- [85] H. Jin, S. Yang, G. He, M. Wang, and R. A. Williams, “The Effect of Gas-Liquid Counter-Current Operation on Gas Hold-Up in Bubble Columns using Electrical Resistance Tomography,” *Journal of Chemical Technology & Biotechnology*, vol. 85, no. 9, pp. 1278–1283, 2010.
- [86] Y. Xu, H. Wang, Z. Cui, and F. Dong, “Application of Electrical Resistance Tomography for Slug Flow Measurement in Gas-Liquid Flow of Horizontal Pipe,” in *Proceedings of the 2009 IEEE International Workshop on Imaging Systems and Techniques*, May 11-12, 2009, Shenzhen, China, 2009, pp. 319–323.
- [87] Z. Meng, Z. Huang, B. Wang, H. Ji, and H. Li, “Flowrate Measurement of Air-Water Two-Phase Flow using An Electrical Resistance Tomography Sensor and A Venturi Meter,” in *Proceedings of the 2009 IEEE International Conference on Instrumentation and Measurement Technology*, May 5-7, 2009, Singapore, 2009, pp. 118–121.
- [88] M. Wang, G. Lucas, Y. Dai, N. Panayotopoulos, and R. A. Williams, “Visualisation of Bubbly Velocity Distribution in A Swirling Flow using Electrical Resistance Tomography,” *Particle & Particle Systems Characterization*, vol. 23, no. 3–4, pp. 321–329, 2006.
- [89] F. Dong, Y. Xu, L. Hua, and H. Wang, “Two Methods for Measurement of Gas-Liquid Flows in Vertical Upward Pipe using ERT System,” *IEEE Transactions on Instrumentation and Measurement*, vol. 55, no. 5, pp. 1576–1586, 2006.
- [90] C. K. Alexander and M. N. Sadiku, *Fundamentals of Electric Circuits*, 2nd. ed. New York: McGraw Hill, 2004.
- [91] J. William H. Hayt and John A. Buck, “Current and Conductors,” in

- Engineering Electromagnetics*, 7th. ed., New York: Mc Graw Hill, 2006, p. 122.
- [92] C. Kuo-Sheng, D. Isaacson, J. C. Newell, and D. G. Gisser, “Electrode Models for Electric Current Computed,” *IEEE Transactions on Biomedical Engineering*, vol. 36, no. 9, pp. 918–924, 1989.
- [93] F. T. Ulaby, E. Michielssen, and U. Ravaioli, “Resistance,” in *Fundamentals of Applied Electromagnetics*, 6th ed., United State of America: Pearson, 2010, p. 196.
- [94] F. R. M. Yunus, R. A. Rahim, S. R. Aw, N. M. N. Ayob, C. L. Goh, and M. J. Pusppanathan, “Simulation Study of Electrode Size in Air-Bubble Detection for Dual-Mode Integrated Electrical Resistance and Ultrasonic Transmission Tomography,” *Powder Technology*, vol. 256, pp. 224–232, 2014.
- [95] D. K. Kalluri, “Quasistatic and Static Approximations,” in *Principles of Electromagnetic Waves and Materials*, Boca Raton: CRC Press, 2013, pp. 11–16.
- [96] J. Larsson, “Electromagnetics from A Quasistatic Perspective,” *American Journal of Physics*, vol. 75, no. 3, p. 230, 2006.
- [97] J. William H. Hayt and John A. Buck, “Energy and Potential,” in *Engineering Electromagnetics*, 7th. ed., New York: McGraw Hill, 2006, p. 99.
- [98] F. T. Ulaby, E. Michielssen, and U. Ravaioli, “Displacement Current,” in *Fundamentals of Applied Electromagnetics*, 6th ed., United State of America: Pearson, 2010, pp. 299–300.
- [99] L. Pakzad, F. Ein-Mozaffari, and P. Chan, “Measuring Mixing Time in the Agitation of Non-Newtonian Fluids through Electrical Resistance Tomography,” *Chemical Engineering & Technology*, vol. 31, no. 12, pp. 1838–1845, 2008.
- [100] M. A. R. Frias, “Electrical Capacitance and Resistance Tomography with Voltage Excitation”, Master Thesis, University of Manchester, 2015.
- [101] K. J. Alme and S. Mylvaganam, “Electrical Capacitance Tomography — Sensor Models , Design , Simulations , and Experimental Verification,” *IEEE Sensors Journal*, vol. 6, no. 5, pp. 1256–1266, 2006.
- [102] Q. Marashdeh, F. L. Teixeira, and L.-S. Fan, “Electrical Capacitance Tomography,” in *Industrial Tomography: Systems and Applications*, Elsevier Ltd, 2015, pp. 3–21.

- [103] D. Y. Chen, L. L. Wang, and Y. Chen, "Analysis and Simulation of Sensor Structure Parameters and Filling Materials Characteristics for Electrical Capacitance Tomography System," *Materials Science Forum*, vol. 575–578, pp. 1217–1221, 2008.
- [104] S. Ridzuan Aw, R. Abdul Rahim, M. H. Fazalul Rahiman, F. R. Mohd Yunus, and C. L. G. Goh, "Electrical Resistance Tomography: A Review of The Application of Conducting Vessel Walls," *Powder Technology*, Jan. 2014.
- [105] M. Wang, "Electrical Impedance Tomography," in *Industrial Tomography: Systems and Applications*, Elsevier Ltd, 2015, pp. 23–59.
- [106] T. K. Bera and J. Nagaraju, "Studying the Resistivity Imaging of Chicken Tissue Phantoms With Different Current Patterns in Electrical Impedance Tomography (EIT)," *Measurement*, vol. 45, no. 4, pp. 663–682, 2012.
- [107] F. Dickin and M. Wang, "Electrical Resistance Tomography for Process Applications," *Measurement Science and Technology*, vol. 7, no. 3, pp. 247–260, 1996.
- [108] G. Dong, J. Zou, R. H. Bayford, X. Ma, S. Gao, W. Yan, M. Ge, and A. G. Equation, "The Comparison Between FVM and FEM for EIT Forward Problem," *IEEE Transactions on Magnetics*, vol. 41, no. 5, pp. 1468–1471, 2005.
- [109] G. R. Shaw, Y. Goussard, and R. Guardo, "Linearization of The Forward Problem in Electrical Impedance Tomography," in *Proceedings of the IEEE International Conference on Engineering in Medicine and Biology Society*, October 28-31, 1993, San Diego, USA, 1993, no. 2, pp. 82–83.
- [110] L. Peng, C. Mou, D. Yao, B. Zhang, and D. Xiao, "Determination of The Optimal Axial Length of The Electrode in An Electrical Capacitance Tomography Sensor," *Flow Measurement and Instrumentation*, vol. 16, no. 2–3, pp. 169–175, Apr. 2005.
- [111] J. Sun and W. Yang, "Evaluation of Fringe Effect of Electrical Resistance Tomography Sensor," *Measurement*, vol. 53, pp. 145–160, 2014.
- [112] J. Lei, S. Liu, X. Wang, and Q. Liu, "An Image Reconstruction Algorithm for Electrical Capacitance Tomography Based on Robust Principle Component Analysis," *Sensors*, vol. 13, no. 2, pp. 2076–92, 2013.
- [113] L. Peng, J. Ye, G. Lu, W. Yang, and A. E. C. T. Sensor, "Evaluation of Effect of Number of Electrodes in ECT Sensors on Image Quality," *IEEE Sensors*

- Journal*, vol. 12, no. 5, pp. 1554–1565, 2012.
- [114] E. J. Mohamad, “A Segmented Capacitance Tomography for Visualising Material Distributions in Pipeline Conveying Crude Plam Oil”, Ph.D. Thesis, Universiti Teknologi Malaysia, 2012.
- [115] W. Q. Yang and L. Peng, “Image reconstruction algorithms for electrical capacitance tomography,” *Measurement Science and Technology*, vol. 14, no. 1, pp. R1–R13, 2003.
- [116] M. H. Fazalul Rahiman, “Ultrasonic Tomography System for Liquid / Gas Bubble Column”, Ph.D. Thesis, Universiti Teknologi Malaysia, 2013.
- [117] S. Zarina, M. Muji, R. Bin, A. Rahim, M. Hafiz, and F. Rahiman, “Sensitivity Map in Parallel and Fan Beam Mode in Optical Tomography,” in *Proceedings of the 6th World Congress on Industrial Process Tomography*, September 6-9, 2010, Beijing, China, 2010, pp. 762–771.
- [118] J. Jamaludin, R. A. Rahim, H. A. Rahim, M. H. Fazalul Rahiman, S. Z. Mohd Muji, and J. M. Rohani, “Charge Coupled Device Based on Optical Tomography System in Detecting Air Bubbles in Crystal Clear Water,” *Flow Measurement and Instrumentation*, vol. 50, pp. 13–25, 2016.
- [119] S. Mohammad, “Segmented Excitation for Electrical Capacitance Tomography”, Master Thesis, Universiti Teknologi Malaysia, 2012.
- [120] Wael Abdelrahman Deabes, “Electrical Capacitance Tomography for Conductive Materials”, Ph.D. Thesis, Tennessee Technological University, 2010.
- [121] Z. Zakaria, “Non-Invasive Imaging of Liquid/Gas Flow using Ultrasonic Transmission Mode Tomography”, Ph.D. Thesis, Universiti Teknologi Malaysia, 2007.
- [122] N. M. N. Ayob, S. Yaacob, Z. Zakaria, M. H. F. Rahiman, R. A. Rahim, and M. R. Manan, “Improving Gas Component Detection of An Ultrasonic Tomography System for Monitoring Liquid / Gas Flow,” in *Proceedings of the 2010 6th IEEE International Colloquium on Signal Processing & Its Applications*, 2010, pp. 278–282.
- [123] M. H. Fazalul Rahiman, R. Abdul Rahim, and J. Puspanathan, “Two-Phase Flow Regime Identification by Ultrasonic,” *Sensors & Transducer*, vol. 116, no. 5, pp. 76–82, 2010.
- [124] M. H. Fazalul Rahiman, R. Abdul Rahim, H. Abdul Rahim, and N. M. Nor

- Ayob, "The Hardware Design Technique for Ultrasonic Process," *Sensors & Transducer*, vol. 141, no. 6, pp. 52–61, 2012.
- [125] S. Z. Mohd Muji, R. Abdul Rahim, M. H. Fazalul Rahiman, Z. Tukiran, N. M. Nor Ayob, E. J. Mohamad, and M. J. Puspanathan, "Optical Tomography: Image Improvement Using Mixed Projection of Parallel and Fan Beam Modes," *Measurement*, vol. 46, no. 6, pp. 1970–1978, Jul. 2013.
- [126] M. Wei, M. Tong, J. Hao, L. Cai, and J. Xu, "Detection of Coal Dust in A Mine Using Optical Tomography," *International Journal of Mining Science and Technology*, vol. 22, no. 4, pp. 523–527, 2012.
- [127] T. Dyakowski, L. F. C. Jeanmeure, and A. J. Jaworski, "Applications of Electrical Tomography for Gas–Solids and Liquid–Solids Flows — A review," *Powder Technology*, vol. 112, no. 3, pp. 174–192, 2000.
- [128] P. Wang, B. Guo, and N. Li, "Multi-Index Optimization Design for Electrical Resistance Tomography Sensor," *Measurement*, vol. 46, no. 8, pp. 2845–2853, 2013.
- [129] Z. Cui, H. Wang, C. Yang, D. Zhang, and Y. Geng, "Development and Application of ECT Digital System for Online Flow Measurement," in *Proceedings of the 2012 IEEE International Conference on Imaging Systems and Techniques Proceedings*, July 16-17, 2012, Manchester, United Kingdom, 2012, pp. 599–604.
- [130] E. J. Mohamad and R. Abdul Rahim, "Multiphase Flow Reconstruction in Oil Pipelines by Portable Capacitance Tomography," in *Proceedings of the 2010 IEEE SENSORS*, November 1-4, 2010, Kona, Hawaii, 2010, pp. 273–278.
- [131] W. Q. Yang and L. Peng, "Image Reconstruction Algorithms for Electrical Capacitance Tomography," *Measurement Science and Technology*, vol. 14, pp. R1–R13, 2003.
- [132] Y. Chen, L. Zhang, and D. Chen, "Two-Phase Flow Parameters Measurement and Image Reconstruction for Electrical Capacitance Tomography," in *Proceedings of the 2008 IEEE International Workshop on Education Technology and Training & International Workshop on Geoscience and Remote Sensing*, December 21-22, 2008, Shanghai, China, 2008, pp. 680–684.
- [133] M. J. Puspanathan, F. R. Yunus, N. M. Nor Ayob, R. Abdul Rahim, and F. A. Phang, "A Novel Electrical Capacitance Sensor Design For Dual

- Modality,” *Jurnal Teknologi*, vol. 5, pp. 45–47, 2013.
- [134] J. Lei and S. Liu, “Dynamic Inversion Approach for Electrical Capacitance Tomography,” *IEEE Transactions on Instrumentation and Measurement*, vol. 62, no. 11, pp. 3035–3049, 2013.
- [135] S. Y. Nor Muzakkir Nor Ayob , Mohd Hafiz Fazalul Rahiman , Zulkarnay Zakaria , Ruzairi Abdul Rahim, “Eminent Pixel Reconstruction Algorithm for Ultrasonic Tomography,” *Jurnal Teknologi*, vol. 55, no. 2, pp. 15–22, 2011.
- [136] W. Warsito and L. Fan, “Measurement of Real-Time Flow Structures in Gas–Liquid and Gas-Liquid–Solid on Systems using Electrical Capacitance Tomography (ECT),” *Chemical Engineering Science*, vol. 56, pp. 6455–6462, 2001.
- [137] G. L. Zeng, “Image Reconstruction - A tutorial,” *Computerized Medical Imaging and Graphics*, vol. 25, pp. 97–103, 2001.
- [138] T. R. Hay, R. L. Royer, H. Gao, X. Zhao, and J. L. Rose, “A Comparison of Embedded Sensor Lamb Wave Ultrasonic Tomography Approaches for Material Loss Detection,” *Smart Materials and Structures*, vol. 15, no. 4, pp. 946–951, Aug. 2006.
- [139] E. V Malyarenko, J. S. Heyman, H. H. Chen-Mayer, and R. E. Tosh, “Time-Resolved Radiation Beam Profiles in Water Obtained by Ultrasonic Tomography,” *Metrologia*, vol. 47, no. 3, pp. 208–218, Jun. 2010.
- [140] D. Sueseenak, T. Chanwimalueang, W. Narkbuekaew, K. Chitsakul, and C. Pintavirooj, “Cone-Beam X-Ray Tomography with Arbitrary-Orientation X-ray Tube,” in *Proceedings of the 2006 1st IEEE Conference on Industrial Electronics and Applications*, May 24-26, 2006, Singapore, 2006, pp. 1–4.
- [141] F. Meng, N. Zhang, and W. Wang, “Virtual Experimentation of Beam Hardening Effect in X-ray CT Measurement of Multiphase Flow,” *Powder Technology*, vol. 194, no. 1–2, pp. 153–157, 2009.
- [142] A. Contin and G. Schena, “3D Structural Analysis of Ground-Wall Insulation of ac Rotating Machines Using X-Ray Tomography,” in *Proceedings of the IEEE International Conference on Solid Dielectrics*, June 30-July 4, 2013, Bologna, Italy, 2013, pp. 279–282.
- [143] K. S. Chan, “Real Time Image Reconstruction for Fan Beam Optical Tomography System”, Master Thesis, Universiti Teknologi Malaysia, 2002.
- [144] M. Madec, J.-B. Fasquel, W. Uhring, and P. Joffre, “Optical Implementation

- of The Filtered Backprojection Algorithm,” *Optical Engineering*, vol. 46, no. 10, pp. 1–16, Oct. 2007.
- [145] D. C. Barber, B. H. Brown, and N. J. Avis, “Image Reconstruction in Electrical Impedance Tomography Using Filtered Back - Projection,” in *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, October 29–November 1, 1992, Paris, France, 1992, pp. 1691–1692.
- [146] R. Giguère, L. Fradette, D. Mignon, and P. A. Tanguy, “ERT Algorithms for Quantitative Concentration Measurement of Multiphase Flows,” *Chemical Engineering Journal*, vol. 141, no. 1–3, pp. 305–317, 2008.
- [147] W. R. B. Lionheart, “Reconstruction Algorithms for Permittivity and Conductivity Imaging,” in *Proceedings of the 2001 World Congress on Industrial Process Tomography*, August 29–31, 2001, Hannover, Germany, 2001, pp. 4–11.
- [148] Y. J. Zhang, Y. Chen, and D. Y. Chen, “Measurement of Two-Phase Flow Parameters and Image Reconstruction Based on Weighted SVD Truncated Conjugate Gradient Algorithm,” *Key Engineering Materials*, vol. 439–440, pp. 1516–1521, Jun. 2010.
- [149] X. Song, Y. Xu, and F. Dong, “An Improved Total Variation Regularization Method for Electrical Resistance Tomography,” in *Proceedings of the 2013 Chinese Intelligent Automation Conference*, , *Lecture Notes in Electrical Engineering 255*, vol. 255, Z. Sun and Z. Deng, Eds. Berlin, Heidelberg: Springer Berlin Heidelberg, 2013, pp. 603–610.
- [150] C. G. Xie, N. Reinecke, M. S. Beck, D. Mewes, and R. A. Williams, “Electrical Tomography Technique for Process Engineering Applications,” *The Chemical Engineering Journal*, vol. 56, pp. 127–133, 1995.
- [151] J. William H.Hayt and J. A. Buck, “Material Constants,” in *Engineering Electromagnetics*, 7th. ed., New York: Mc Graw Hill, 2006, pp. 552–553.
- [152] R. Thorn, G. A. Johansen, and B. T. Hjertaker, “Three-Phase Flow Measurement in The Petroleum Industry,” *Measurement Science and Technology*, vol. 24, no. 1, p. 12003, 2013.
- [153] M. Cheney and D. Isaacson, “Issues in Electrical Impedance Imaging,” *IEEE Computational Science and Engineering*, vol. 2, no. 4, pp. 53–62, 1995.
- [154] M. A. Zimam, E. J. Mohamad, R. Abdul Rahim, and L. P. Ling, “Sensor

- Modeling for Electrical Capacitance Tomography System using COMSOL Multiphysics,” *Jurnal Teknologi*, vol. 55, no. 2, pp. 33–47, 2011.
- [155] Clean Water Team (CWT), “Electrical Conductivity / Salinity Fact Sheet, FS-3.1.3.0(EC),” in *The Clean Water Team Guidance Compendium for Watershed Monitoring and Assessment, Version 2.0. Division of Water Quality, California State Water Resources Control Board (SWRCB), Sacramento, CA, 2004*, pp. 1–5.
- [156] P. R. Tortora, S. L. Ceccio, S. M. Trujillo, T. J. O’Hern, and K. a. Shollenberger, “Capacitance Measurements of Solid Concentration in Gas-Solid Flows,” *Powder Technology*, vol. 148, no. 2–3, pp. 92–101, Nov. 2004.
- [157] J. Sun and W. Yang, “A Dual-Modality Electrical Tomography Sensor for Measurement of Gas–Oil–Water Stratified Flows,” *Measurement*, vol. 66, pp. 150–160, 2015.
- [158] J. C. Sparks, *The Pythagorean Theorem: Crown Jewel of Mathematics*. Bloomington, Indiana: AuthorHouse, 2008.
- [159] Y. S. Kim, S. H. Lee, U. Z. Ijaz, K. Y. Kim, and B. Y. Choi, “Sensitivity Map Generation in Electrical Capacitance Tomography using Mixed Normalization Models,” *Measurement Science and Technology*, vol. 18, no. 7, pp. 2092–2102, 2007.
- [160] Z. Guo, F. Shao, and D. Lv, “Sensitivity Matrix Construction for Electrical Capacitance Tomography Based on The Difference Model,” *Flow Measurement and Instrumentation*, vol. 20, no. 3, pp. 95–102, 2009.
- [161] T. Dyakowski, L. F. C. Jeanmeure, and A. J. Jaworski, “Applications of Electrical Tomography for Gas–Solids and Liquid–Solids Flows — A review,” *Powder Technology*, vol. 112, no. 3, pp. 174–192, 2000.
- [162] M. H. Fazalul Rahiman and R. Abdul Rahim, “A Novel Hybrid Binary Reconstruction Algorithm for Ultrasonic Tomography,” *Sensors & Transducer*, vol. 89, no. 3, pp. 93–100, 2008.
- [163] Z. Wang, A. C. Bovik, H. R. Sheikh, and E. P. Simoncelli, “Image Quality Assessment: From Error Visibility to Structural Similarity,” *IEEE Transactions on Image Processing*, vol. 13, no. 4, pp. 600–612, 2004.
- [164] Z. Wang, A. C. Bovik, and L. Lu, “Why is Image Quality Assessment So Difficult?,” in *Proceedings of the 2002 IEEE International Conference on Acoustics Speech and Signal Processing*, May 13-17, 2002, Florida, USA,

- 2002, vol. 303, pp. 3313–3316.
- [165] Z. Wang and A. C. Bovik, “Error: Love It or Leave It?,” *IEEE Signal Processing Magazine*, vol. 26, no. January, pp. 98–117, 2009.
- [166] C. G. Xie, S. M. Huang, C. P. Lenn, A. L. Stott, and M. S. Beck, “Experimental Evaluation of Capacitance Tomographic Flow Imaging Systems using Physical Models,” in *IEE Proceedings- Circuits, Devices and Systems*, 1994, vol. 141, no. 5, pp. 357–368.
- [167] R. Abdul Rahim and M. H. Fazalul Rahiman, *Ultrasonic Tomography Non-Invasive Techniques for Flow Measurement*. Johor Bharu, Malaysia: Penerbit UTM Press, 2012.
- [168] R. C. Gonzalez and R. E. Woods, *Digital Image Processing*, 2nd. ed. New Jersey: Prentice Hall, 2010.
- [169] I. L. S. Mei, I. B. Ismail, A. Bawadi, and A. Shafquet, “Evaluation of Electrical Capacitance Tomography Thresholding Techniques for Void Fraction Measurement of Gas-Liquid System,” *Applied Mechanics and Materials*, vol. 625, pp. 439–443, 2014.
- [170] C. H. Radewan, “Digital Image Processing with Pseudo-Color,” *Acquisition and Analysis of Pictorial Data*, vol. 48, p. 50, 1975.
- [171] L. Riordan, “Programming the AD9833/AD9834,” in *An-1070 Application Note*, USA: Analog Device, 2010, pp. 1–4.
- [172] R. L. Boylestad and L. Nashelsky, “Operational Amplifiers,” in *Electronic Devices and Circuit Theory*, 9th. ed., United State of America: Pearson, 2006, pp. 608–611.
- [173] G. King, “Ask the Applications Engineer-Capacitive Loads on Op Amps,” *Analog Dialogue*, vol. 31, no. 2, pp. 19–20, 1997.
- [174] R. L. Boylestad and L. Nashelsky, “Op-Amp Applications,” in *Electronic Devices and Circuit Theory*, 9th. ed., United State of America: Pearson, 2006, p. 640.
- [175] M. A. Shayib, *Applied Statistics*, 1st. ed. London: Bookboon, 2013.
- [176] J. Sauro and J. R. Lewis, “Six Enduring Controversies in Measurement and Statistics,” in *Quantifying the User Experience Practical Statistics for User Research*, 2nd. ed., Cambridge: Elsevier Inc., 2012, pp. 241–267.
- [177] X. Liu and X. Xu, “A New Generalized P-Value Approach for Testing The Homogeneity of Variances,” *Statistics & Probability Letters*, vol. 80, no. 19–

20, pp. 1486–1491, 2010.

- [178] S. M. Ross, *Introduction to Probability and Statistics for Engineers and Scientists*, 4rd. ed. Canada: Elsevier Academic Press, 2009.
- [179] J. Sauro and J. R. Lewis, “Did We Meet or Exceed Our Goal?,” in *Quantifying the User Experience Practical Statistics for User Research*, 2nd. ed., Cambridge: Elsevier Inc., 2012, pp. 41–62.